

CONFIRMATION

Notification: 0000059865 01 02 on changes according to EN 15267 regarding certificate 0000059865_01 dated 22 March 2023

AMS designation:

CEMS II ef for CO, NO, NO₂, N₂O, SO₂, HCI, HF, NH₃, CH₄, CH₂O, TOC, O₂,

H₂O and CO₂

Manufacturer:

Gasmet Technologies Oy

Mestarintie 6 01730 Vantaa Finland

German Federal Environmental Agency (UBA)

Announcement regarding the uniform practice in monitoring emissions and ambient air dated 05 July 2023 Publication: BAnz AT 02.08.2023 B7

- Notifications as regards the uniform practice in monitoring III. emissions and ambient air:
- Notification as regards Federal Environment Agency (UBA) notices of 21 February 2018 (BAnz AT 26.03.2018 B8, chapter I number 3.2) and of 21 February 2023 (BAnz AT 20.03.2023 B6, chapter IV notification 27)

The oxygen measuring module OXITEC 500E SME5 from the CEMS II ef measuring system for the components O_2 , CO, NO, NO_2 , N_2O , SO_2 , HCI, HF, NH_3 , H_2O , CO_2 , H₂CO, CH₄ and TOC from Gasmet Technology Oy can now also be equipped with the display board Rev. 10e. For additional implementation of the display board Rev. 10e the software of the oxygen measuring module has been updated.

The current software versions of the CEMS II ef measuring system are now:

Calcmet: 12.240 with evaluation module 4.42.2

GFID: v2.22 (Calculation Process) and v3.8.c (Display Process)

OXITEC 500E: 4.14.

Statement issued by TÜV Rheinland Energy GmbH dated 31 March 2023

TÜV Rheinland Energy GmbH Cologne, 02 August 2023

i. V. Mbl.-Ing. Guido Baum

A. Dipl.-Ing. Carsten Röllig

www.umwelt-tuv.eu TRE@umwelt-tuv.eu

Tel. +49 221 806-5200

TÜV Rheinland Energy GmbH

Am Grauen Stein

51105 Cologne

Test institute accredited to EN ISO/IEC 17025 by DAkkS (German Accreditation Body). This accreditation is limited to the accreditation scope defined in the enclosure to the certificate D-PL-11120-02-00.