



Certificate number: 3055767-ts



Industrie Service

# CERTIFICATE

of product conformity (QAL 1)

Certificate number: 3055767-ts

|                      |   |
|----------------------|---|
| <b>Certified AMS</b> | Dust Monitor S305QAL for dust                             |
| <b>Manufacturer</b>  | Sintrol Oy<br>Ruosilantie 15<br>00390 Helsinki<br>Finland |

**Test institute** TÜV SÜD Industrie Service GmbH

**This is to certify that the AMS has been tested and found to comply with the standards  
DIN EN 15267-1 (2009), DIN EN 15267-2 (2009), DIN EN 15267-3 (2008) and  
DIN EN 14181 (2015).**

**Certification applies to the conditions listed in this certificate  
(the certificate consists of 6 pages).**



Certificate No.: 3055767-ts

**Publication in the German Federal Gazette  
(BAnz) of 03 May 2021**

**This certificate will expire on:  
02 May 2026**

Umweltbundesamt  
Dessau, 05 May 2021

TÜV SÜD Industrie Service GmbH  
Testing laboratory emission measurement/  
calibration  
Munich, 04 May 2021

Dr. Marcel Langner  
Head of Section II 4.1

Hans-Jörg Eisenberger

|                                     |   |
|-------------------------------------|---|
| <b>Test report</b>                  | 3055767 from 16 September 2020            |
| <b>Initial certification</b>        | 03 May 2021                               |
| <b>Certification validity until</b> | 02 May 2026 (5 years)                     |
| <b>Publication</b>                  | BAnz AT 03 May 2021 B9, chapter I, no 1.5 |

#### **Approved application**

The tested AMS is suitable for use at plants requiring authorisation and plants in accordance with the 27. BImSchV, the 30. BImSchV and the 44. BImSchV. The suitability for this application was assessed on the basis of a laboratory test and a field test of the AMS Dust Monitor S305QAL lasting over more than three months at plant according to Directive 2010/75/EU, chapter III (13. BImSchV). The measuring system is approved for ambient temperatures between -20 °C bis +50 °C.

The AMS publication, the suitability test and the performance of the uncertainty calculations were conducted based on the provisions valid at the time of testing. Due to possible amendments to legal foundations, every user should ensure before use of the AMS that it is suitable for monitoring the applicable values.

The operator should consult the manufacturer to ensure that the AMS is suitable for the plant at which it is to be installed.

#### **Certification basis**

This certificate is based on:

- TÜV SÜD Industrie Service GmbH test report 3055767 from 16 September 2020
- Suitability announcement by the German Federal Environmental Agency as relevant body
- The ongoing surveillance of the product and the manufacturing process

- Publication in the German Federal Gazette (BAnz AT 03 May 2021 B9, chapter I, no. 1.5, UBA publication from 31 March 2021)

**AMS:** Dust Monitor S305QAL for dust

**Manufacturer:** Sintrol Oy  
Helsinki, Finland

**Suitability:** For plants requiring authorisation and plants in compliance with the 27. BImSchV, the 30 BImSchV and the 44. BImSchV

**Measurement ranges in the suitability test:**

| Component | Certification range | Supplementary measurement ranges |                     | Unit              |
|-----------|---------------------|----------------------------------|---------------------|-------------------|
|           |                     | Measurement range 2              | Measurement range 3 |                   |
| Dust      | 0 – 7,5             | 0 – 15                           | 0 – 100             | mg/m <sup>3</sup> |

**Software version:**

Version: 3.2.4

**Restrictions:**

1. It cannot be used in steam saturated flue gases. Droplet emissions also influence the measured dust concentration.
2. It cannot be used directly after electric filters.
3. It can be used for flue gas flow rates in the range from 3 – 40 m/s

**Notes:**

1. The maintenance interval is three months.
2. The AMS can only be aligned using the automatic alignment function at zero and span point.
3. At a flue gas flow rate in the 3 – 40 m/s range the dependence of the flue gas flow rate is eliminated by the integrated flow rate compensation. To this end the analogue input 4 – 20 mA shall be covered with a signal to represent the flue gas flow rate.
4. At constant flue gas flow rates ( $\pm 10\%$  of the average flow rate) a fixed value can also be entered for the flue gas flow rate.
5. When using a purge air feature, adherence to the stipulated purge air amount should be checked.
6. The AMS shall be operated at an interval of 24 h for the automatic control cycle.
7. The manufacturer's recommendations on probe lengths should be followed. Probe lengths from 250 mm to 1000 mm can be used.
8. The power supply can be 230 V AC or 24 V DC.
9. The AMS has a digital Modbus interface (serial RS 485), corresponding to VDI 4201 page 1 and 3.

**Test report:** TÜV SÜD Industrie Service GmbH, Munich  
Report no.: 3055767 from 16 September 2020

### **Certified Product**

The certificate applies to AMS that comply with the following description:

The entire tested measuring system Dust Monitor S305QAL consists of the probe, the probe extension, the purge air adapter and the electronic unit attached to the probe.

The AMS Dust Monitor S305QAL is used to detect the dust mass concentration in flue and process gases. The measuring system works according to the principle of the triboelectric effect, in which an insulated probe is electrically charged by dust particles that hit it or flow past it in the vicinity, and this charge is detected. The measuring system also has a digital interface of the Modbus type, which is implemented serially as EIA-485.

The entire AMS consist of the following components:

#### **Entire system**

|                               |                      |
|-------------------------------|----------------------|
| <b>Manufacturer:</b>          | Sintrol Oy           |
| <b>Type:</b>                  | Dust Monitor S305QAL |
| <b>Software:</b>              | 3.2.4                |
| <b>Measurement principle:</b> | triboelectric effect |

#### **Accessories:**

Welded adapter MC900229  
Blind cap MC900033  
Tri-Clamp damp MC900034  
Tri-Clamp Teflon seal MC900007  
Purge air dapter MC900203  
Probe extension 250, 500 mm  
User software Dust Tool

#### **Optional accessories:**

Interface adapter RS 485 USB EC900041  
Sintrol reference signal generator

This certificate is based on the analyser tested. The manufacturer is responsible for the continuous compliance of the production to the DIN EN 15267 requirements. The manufacturer is required to maintain an approved quality management system to control the manufacture of the certified product. Regular monitoring must be conducted on both the product and the quality management systems.

If the product from the current production series no longer comply with the certified product, the Environmental Service Department of TÜV SÜD Industrie Service GmbH must be informed (address see footnote).

A certification mark with an ID-Number that is specific to the certified product is presented on page 1 of this certificate. This can be applied on the product or used in publicity material for the certified product.

This document and the certification mark shall remain the property of TÜV SÜD Industrie Service GmbH.

Should the publication be revoked, this certificate will become invalid. This document must be returned when the period of validity has elapsed and at the request of TÜV SÜD Industrie Service GmbH and the certification mark may no longer be used.

The current version of the certificate and its expiration is also accessible on the internet at **qal1.de**.

The certification of the Dust Monitor S305QAL measuring system is based on the following documents and the regular continuous monitoring of the manufacturer's quality management system:

**Initial certification in accordance with DIN EN 15267:**

|                            |                       |
|----------------------------|-----------------------|
| Certificate no. 3055767-ts | 03 May 2021           |
| Certificate validity until | 02 May 2026 (5 years) |

Report no.: 3055767 from 16 September 2020,  
TÜV SÜD Industrie Service GmbH  
Publication: BAnz AT 03 May 2021 B9, chapter I no. 1.5  
UBA publication from 31 March 2021



**Calculation of total uncertainty for QAL1 testing according to DIN EN 14181 and DIN EN 15267-3 for the measuring system Dust Monitor S305QAL**

**Total uncertainty for the measurement component dust in the measuring range 0 – 7,5 mg/m<sup>3</sup>**

| <i>Performance characteristic</i>                                  | <i>Uncertainty</i>             | <i>Value standard uncertainty mg/m<sup>3</sup></i> | <i>Quadrat der Standardunsicherheit in (mg/m<sup>3</sup>)<sup>2</sup></i> |
|--|--------------------------------|--|---|
| Lack-of-fit  | $u_{lof}$                      | -0,046   | 0,0021  |
| Zero drift from field test   | $u_{d,z}$                      | -0,003   | 0,00001   |
| Span drift from field test   | $u_{d,s}$                      | -0,027   | 0,00073   |
| Influence of ambient temperature at span                           | $u_t$                          | 0,047  | 0,00221   |
| Influence of sample gas pressure                                   | $u_p$                          |  |   |
| Influence of sample gas flow                                       | $u_f$                          |  |   |
| Influence of voltage supply  | $u_v$                          | 0,019  | 0,00036   |
| Cross-sensitivity (interference)                                   | $u_i$                          |  |   |
| Repeatability standard deviation at span                           | $u_r = s_r$                    | 0,028  | $u_r < u_d$   |
| Standard deviation from paired measurements under field cond.      | $u_d = s_d$                    | 0,042  | 0,00176   |
| Uncertainty of reference material 3 % at 80% of CR                 | $u_{rm}$                       | 0,104  | 0,0108  |
| Excursion of measurement beam                                      | $u_{mb}$                       |  |   |
| Converter efficiency for AMS measuring NOx                         | $u_{ce}$                       |  |   |
| Variation of response factors (TOC)                                | $u_{rf}$                       |  |   |
|  |                                | total  | 0,01799   |
| Combined standard uncertainty                                      | $u_c = \sqrt{\sum (u_i)^2}$    | 0,134  | mg/m <sup>3</sup>   |
| Total expanded uncertainty   | $U_{0,95} = 1,96 \times u_c$   | 0,263  | mg/m <sup>3</sup>   |
| Relativ expanded uncertainty                                       | $U$                            | 5,3  | % ELV   |
| Permissible uncertainty of EN 15267-3                              | ( of ELV 5 mg/m <sup>3</sup> ) | 22,5   | % ELV   |
| Complied with requirements relating to the measurement uncertainty |                                | yes  | regarding EN 15267-3  |
| Permissible uncertainty 13. / 17. BImSchV                          | ( of ELV 5 mg/m <sup>3</sup> ) | 30   | % ELV   |
| Complied with requirements relating to the measurement uncertainty |                                | yes  | regarding 13. / 17. BImSchV   |