

CERTIFICATE

of Product Conformity (QAL1)

Certificate No: 0000050625_02

Certified AMS: AO2000-Fidas24 for TOC

Manufacturer: ABB AG
Stierstädter Str. 5
60488 Frankfurt/Main
Germany

Test Institute: TÜV Rheinland Energy & Environment GmbH

**This is to certify that the AMS has been tested
and found to comply with the standards
EN 15267-1 (2009), EN 15267-2 (2023), EN 15267-3 (2007)
as well as EN 14181 (2014).**

Certification is awarded in respect of the conditions stated in this certificate
(this certificate contains 11 pages).
The present certificate replaces certificate 0000050625_01 dated 13 March 2021.



Suitability Tested
EN 15267
QAL1 Certified
Regular
Surveillance

www.tuv.com
ID 0000050625


Publication in the German Federal Gazette
(BAnz) of 14 March 2016

German Environment Agency


Dessau, 12 March 2026

This certificate will expire on:
13 March 2031

TÜV Rheinland
Energy & Environment GmbH
Cologne, 10 March 2026



Dr. Marcel Langner
Head of Section II 4



i.V. Guido Baum

www.umwelt-tuv.eu
qal1-info@tuv.com
Tel. + 49 221 806-5200

TÜV Rheinland Energy & Environment GmbH
Am Grauen Stein
51105 Köln

Test institute accredited to EN ISO/IEC 17025 by DAkkS (German Accreditation Body).
This accreditation is limited to the accreditation scope defined in the enclosure to the certificate D-PL-11120-02-00.

Test report:	936/21228173/A dated 21 October 2015
Initial certification:	14 March 2016
Expiry date:	13 March 2031
Certificate:	Renewal (of previous certificate 0000050625_01 of 13 March 2021 valid until 13 March 2026)
Publication:	BAnz AT 14.03.2016 B7, chapter I No. 3.1

Approved application

The tested AMS is suitable for use at plants according to Directive 2010/75/EC, chapter III (combustion plants / 13th BImSchV:2015), chapter IV (waste incineration plants / 17th BImSchV:2013), Directive 2015/2193/EC (44th BImSchV:2019), TA Luft:2002, 30th BImSchV:2009 and 27th BImSchV:2013. The measured ranges have been selected so as to ensure as broad a field of application as possible.

The suitability of the AMS for this application was assessed on the basis of a laboratory test and a three month field test at a waste incineration.

The AMS is approved for an ambient temperature range of +5 °C to +45 °C.

The notification of suitability of the AMS, performance testing and the uncertainty calculation have been effected on the basis of the regulations applicable at the time of testing. As changes in legal provisions are possible, any potential user should ensure that this AMS is suitable for monitoring the emission limit values relevant to the application.

Any potential user should ensure, in consultation with the manufacturer, that this AMS is suitable for the installation at which it will be installed.

Note

The legal regulations mentioned correspond to the current state of legislation during certification. Each user should, if necessary, in consultation with the competent authority, ensure that this AMS meets the legal requirements for the intended use. In addition, it cannot be ruled out that legal regulations governing the use of a measuring device for emission monitoring may change during the lifetime of the certificate.

Basis of the certification

This certification is based on:

- Test report 936/21228173/A dated 21 October 2015 of TÜV Rheinland Energie und Umwelt GmbH
- Suitability announced by the German Federal Environment Agency (UBA) as the relevant body
- The ongoing surveillance of the product and the manufacturing process

Publication in the German Federal Gazette: BAnz AT 14.03.2016 B7, chapter I No. 3.1,
Announcement by UBA dated 18 February 2016:

AMS designation:

AO2000-Fidas24 for TOC

Manufacturer:

ABB Automation GmbH, Frankfurt am Main

Field of application:

For measurements at plants requiring official approval and plants in accordance with the 27th BImSchV

Measuring ranges during the performance test:

Component	Certification range	Supplementary measurement ranges			Unit
TOC	0 - 15	0 - 50	0 - 150	0 - 500	mg/m ³

Software version:

Fidas24: 3.4.2

Syscon: 5.1.4

Restrictions:

none

Notes:

1. The maintenance interval is four weeks.
2. It is possible to use the analyser in its versions AO2020 (19" housing for rack mounting) and AO2040 (housing for wall mounting).

Test Institute: TÜV Rheinland Energie und Umwelt GmbH, Cologne

Report No.: 936/21228173/A dated 21 October 2015

Publication in the German Federal Gazette: BAnz AT 01.08.2016 B11, Chap. V notification 3,
Announcement by UBA dated 14 July 2016:

3 Notification as regards Federal Environment Agency (UBA) notice of 18 February 2016 (BAnz AT 14.03.2016 B7, chapter I number 3.1)

The current software versions of the AO2000-Fidas24 measuring TOC manufactured by ABB Automation GmbH are:

Fidas24 (AMC-Board): 3.4.4
Syscon: 5.1.4

Statement issued by TÜV Rheinland Energie und Umwelt GmbH dated 29-02-2016

Publication in the German Federal Gazette: BAnz AT 15.03.2017 B6, Chap. V notification 18,
Announcement by UBA dated 22 February 2017:

18 Notification as regards Federal Environment Agency (UBA) notices of 18 February 2016 (BAnz AT 14.03.2016 B7, chapter I number 3.1) and of 14 July 2016 (BAnz AT 01.08.2016 B11, chapter V notification 3)

The latest software versions of the AO2000-Fidas24 measuring system for TOC are:

Fidas24 (AMC-Board): 3.4.4
Syscon: 5.1.8
Software versions 5.1.6 and 5.1.7 may also be used.

Statement issued by TÜV Rheinland Energy GmbH dated 6 October 2016

Publication in the German Federal Gazette: BAnz AT 31.07.2017 B12, Chap. II notification 3,
Announcement by UBA dated 13 July 2017:

3 Notification as regards Federal Environment Agency (UBA) notices of 18 February 2016 (BAnz AT 14.03.2016 B7, chapter I number 3.1) and of 22 July 2017 (BAnz AT 15.03.2017 B6, chapter V notification 18)

The current software versions of the AO2000-Fidas24 measuring TOC manufactured by ABB Automation GmbH are:

Fidas24 (AMC-Board): 3.8.0
Syscon: 5.1.12

The Impotron power supply type PSU-0261-12-14 may be used as 24V power supply.

Statement issued by TÜV Rheinland Energy GmbH dated 7 March 2017

Publication in the German Federal Gazette: BAnz AT 26.03.2018 B8, Chap. V notification 17,
Announcement by UBA dated 21 February 2018:

17 Notification as regards Federal Environment Agency (UBA) notices of 18 February 2016 (BAnz AT 14.03.2016 B7, chapter I number 3.1) and of 13 July 2017 (BAnz AT 31.07.2017 B12, chapter II notification 3)

The current software versions of the AO2000-Fidas24 measuring TOC manufactured by ABB Automation GmbH are:

Fidas24 (AMC-Board): 3.8.2
Syscon: 5.1.14

The U-remote modules manufactured by Weidmüller or the KL series manufactured by Beckhoff may be used as analogue output signals.

Statement issued by TÜV Rheinland Energy GmbH dated 7 December 2017

Publication in the German Federal Gazette: BAnz AT 17.07.2018 B9, Chap. III notification 5,
Announcement by UBA dated 3 July 2018:

5 Notification as regards Federal Environment Agency (UBA) notices of 18 February 2016 (BAnz AT 14.03.2016 B7, chapter I number 3.1) and of 21 February 2018 (BAnz AT 26.03.2018 B8, chapter V notification 17)

The latest software versions of the AO2000-Fidas24 measuring TOC manufactured by ABB Automation GmbH are:

Fidas24 (AMC-Board): 3.8.6
Syscon: 5.1.16

This includes software version 3.8.4. for the AMC-board.

Statement issued by TÜV Rheinland Energy GmbH dated 2 May 2018

Publication in the German Federal Gazette: BAnz AT 24.03.2020 B7, Chap. IV notification 1,
Announcement by UBA dated 24 February 2020:

1 Notification as regards Federal Environment Agency notices of 18 February 2016 (BAnz AT 14.03.2016 B7, chapter I number 3.1) and of 3 July 2018 (BAnz AT 17.07.2018 B9, chapter III notification 5)

The latest software versions of the AO2000-Fidas24 measuring system for TOC manufactured by ABB Automation GmbH are:

Fidas24 (AMC-Board): 3.9.0
Syscon: 5.1.18

Statement issued by TÜV Rheinland Energy GmbH dated 13 September 2019

Publication in the German Federal Gazette: BAnz AT 05.08.2021 B5, Chap. IV notification 19, Announcement by UBA dated 29 June 2021:

19 Notification as regards Federal Environment Agency (UBA) notices of 18 February 2016 (BAnz AT 14.03.2016 B7, chapter I number 3.1) and of 24 February 2020 (BAnz AT 24.03.2020 B7, chapter IV notification 1)

The latest software versions of the measuring system AO2000-Fidas24 for TOC manufactured by ABB Automation GmbH are:

Fidas24 (AMC-Board): 3.9.2
Syscon: 5.1.20

Statement issued by TÜV Rheinland Energy GmbH dated 19 February 2021

Publication in the German Federal Gazette: BAnz AT 11.04.2022 B10, Chap. VI notification 22, Announcement by UBA dated 9 March 2022:

22 Notification as regards Federal Environment Agency (UBA) notices of 18 February 2016 (BAnz AT 14.03.2016 B7, chapter I number 3.1) and of 29 June 2021 (BAnz AT 05.08.2021 B5, chapter IV notification 19)

The current software versions of the measuring device AO2000-Fidas24 for TOC of the company ABB Automation GmbH are:

Fidas24 (AMC-Board): 3.9.4
Syscon: 5.1.22

The company ABB Automation GmbH merged with its parent company ABB AG on September 13. The new name of the manufacturer is ABB AG.

Statement issued by TÜV Rheinland Energy GmbH dated 14 September 2021

Publication in the German Federal Gazette: BAnz AT 20.03.2023 B6, Chap. IV notification 2, Announcement by UBA dated 21 February 2023:

2 Notification as regards Federal Environment Agency (UBA) notices of 18 February 2016 (BAnz AT 14.03.2016 B7, chapter I number 3.1) and of 9 March 2022 (BAnz AT 11.04.2022 B10, chapter VI notification 22)

The current software versions of the AO2000-Fidas24 measuring system for TOC from ABB AG are:

Fidas24 (AMC board): 3.9.8, Syscon: 5.1.22

The software version 3.9.6 for the AMC board is included.

The measuring system may also be operated with hardware version 08 of the printed circuit board.

Statement issued by TÜV Rheinland Energy GmbH dated 05 September 2022

Publication in the German Federal Gazette: BAnz AT 31.10.2024 B9, Chap. IV notification 5, Announcement by UBA dated 21 August 2024:

5 Notification as regards Federal Environment Agency (UBA) notices of 18 February 2016 (BAnz AT 14.03.2016 B7, chapter I number 3.1) and of 21 February 2023 (BAnz AT 20.03.2023 B6, chapter IV notification 2)

The current software versions for the AO2000-Fidas24 measuring system for TOC from the company ABB AG are:

Fidas24 (AMC board): 3.9.10
Syscon: 5.1.24

Statement issued by TÜV Rheinland Energy GmbH dated 20 February 2024

Publication in the German Federal Gazette: BAnz AT 19.05.2025 B3, Chap. IV notification 29, Announcement by UBA dated 2 April 2025:

29 Notification as regards Federal Environment Agency (UBA) notices of 18 February 2016 (BAnz AT 14.03.2016 B7, chapter I number 3.1) and of 21 August 2024 (BAnz AT 31.10.2024 B9, chapter IV notification 5)

The current software versions of the AO2000-Fidas24 measuring system for TOC of ABB AG are as follows:

Fidas24 (AMC board): 3.9.12
Syscon: 5.1.24

Statement issued by TÜV Rheinland Energy & Environment GmbH dated 20 December 2024

Certified product

This certificate applies to automated measurement systems conforming to the following description:

The AMS is a flame ionization detector (FID) designed to determine total organic carbon (TOC) in exhaust gases.

The extractive AO2000-Fidas24 analyser consists of the following components:

- Analyser AO2000-Fidas24
- heated sensor (180°C) incl. controller, ABB PFE 3
- heated measurement gas line (180°C), (max. 60 m) incl. controller, Teflon sample gas line

For operation, combustion air with a TOC concentration of < 1% of the measurement range is required.

The Fidas24 analyser is an analyser module integrated in a universal housing type AO2000 which is part of the AO2000 series. This housing accommodates the display and control unit, the evaluation unit, the analyser module and the power supply unit. Analogue outputs and data interfaces are also located here.

It is possible to use the analyser in its versions AO2020 (19" housing for rack mounting) and AO2040 (housing for wall mounting).

General notes

This certificate is based upon the equipment tested. The manufacturer is responsible for ensuring that on-going production complies with the requirements of the EN 15267. The manufacturer is required to maintain an approved quality management system controlling the manufacture of the certified product. Both the product and the quality management systems shall be subject to regular surveillance.

If a product of the current production does not conform to the certified product, TÜV Rheinland Energy & Environment GmbH must be notified at the address given on page 1.

A certification mark with an ID-Number that is specific to the certified product is presented on page 1 of this certificate. This certification mark may be applied to the product or used in advertising materials for the certified product.

This document as well as the certification mark remains property of TÜV Rheinland Energy & Environment GmbH. With revocation of the publication the certificate loses its validity. After the expiration of the certificate and on requests of the TÜV Rheinland Energy & Environment GmbH this document shall be returned and the certificate mark must not be employed anymore.

The relevant version of this certificate and its expiration is also accessible on the internet: [**gal1.de**](http://gal1.de).

History of documents

Certification of AO2000-Fidas24 is based on the documents listed below and the regular, continuous monitoring of the Quality Management System of the manufacturer:

Initial certification according to EN 15267

Certificate No. 0000050625_00: 25 April 2016
Expiry date of the certificate: 13 March 2021
Test report: 936/21228173/A dated 21 October 2015
TÜV Rheinland Energie und Umwelt GmbH
Publication: BAnz AT 14.03.2016 B7, chapter I number 3.1
UBA announcement dated 18 February 2016

Notifications in accordance with EN 15267

Statement issued by TÜV Rheinland Energie und Umwelt GmbH dated 28 February 2016
Publication: BAnz AT 01.08.2016 B11, chapter V notification 3
UBA announcement dated 14 July 2016
(Software changes)

Statement issued by TÜV Rheinland Energy GmbH dated 6 October 2016
Publication: BAnz AT 15.03.2017 B6, chapter V notification 18
UBA announcement dated 22 February 2017
(Software changes)

Statement issued by TÜV Rheinland Energy GmbH dated 7 March 2017
Publication: BAnz AT 31.07.2017 B12, chapter II notification 3
UBA announcement dated 13 July 2017
(Soft- and hardware changes)

Statement issued by TÜV Rheinland Energy GmbH dated 7 December 2017
Publication: BAnz AT 26.03.2018 B8, chapter V notification 17
UBA announcement dated 21 February 2018
(Soft- and hardware changes)

Statement issued by TÜV Rheinland Energy GmbH dated 2 May 2018
Publication: BAnz AT 17.07.2018 B9, chapter III notification 5
UBA announcement dated 3 July 2018
(Software changes)

Statement issued by TÜV Rheinland Energy GmbH dated 13 September 2019
Publication: BAnz AT 24.03.2020 B7, chapter IV notification 1
UBA announcement dated 24 February 2020
(Software changes)

Renewal of certificates

Certificate No. 0000050625_01: 13 March 2021
Expiry date of the certificate: 13 March 2026

Notifications

Statement issued by TÜV Rheinland Energy GmbH dated 19 February 2021
Publication: BAnz AT 05.08.2021 B5, chapter IV notification 19
UBA announcement dated 29 June 2021
(Software change Softwareänderung)

Statement issued by TÜV Rheinland Energy GmbH dated 14 September 2021
Publication: BAnz AT 11.04.2022 B10, chapter VI notification 22
UBA announcement dated 9 March 2022
(Software changes and new producer name formerly ABB Automation GmbH)

Statement issued by TÜV Rheinland Energy GmbH dated 5 September 2022
Publication: BAnz AT 20.03.2023 B6, chapter IV notification 2
UBA announcement dated 21 February 2023
(Soft- and hardware changes)

Statement issued by TÜV Rheinland Energy & Environment GmbH dated 20 February 2024
Publication: BAnz AT 31.10.2024 B9, chapter IV notification 5
UBA announcement dated 21 August 2024
(Software changes)

Statement issued by TÜV Rheinland Energy & Environment GmbH dated 20 December 2024
Publication: BAnz AT 19.05.2025 B3, chapter IV notification 29
UBA announcement dated 2 April 2025
(Software changes)

Renewal of certificates

Certificate No. 0000050625_02: 12 March 2026
Expiry date of the certificate: 13 March 2031

Calculation of overall uncertainty according to EN 14181 and EN 15267-3

Measuring system

Manufacturer	ABB Automation GmbH
AMS designation	AO2000-Fidas24
Serial number of units under test	33481024 / 33481014
Measuring principle	FID

Test report

Test laboratory	TÜV Rheinland
Date of report	2015-10-21

Measured component

Certification range	TOC	0 - 15 mg/m ³
---------------------	-----	--------------------------

Evaluation of the cross-sensitivity (CS)

(system with largest CS)

Sum of positive CS at zero point	0.55 mg/m ³
Sum of negative CS at zero point	0.00 mg/m ³
Sum of positive CS at span point	0.32 mg/m ³
Sum of negative CS at span point	-0.49 mg/m ³
Maximum sum of cross-sensitivities	0.55 mg/m ³
Uncertainty of cross-sensitivity	u_i 0.320 mg/m ³

Calculation of the combined standard uncertainty

Tested parameter

			u^2
Standard deviation from paired measurements under field conditions *	u_D	0.048 mg/m ³	0.002 (mg/m ³) ²
Lack of fit	u_{lof}	0.041 mg/m ³	0.002 (mg/m ³) ²
Zero drift from field test	$u_{d,z}$	-0.225 mg/m ³	0.051 (mg/m ³) ²
Span drift from field test	$u_{d,s}$	0.260 mg/m ³	0.068 (mg/m ³) ²
Influence of ambient temperature at span	u_t	0.058 mg/m ³	0.003 (mg/m ³) ²
Influence of supply voltage	u_v	0.040 mg/m ³	0.002 (mg/m ³) ²
Cross-sensitivity (interference)	u_i	0.320 mg/m ³	0.102 (mg/m ³) ²
Influence of sample gas flow	u_p	0.118 mg/m ³	0.014 (mg/m ³) ²
Uncertainty of reference material at 70% of certification range	u_{rm}	0.121 mg/m ³	0.015 (mg/m ³) ²
Variation of response factors (TOC)	u_{rf}	0.042 mg/m ³	0.002 (mg/m ³) ²

* The larger value is used :

"Repeatability standard deviation at set point" or

"Standard deviation from paired measurements under field conditions"

Combined standard uncertainty (u_c)

$$u_c = \sqrt{\sum (u_{max,j})^2} \quad 0.51 \text{ mg/m}^3$$

Total expanded uncertainty

$$U = u_c * k = u_c * 1.96 \quad 1.00 \text{ mg/m}^3$$

Relative total expanded uncertainty

U in % of the ELV 10 mg/m³ 10.0

Requirement of 2010/75/EU

U in % of the ELV 10 mg/m³ 30.0

Requirement of EN 15267-3

U in % of the ELV 10 mg/m³ 22.5