

CERTIFICATE

of Product Conformity (QAL1)

Certificate No: 0000028749_03

Certified AMS: D-R 290 for dust

Manufacturer: DURAG GmbH
Kollaustr. 105
22453 Hamburg
Germany

Test Institute: TÜV Rheinland Energy & Environment GmbH

**This is to certify that the AMS has been tested
and found to comply with the standards
EN 15267-1 (2009), EN 15267-2 (2023), EN 15267-3 (2007),
as well as EN 14181 (2014).**

Certification is awarded in respect of the conditions stated in this certificate
(this certificate contains 12 pages).

The present certificate replaces certificate 0000028749_02 dated 25 January 2021.




Publication in the German Federal Gazette
(BAnz) of 26 January 2011


This certificate will expire on:
25 January 2031

German Environment Agency

TÜV Rheinland
Energy & Environment GmbH
Cologne, 21 January 2026

Dessau, 22 January 2026


Dr. Marcel Langner
Head of Section II 4


ppa. Dr. Peter Wilbring

Test report:	936/21212470/B dated 1 October 2010
Initial certification:	26 January 2011
Expiry date:	25 January 2031
Certificate:	Renewal (of previous certificate 0000028749_02 of 25 January 2021 valid until 25 January 2026)
Publication:	BAz. 26 January 2011, No. 14, p. 294, chapter I No. 1.2

Approved application

The tested AMS is suitable for use at plants according to Directive 2010/75/EC, chapter III (combustion plants / 13th BImSchV:2009), chapter IV (waste incineration plants / 17th BImSchV:2009), Directive 2015/2193/EC (44th BImSchV:2019), TA Luft:2021, 30th BImSchV:2009 and 27th BImSchV:1997. The measured ranges have been selected so as to ensure as broad a field of application as possible.

The suitability of the AMS for this application was assessed on the basis of a laboratory test and a three month field test at a waste incineration.

The AMS is approved for an ambient temperature range „D-R 290 M of -20 °C to +50 °C and D-R 290 M EC2: -40 °C to +60 °C.

The notification of suitability of the AMS, performance testing and the uncertainty calculation have been effected on the basis of the regulations applicable at the time of testing. As changes in legal provisions are possible, any potential user should ensure that this AMS is suitable for monitoring the emission limit values relevant to the application.

Any potential user should ensure, in consultation with the manufacturer, that this AMS is suitable for the installation at which it will be installed.

Note

The legal regulations mentioned correspond to the current state of legislation during certification. Each user should, if necessary, in consultation with the competent authority, ensure that this AMS meets the legal requirements for the intended use. In addition, it cannot be ruled out that legal regulations governing the use of a measuring device for emission monitoring may change during the lifetime of the certificate.

Basis of the certification

This certification is based on:

- Test report 936/21212470/B dated 1 October 2010 of TÜV Rheinland Energie und Umwelt GmbH
- Suitability announced by the German Federal Environment Agency (UBA) as the relevant body
- The ongoing surveillance of the product and the manufacturing process

Publication in the German Federal Gazette: BAnz. 26 January 2011, No. 14, p. 294, chapter I No. 1.2, Announcement by UBA dated 10 January 2011:

AMS designation:

D-R 290 for dust

Manufacturer:

DURAG GmbH, Hamburg

Field of application:

For plants requiring official approval and for plants according to the 27th BImSchV

Measuring ranges during the performance test:

Component	Certification range	supplementary ranges			Unit
Dust (optical transmission)	0 – 15				mg/m ³
		0 - 0.2	0 - 0.5	0 - 1.6	Ext.
		0 – 100%	.		Opac

0 – 0.1 Ext. equals 0 – 16 mg/m³ with an optical length of 5 m.

Software version:

3.21 (measurement head),
4.37 (evaluation unit)

Restrictions:

The measuring system may only be employed if the temperature does not fall below dew point.

Notes:

1. The dust concentration is determined in wet flue gas under operational conditions.
2. The maintenance interval is four weeks.
3. The measuring path length of 5 m and the measuring range of 16 mg/m³ determined during the calibration results in a product of 80 mg m/m³ for the field test plant.
4. Supplementary test as regards Federal Agency Notice of 22 April 2003 (BAnz. p. 10742, chapter I number 1.1) related to the applicability of standard EN 15267.
5. The requirement for the determination coefficient R² of the calibration function in accordance with EN 15267-3 was not satisfied.

Test Institute: TÜV Rheinland Energie und Umwelt GmbH, Cologne

Report No.: 936/21212470/B dated 1 October 2010

Publication in the German Federal Gazette: BAnz AT 02.04.2015 B5, Chap. IV notification 28, Announcement by UBA dated 25 February 2015:

28 Notification as regards Federal Environment Agency (UBA) notice of 10 January 2011 (BAnz. p. 294, chapter I number 1.2)

The D-R 290 measuring system manufactured by DURAG GmbH may also be equipped with the L3-W32 light source instead of the L3-W30 light source.

Statement issued by TÜV Rheinland Energie und Umwelt GmbH
dated 30 September 2014

Publication in the German Federal Gazette: BAnz AT 26.08.2015 B4, Chap. V notification 28, Announcement by UBA dated 22 July 2015:

28 Notification as regards Federal Environment Agency (UBA) notices of 10 February 2011 (BAnz. p. 294, chapter I number 1.2) and of 25 July 2015 (BAnz AT 02.04.2015 B5, chapter IV notification 28)

The D-R 290 measuring system for dust manufactured by DURAG GmbH has been equipped with a re-designed measuring head, which now carries the designation D-R 290 M EC2. With the new measuring head it is no longer necessary to use the D-R 290 AW evaluation unit to operate the AMS.

The following points apply when using AMS with the D-R 290 M EC2 measuring head:

- The AMS can be used with the D-ISC 100 evaluation unit or the D-TB 100 supply unit.
- The D-ISC 100 universal control unit has a digital Modbus RTU interface and a Modbus TCP in accordance with VDI 4201 parts 1 and 3 (EIA-485, serial and TCP/IP, Ethernet).
- The D-R 290 measuring system has a digital Modbus RTU interface in accordance with VDI 4201 parts 1 and 3.
- When using the D-R 290 measuring system with the D-ISC 100 universal control unit, the Modbus interface of the D-R 290 measuring system cannot be used. Instead, the Modbus digital interface of the D-ISC 100 universal control unit is used.
- When using the AMS without the D-ISC 100 evaluation unit, the AMS shall be operated by means of the D-ESI 100 software on a customary PC/notebook/tablet.
- The permissible ambient temperature range for the AMS is -40 °C – 60 °C.

Regardless of the changes, the use of the D-R 290 R reflector and a suitable purge air supply with the AMS is still mandatory.

The latest software versions of the D-R 290 measuring system for dust manufactured by DURAG GmbH are:

D-R 290: 05.00R0000

D-ISC 100: 01.03R0000

D-ESI 100: 1.1.015

Statement and test report No. 936/21226948/A dated 26 March 2015
issued by TÜV Rheinland Energie und Umwelt GmbH

Publication in the German Federal Gazette: BAnz AT 26.03.2018 B8, Chap. V notification 23, Announcement by UBA dated 21 February 2018:

**23 Notification as regards Federal Environment Agency notices
of 10 February 2011 (BAnz. p. 294, chapter I number 1.2) and
of 22 July 2015 (BAnz AT 26.08.2015 B4, chapter V notification 28)**

The latest software versions of D-R 290 measuring system for dust manufactured by DURAG GmbH are:

D-R 290: 05.01.R004
D-ISC 100: 01.04R0007
D-ESI 100: 01.10R0007

The following intermediary versions have also been approved:

D-ISC 100: 01.04R0001; 01.04R0004; 01.04R0006
D-ESI 100: 1.1.016; 1.1.017; 1.2.003

In the D-ISC 100, the Phoenix Contact QUINT4-PS/1AC/24DC/10 power supply unit may be used instead of the XPPower DNR240PS24-I power supply unit used up to now.

Statement issued by TÜV Rheinland Energy GmbH dated 8 December 2017

Publication in the German Federal Gazette: BAnz AT 26.03.2019 B7, Chap. IV notification 3, Announcement by UBA dated 27 February 2019:

3 Notification as regards Federal Environment Agency (UBA) notices of 10 January 2011 (BAnz. p. 294, chapter I number 1.2) and of 21 February 2018 (BAnz AT 26.03.2018 B8, chapter V notification 23)

The latest software versions of D-R 290 measuring system for dust manufactured by DURAG GmbH are:

D-R 290: 05.10.R004
D-ISC 100: 02.02R0066
D-ESI 100: 01.10R0007

Thus, the following software versions have also been approved:
D-ISC 100: 02.00R0048, 02.02R0020

The measuring system (D-R 290 measuring head MEC2) may be equipped with a revised version of the D-ISC 100 control unit. It is available in the following model versions:

- D-ISC 100 M (standard)
- D-ISC 100 C (compact housing)
- D-ISC 100 P (c/w purge air blower)
- D-ISC 100 R (housing for 19“ rack mounting)

The D-ISC 100 control unit also provides a digital Modbus interface which complies with VDI standard 4201, parts 1 and 3.

Report No. 936/21242380/A dated 14 September 2018 prepared by TÜV Rheinland Energy GmbH presents the test results for the revised D-ISC 100 control unit.

An alternative supplier was qualified for the retroreflector or triple mirror. The technical specifications of the components remain unchanged.

An alternative manufacturer qualified for the PK243-03A-C22 step engine, art. No. 1107147. The SECM243-S0.3A engine manufactured by EC Motors may also be used in the future.

Statement issued by TÜV Rheinland Energy GmbH dated 14 January 2019

Publication in the German Federal Gazette: BAnz AT 24.03.2020 B7, Chap. IV notification 13, Announcement by UBA dated 24 February 2020:

13 Notification as regards Federal Environment Agency (UBA) notices of 10 January 2011 (BAnz. p. 294, chapter I number 1.2) and of 27 February 2019 (BAnz AT 26.03.2019 B7, chapter IV notification 3)

The latest software versions of D-R 290 measuring system for dust manufactured by DURAG GmbH are:

D-R 290: 05.10.R004
D-ISC 100: 02.02R0066
D-ESI 100: 01.11R0018

D-ESI 100 software version 01.11R0017 may also be used.

Statement issued by TÜV Rheinland Energy GmbH dated 1 October 2019

Publication in the German Federal Gazette: BAnz AT 28.07.2022 B4, Chap. III notification 4, Announcement by UBA dated 28 June 2022:

4 Notification as regards Federal Environment Agency (UBA) notices of 10 January 2011 (BAnz. p. 294, chapter I number 1.2) and of 24 February 2020 (BAnz AT 24.03.2020 B7, chapter IV notification 13)

The current software versions of the measuring device D-R 290 for dust of the company DURAG GmbH are:

D-R 290: 05.10.R004
D-ISC 100: 02.02R0073
D-ESI 100: 01.11R0018

The D-ESI 100 software version 01.11R0017 can also be used.

Statement issued by TÜV Rheinland Energy GmbH dated 13 April 2022

Publication in the German Federal Gazette: BAnz AT 31.10.2025 B5, Kap. IV Mitteilung 5,
Announcement by UBA dated 31 October 2025

**4 Notification as regards Federal Environment Agency (UBA) notices
of 10 January 2011 (BAnz. p. 294, chapter I number 1.2) and
of 28 June 2022 (BAnz AT 28.07.2022 B4, chapter III notification 4)**

The current software versions of the D-R 290 measuring system for dust from
DURAG GmbH are as follows:

D-R290 M:	3.21
D-R 290 (Mxx2):	05.10.R004
D-ISC 100:	02.02R0073
D-ESI 100:	01.11R0025

Statement issued by TÜV Rheinland Energy & Environment GmbH dated
20 Februaray 2025

Certified product

This certificate applies to automated measurement systems conforming to the following description:

The D-R 290 measuring system uses the two-beam alternate light method, based on the autocollimation principle. The measuring light crosses the measuring path twice. The attenuation of the measuring light beam caused by the dust concentration is measured.

An optical recorder receives the measuring and comparison light beams alternately. The changeover between measuring light and comparison light is performed using a step motor every 2 min for 2 s. There is a common amplifier for signal processing of measuring and comparison light, temperature influences and long-term drift effects of the amplifier are compensated. The measuring light beam is generated by a Super Wide Band Diode (SWBD) without any influence of d.c. light (daylight). With the Wide Band performance of the SWBD the measuring result is independent against temperature and other influences and provides a very stable measuring.

The measurement system D-R 290 has two analogue outputs. Each of these outputs has two freely selectable extinction and opacity measuring ranges, which are external changeable. The ranges are freely adjustable from 0.1 to 1.6 Extinction and from 20 to 100% Opacity.

To check proper functioning of the D-R 290, a control cycle is performed at adjustable periodic intervals. In this cycle, the contamination of the optical interfaces, the span and the zero point are automatically measured and displayed. The results of the following measurements are corrected by the magnitude of the measured difference (contamination). If the contamination exceeds 6% a status signal is produced. By heating the optical discs, condensation and contamination are reduced as far as possible.

The D-R 290 measuring system with measuring head D-R 290 M (see type plate) consists of:

- D-R 290 M measuring head
- the D-R 290 R reflector
- D-R 290 AW evaluation unit
- Welding flanges and
- Blower purge

The D-R 290 measuring system with measuring head type D-R 290 M EC2 consists of:

- D-R 290 M EC2 measuring head
- the D-R 290 R reflector
- a suitable purge air supply
- D-ISC 100 evaluation unit with digital interface Modbus RTU and Modbus TCP
- according to VDI 4201 page 1 and 3 (EIA-485, serial and TCP/IP, Ethernet)
- or
- D-TB 100 supply unit with digital interface Modbus RTU
according to VDI 4201 page 1 and 3 (EIA-485, serial)

When combining the D-R 290 measuring system (equipped with the D-R 290 M EC2 measuring head) with the D-ISC 100 universal control unit, the digital Modbus interface on the D-ISC 100 universal control unit must be used. When using the AMS without the D-ISC 100 evaluation unit, the AMS shall be operated by means of the D-ESI 100 software on a customary PC/notebook/tablet.

General notes

This certificate is based upon the equipment tested. The manufacturer is responsible for ensuring that on-going production complies with the requirements of the EN 15267. The manufacturer is required to maintain an approved quality management system controlling the manufacture of the certified product. Both the product and the quality management systems shall be subject to regular surveillance.

If a product of the current production does not conform to the certified product, TÜV Rheinland Energy & Environment GmbH must be notified at the address given on page 1.

A certification mark with an ID-Number that is specific to the certified product is presented on page 1 of this certificate. This certification mark may be applied to the product or used in advertising materials for the certified product.

This document as well as the certification mark remains property of TÜV Rheinland Energy & Environment GmbH. With revocation of the publication the certificate loses its validity. After the expiration of the certificate and on requests of the TÜV Rheinland Energy & Environment GmbH this document shall be returned and the certificate mark must not be employed anymore.

The relevant version of this certificate and its expiration is also accessible on the internet:
qal1.de

History of documents

Certification of D-R 290 is based on the documents listed below and the regular, continuous monitoring of the Quality Management System of the manufacturer:

Basic test

Test report: 936/801017/A dated 31 January 2003

TÜV Immissionsschutz und Energiesysteme GmbH

Publication: BAnz. 15 May 2003, No. 90, p. 10742, chapter I number 1.1

UBA announcement dated 22 April 2003

Notifications

Statement issued by TÜV Immissionsschutz und Energiesysteme GmbH dated 30 June 2006

Publication: BAnz. 14 October 2006, No. 194, p. 6715, chapter V notification 1

UBA announcement dated 12 September 2006

(Extension of the scope to include the 27th BlmSchV)

Statement of TÜV Immissionsschutz und Energiesysteme GmbH dated 22 October 2009

Publication: BAnz. 12 February 2010, No. 24, p. 553, chapter IV notification 12

UBA announcement dated 25 January 2010

(Software changes)

Initial certification according to EN 15267

Certificate No. 0000028749_00: 9 February 2011

Expiry date of the certificate: 25 January 2016

Test report: 936/21212470/B dated 1 October 2010

TÜV Rheinland Energie und Umwelt GmbH

Publication: BAnz. 26 January 2011, No. 14, p. 294, chapter I number 1.2

UBA announcement dated 10 January 2011

Notifications

Statement issued by TÜV Rheinland Energie und Umwelt GmbH dated 30 September 2014

Publication: BAnz AT 02.04.2015 B5, chapter IV notification 28

UBA announcement dated 25 February 2015

(Hardware changes)

Statement issued by TÜV Rheinland Energie und Umwelt GmbH dated 26 March 2015

Test report: 936/21226948/A dated 26 March 2015

Publication: BAnz AT 26.08.2015 B4, chapter V notification 28

UBA announcement dated 22 July 2015

(Soft- and hardware changes)

Renewal of certificates

Certificate No. 0000028749_01: 21 January 2016

Expiry date of the certificate: 25 January 2021

Notifications

Statement issued by TÜV Rheinland Energy GmbH dated 8 December 2017

Publication: BAnz AT 26.03.2018 B8, chapter V notification 23

UBA announcement dated 21 February 2018

(Soft- and hardware changes)

Statement issued by TÜV Rheinland Energy GmbH dated 14 January 2019

Test report: 936/21242380/A dated 14 September 2018

Publication: BAnz AT 26.03.2019 B7, chapter IV notification 3

UBA announcement dated 27 February 2019

(Hard and software changes)

Statement issued by TÜV Rheinland Energy GmbH dated 1 October 2019

Publication: BAnz AT 24.03.2020 B7, chapter IV notification 13

UBA announcement dated 24 February 2020

(Software changes)

Renewal of certificates

Certificate No. 0000028749_02: 25 January 2021

Expiry date of the certificate: 25 January 2026

Notifications

Statement issued by TÜV Rheinland Energy GmbH dated 13 April 2022

Publication: BAnz AT 28.07.2022 B4, chapter III notification 4

UBA announcement dated 28 June 2022

(Software changes)

Statement issued by TÜV Rheinland Energy & Environment GmbH dated 31 October 2025

Publication: BAnz AT 31.10.2025 B5, chapter IV notification 5

UBA announcement dated 27 August 2025

(Software changes)

Renewal of certificates

Certificate No. 0000028749_03: 26 January 2026
Expiry date of the certificate: 25 January 2031

Calculation of overall uncertainty according to EN 14181 and EN 15267-3

Measuring system

Manufacturer	DURAG GmbH		
Name of measuring system	D-R 290		
Serial number of the candidates	406752 (142) / 406753 (158) / 1214444 / 1214434		
Measuring principle	optische Transmission		

Test report

Test laboratory	936 / 21212470/B		
Date of report	TÜV Rheinland		
	2010-10-01		

Measured component

Certification range	Dust	0 - 15 mg/m³
---------------------	------	--------------

Calculation of the combined standard uncertainty

Tested parameter

	<i>u</i>	<i>u</i> ²
Standard deviation from paired measurements under field conditions *	<i>u_D</i> 0.143 mg/m³	0.020 (mg/m³) ²
Lack of fit	<i>u_{lof}</i> 0.058 mg/m³	0.003 (mg/m³) ²
Zero drift from field test	<i>u_{d,z}</i> 0.012 mg/m³	0.000 (mg/m³) ²
Span drift from field test	<i>u_{d,s}</i> 0.017 mg/m³	0.000 (mg/m³) ²
Influence of ambient temperature at span	<i>u_t</i> 0.052 mg/m³	0.003 (mg/m³) ²
Influence of supply voltage	<i>u_v</i> 0.040 mg/m³	0.002 (mg/m³) ²
Uncertainty of reference material at 70% of certification range	<i>u_{rm}</i> 0.121 mg/m³	0.015 (mg/m³) ²
Excursion of measurement beam	<i>u_{mb}</i> 0.167 mg/m³	0.028 (mg/m³) ²

* The larger value is used :

"Repeatability standard deviation at span" or

"Standard deviation from paired measurements under field conditions"

Combined standard uncertainty (<i>u_c</i>)	$u_c = \sqrt{\sum (u_{max,j})^2}$	0.27 mg/m³
Total expanded uncertainty	$U = u_c * k = u_c * 1.96$	0.52 mg/m³

Relative total expanded uncertainty

Requirement of 2000/76/EC and 2001/80/EC

Requirement of EN 15267-3

U in % of the ELV 10 mg/m³ 5.2

U in % of the ELV 10 mg/m³ 30.0

U in % of the ELV 10 mg/m³ 22.5