

CERTIFICATE

of Product Conformity (QAL1)

Certificate No.: 0000040204_01

Certified AMS: Serinus 40 for NO, NO₂ and NO_x

Manufacturer: Ecotech Pty Ltd.
1492 Ferntree Gully Road
Knoxfield, VIC, 3180
Australia

Test Institute: TÜV Rheinland Energy GmbH

**This is to certify that the AMS has been tested
and found to comply with:**

**VDI 4202-1: 2010, VDI 4203-3: 2010, EN 14211: 2012,
EN 15267-1: 2009 and EN 15267-2: 2009**

Certification is awarded in respect of the conditions stated in this certificate
(see also the following pages).

The present certificate replaces certificate 0000040204 of 29 April 2014




Suitability Tested
Complying with
2008/50/EC
EN 15267
Regular
Surveillance
www.tuv.com
ID 0000040204


Publication in the German Federal Gazette
(BAz.) of 1 April 2014

This certificate will expire on:
30 June 2020

German Federal Environment Agency
Dessau, 1 April 2019

TÜV Rheinland Energy GmbH
Cologne, 31 March 2019


Dr. Marcel Langner
Head of Section II 4.1


ppa. Dr. Peter Wilbring

www.umwelt-tuv.de / www.eco-tuv.com
teu@umwelt-tuv.de
Tel. +49 221 806-5200

TÜV Rheinland Energy GmbH
Am Grauen Stein
51105 Cologne

Accreditation according to EN ISO/IEC 17025:2018 and certified according to ISO 9001:2015.

Certificate:
0000040204_01 / 1 April 2019

Test report: 936/21221977/A of 08 October 2013

Initial certification: 01 April 2014

Date of expiry: 30 June 2020

Publication: BAuz AT 01 April 2014 B12, chapter IV, No. 4.1

Approved application

The tested AMS is suitable for the continuous measurement of concentrations of nitrogen oxides in ambient air (stationary operation).

The suitability of the AMS for this application was assessed on the basis of a laboratory test and a three-month field test.

The AMS is approved for a temperature range of 0 °C to +30 °C.

Any potential user should ensure, in consultation with the manufacturer, that this AMS is suitable for ambient air applications at which it will be installed.

Basis of the certification

This certification is based on:

- test report 936/21221977/A of 08 October 2013 of TÜV Rheinland Energie und Umwelt GmbH
- suitability announced by the German Federal Environment Agency (UBA) as the relevant body
- the ongoing surveillance of the product and the manufacturing process
- publication in the German Federal Gazette: BAuz AT 01 April 2014 B12, chapter IV, No. 4.1
Announcement by UBA from 27 February 2014

AMS designation:

Serinus 40 for NO, NO₂ and NO_x

Manufacturer:

Ecotech Pty Ltd., Knoxfield, Australia

Field of application:

Continuous measurement of concentrations of nitrogen oxides in ambient air (stationary operation)

Measuring range during the performance test:

Component	Certification range	Unit
Nitrogen monoxide	0 - 1200	µg/m ³
Nitrogen dioxide	0 - 500	µg/m ³

Software version:

Firmware: 2.09.0005

Restrictions:

None

Notes:

1. The measuring system has to be operated in a lockable measuring cabinet or container.
2. The test report on the performance test is available online at www.qal1.de.

Test institute:

TÜV Rheinland Energie und Umwelt GmbH, Cologne

Report No.: 936/21221977/A of 8 October 2013

Certified product

This certificate applies to automated measurement systems conforming to the following description:

The Serinus 40 ambient air measuring system continuously monitors concentrations of nitrogen oxides by means of the chemiluminescence method. The instrument is designed for the continuous measuring of NO, NO₂ and NO_x in ambient air.

Nitric oxides are measured on the basis of chemiluminescence detection of gas phases. Sample air enters the reaction cell via two separate (alternating) paths; the NO and NO_x channels.

In the first path, NO reacts with ozone as follows:



In the second path, the gas first passes through the delay coil and then through the NO₂/NO converter so that it reaches the reaction cell after the gas in the first path. At that time NO_x (total concentration of NO and NO₂) is measured.

The NO₂ concentration is then calculated by subtracting the NO value from the measured NO_x value.

This reaction releases energy in the form of chemiluminescent radiation at a wavelength of 1100 nm, which is filtered by the optical band-pass filter and detected by the photomultiplier tube (PMT).

The detected level of chemiluminescence is directly proportional to the NO concentration in the sample.

The nitrogen oxides analyser consists of five main modules:

- pneumatics for channelling sample and exhaust gas (incl. valve manifolds)
- sensors for measuring nitrogen oxides (reaction cell module) and other relevant parameters
- control unit consisting of printed circuit boards for the control of sensors and pneumatics
- power supply for all processes in the analyser
- communication module for data access

Particulate filter

The particle filter is a 5 µm Teflon filter with a diameter of 47 mm. This filter removes all particles > 5 µm, which might affect the measurements.

Permeation dryer for drying the sample gas

After the particle filter there are two serially-connected permeation dryers in the sample gas line. During performance testing, these dryers were integrated into the analyser and therefore shall be considered as part of the tested measuring system. The dryers remove moisture from the sample gas and thereby reduce the amount of interference caused by moisture.

Sample gas pump

Manufacturer: Thomas, type: 617CD22-194 C

During performance testing the above-mentioned sample gas pump was used in the laboratory as well as in the field test. As far as the models Serinus 10 (ozone), Serinus 30 (CO) and Serinus 50 (SO₂) are concerned, one pump can be operated with up to two analysers. However, for the Serinus 40 (NO_x) one sample gas pump per analyser is required.

Certificate:

0000040204_01 / 1 April 2019

General notes

This certificate is based upon the equipment tested. The manufacturer is responsible for ensuring that on-going production complies with the requirements of the EN 15267. The manufacturer is required to maintain an approved quality management system controlling the manufacture of the certified product. Both the product and the quality management systems shall be subject to regular surveillance.

If a product of the current production does not conform to the certified product, TÜV Rheinland Energy GmbH must be notified at the address given on page 1.

A certification mark with an ID-Number that is specific to the certified product is presented on page 1 of this certificate. This can be applied to the product or used in publicity material for the certified product is presented on page 1 of this certificate.

This document as well as the certification mark remains property of TÜV Rheinland Energy GmbH. With revocation of the publication the certificate loses its validity. After the expiration of the certificate and on requests of the TÜV Rheinland Energy GmbH this document shall be returned and the certificate mark must not be employed anymore.

The relevant version of this certificate and the validity is also accessible on the internet: qal1.de.

Certification of Serinus 40 nitrogen oxide Analyzer is based on the documents listed below and the regular, continuous monitoring of the Quality Management System of the manufacturer:

Initial certification according to EN 15267:

Certificate No. 0000040204: 29 April 2014

Validity of the certificate until: 31 March 2019

Test report: 936/21221977/A of 08 October 2013

TÜV Rheinland Energie und Umwelt GmbH, Cologne

Publication: BAnz AT 01 April 2014 B12, chapter IV, No. 4.1

Announcement by UBA from 27 February 2014

Renewal of the certificate according to EN 15267:

Certificate No. 0000040204_01: 1 April 2019

Validity of the certificate until: 30 June 2020

Certificate:
 0000040204_01 / 1 April 2019

Expanded uncertainty based on the results of the laboratory testing of Device 1

Measured component	Ecotech Serinus 40	Serial-No.:	13-0095 (Device 1)
NO ₂		1h-limit value:	104.6 nmol/mol
No.	Performance characteristic	Performance criterion	Result
1	Repeatability standard deviation at zero	≤ 1.0 nmol/mol	0.310
2	Repeatability standard deviation at 1h-limit value	≤ 3.0 nmol/mol	0.270
3	"ack of fit" at 1h-limit value	≤ 4.0% of measured value	0.720
4	Sensitivity coefficient of sample gas pressure at 1h-limit value	≤ 8.0 nmol/mol/kPa	1.290
5	Sensitivity coefficient of sample gas temperature at 1h-limit value	≤ 3.0 nmol/mol/K	0.040
6	Sensitivity coefficient of surrounding temperature at 1h-limit value	≤ 3.0 nmol/mol/K	0.620
7	Sensitivity coefficient of electrical voltage at 1h-limit value	≤ 0.30 nmol/mol/V	0.045
8a	Interferent H ₂ O with 21 nmol/mol	≤ 10 nmol/mol (Zero)	0.90
8b	Interferent CO ₂ with 500 µmol/mol	≤ 10 nmol/mol (Span)	-0.300
8c	Interferent NH ₃ mit 200 nmol/mol	≤ 5.0 nmol/mol (Zero)	0.000
9	Averaging effect	≤ 7.0% of measured value	-1.350
18	Difference sample/calibration port	≤ 1.0%	-0.260
21	Converter efficiency	≥ 98	98.90
23	Uncertainty of test gas	≤ 3.0%	2.000
		Combined standard uncertainty	U _c
		Expanded uncertainty	U
		Relative expanded uncertainty	W
		Maximum allowed expanded uncertainty	W _{req}

Certificate:
0000040204_01 / 1 April 2019

Expanded uncertainty based on the results of the laboratory testing of Device 2

Measured component	Ecotech Serinus 40	Serial-No.:	13-0094 (Device 2)
NO ₂		1h-limit value:	104.6 nmol/mol
Performance characteristic			
No.	Performance criterion	Result	Partial uncertainty
1	Repeatability standard deviation at zero	≤ 1.0 nmol/mol	0.190 $U_{r,z}$ 0.04
2	Repeatability standard deviation at 1h-limit value	≤ 3.0 nmol/mol	0.190 $U_{r,h}$ 0.01
3	"lack of fit" at 1h-limit value	≤ 4.0% of measured value	0.840 $U_{l,h}$ 0.51
4	Sensitivity coefficient of sample gas pressure at 1h-limit value	≤ 8.0 nmol/mol/kPa	1.970 U_{sp} 5.37
5	Sensitivity coefficient of sample gas temperature at 1h-limit value	≤ 3.0 nmol/mol/K	0.450 U_{gt} 1.02
6	Sensitivity coefficient of surrounding temperature at 1h-limit value	≤ 3.0 nmol/mol/K	0.240 U_{st} 0.60
7	Sensitivity coefficient of electrical voltage at 1h-limit value	≤ 0.30 nmol/mol/V	0.006 U_V 0.02
8a	Interferent H ₂ O with 21 nmol/mol	≤ 10 nmol/mol (Zero)	0.000 U_{H2O} -0.26
8b	Interferent CO ₂ with 500 μmol/mol	≤ 10 nmol/mol (Span)	0.000 U_{CO2} 0.0700
8c	Interferent NH ₃ mit 200 nmol/mol	≤ 5.0 nmol/mol (Zero)	0.600 U_{NH3} 0.1434
9	Averaging effect	≤ 5.0 nmol/mol (Span)	0.430 or 0.38
18	Difference sample/calibration port	≤ 7.0% of measured value	-1.310 U_{av} -0.79
21	Converter efficiency	≤ 1.0%	0.230 U_{asc} 0.24
23	Uncertainty of test gas	≤ 3.0%	98.80 U_{ec} 1.26
Combined standard uncertainty			
		U_c	5.8345 nmol/mol
	Expanded uncertainty	U	11.6690 nmol/mol
	Relative expanded uncertainty	W	11.16 %
	Maximum allowed expanded uncertainty	W_{req}	15 %

Certificate:
0000040204_01 / 1 April 2019

Expanded uncertainty based on the results of the laboratory and field testing of Device 1

Measured component	Measuring device:	NO ₂	Serial-No.:	13-0095 (Device 1)
No.	Performance characteristic	Performance criterion	Result	Partial uncertainty
1	Repeatability standard deviation at zero	≤ 1.0 nmol/mol	0.310	U _{LZ} 0.06
2	Repeatability standard deviation at 1h-limit value	≤ 3.0 nmol/mol	0.270	U _{L,H} not considered, as $\sqrt{2} \cdot u_{r,h} = 0.01 < u_{r,f}$
3	"lack of fit" at 1h-limit value	≤ 4.0% of measured value	0.720	U _{L,H} 0.43
4	Sensitivity coefficient of sample gas pressure at 1h-limit value	≤ 8.0 nmol/mol/kPa	1.290	U _{SP} 3.56
5	Sensitivity coefficient of sample gas temperature at 1h-limit value	≤ 3.0 nmol/mol/K	0.040	U _{GT} 0.09
6	Sensitivity coefficient of surrounding temperature at 1h-limit value	≤ 3.0 nmol/mol/K	0.620	U _{ST} 1.55
7	Sensitivity coefficient of electrical voltage at 1h-limit value	≤ 0.30 nmol/mol/V	0.045	U _V 0.13
8a	Interferent H ₂ O with 21 nmol/mol	≤ 10 nmol/mol (Zero)	0.090	U _{H2O} 0.01
8b	Interferent CO ₂ with 500 nmol/mol	≤ 5.0 nmol/mol (Zero)	-0.470	U _{Int,Pos} or U _{Int,Neg} 0.09
8c	Interferent NH ₃ mit 200 nmol/mol	≤ 5.0 nmol/mol (Zero)	0.000	0.0086
9	Averaging effect	≤ 7.0% of measured value	-1.350	U _{av} -0.82
10	Reproducibility standard deviation under field conditions	≤ 5.0% of average over 3 months	3.550	U _f 3.71
11	Long term drift at zero level	≤ 5.0 nmol/mol	-0.510	U _{L,Z} -0.29
12	Long term drift at span level	≤ 5.0% of max. of certification range	2.510	U _{L,H} 1.52
18	Difference sample/calibration port	≤ 1.0%	-0.260	U _{SC} -0.27
21	Converter efficiency	≤ 98	98.900	U _{EC} 1.15
23	Uncertainty of test gas	≤ 3.0%	2.000	U _{Eq} 1.05
		Combined standard uncertainty	U _c	5.8861 nmol/mol
		Expanded uncertainty	U	11.7723 nmol/mol
		Relative expanded uncertainty	W	11.25%
		Maximum allowed expanded uncertainty	W _{req}	15%

Certificate:
0000040204_01 / 1 April 2019

Expanded uncertainty based on the results of the laboratory and field testing of Device 2

Measured component	Measuring device:	Ecotech Serinus 40	Serial-No.:	13-0094 (Device 2)	
		NO ₂	1h-limit value:	104,6 nmol/mol	
No.	Performance characteristic	Performance criterion	Result	Partial uncertainty	Square of partial uncertainty
1	Repeatability standard deviation at zero	≤ 1.0 nmol/mol	0.190	u _{r,z}	0.04
2	Repeatability standard deviation at 1h-limit value	≤ 3.0 nmol/mol	0.190	u _{r,h}	not considered, as $\sqrt{2} \cdot u_{r,h} = 0.01 < u_{r,f}$
3	"lack of fit" at 1h-limit value	≤ 4.0% of measured value	0.840	u _{i,h}	0.51
4	Sensitivity coefficient of sample gas pressure at 1h-limit value	≤ 8.0 nmol/mol/kPa	1.970	u _{gp}	5.37
5	Sensitivity coefficient of sample gas temperature at 1h-limit value	≤ 3.0 nmol/mol/K	0.450	u _{gt}	1.02
6	Sensitivity coefficient of surrounding temperature at 1h-limit value	≤ 3.0 nmol/mol/K	0.240	u _{st}	0.60
7	Sensitivity coefficient of electrical voltage at 1h-limit value	≤ 0.30 nmol/mol/V	0.006	u _v	0.02
8a	Interferent H ₂ O with 21 nmol/mol	≤ 10 nmol/mol (Zero)	-0.300	u _{H2O}	-0.26
8b	Interferent CO ₂ with 500 µmol/mol	≤ 5.0 nmol/mol (Span)	-0.570	u _{int, pos}	0.0700
8c	Interferent NH ₃ mit 200 nmol/mol	≤ 5.0 nmol/mol (Zero)	-0.240	or u _{int, neg}	0.1434
9	Averaging effect	≤ 7.0% of measured value	-1.310	u _{av}	-0.79
10	Reproducibility standard deviation under field conditions	≤ 5.0% of average over 3 months	3.550	u _{r,f}	3.71
11	Long term drift at zero level	≤ 5.0 nmol/mol	0.580	u _{d,z}	0.33
12	Long term drift at span level	≤ 5.0% of max. of certification range	2.550	u _{d,h}	1.54
18	Difference sample/calibration port	≤ 1.0%	0.230	u _{ssc}	0.24
21	Converter efficiency	≤ 98	98.800	u _{EC}	1.26
23	Uncertainty of test gas	≤ 3.0%	2.000	u _{sg}	1.05
		Combined standard uncertainty	u _c	7.0932 nmol/mol	
		Expanded uncertainty	U	14.1864 nmol/mol	
		Relative expanded uncertainty	W	13.56 %	
		Maximum allowed expanded uncertainty	W _{ref}	15 %	