

CONFIRMATION

Notification: 0000040204_02_05
on changes according to EN 15267 regarding certificate 0000040204_02 dated 1 July 2020

AMS designation: Serinus 40 for NO, NO₂ and NO_x

Manufacturer: ACOEM Australasia (Ecotech Pty Ltd)
1492 Ferntree Gully Road,
Knoxfield, VIC, 3180
Australia

German Federal Environmental Agency (UBA)

**Announcement regarding the uniform practice in
monitoring emissions and ambient air
dated 2 April 2025
Publication: BAnz AT 19.05.2025 B3**

**IV. Notifications as regards the uniform practice in monitoring
emissions and ambient air:**

**96 Notification as regards Federal Environment Agency (UBA) notices
of 27 February 2014 (BAnz AT 01.04.2014 B12, chapter IV number 4.1) and
of 19 March 2024 (BAnz AT 10.05.2024 B7, chapter V notification 45)**

The current software version of the Serinus 40 measuring system for NO, NO₂ and NO_x
from ACOEM Australasia is unchanged for devices with the microprocessor board
(C010001): V 2.35.0001


The current software version of the Serinus 40 measuring system for NO, NO₂ and NO_x
from ACOEM Australasia is for devices with the microprocessor board (C010014):
V 4.28.0000

The following software versions are also authorised for this device version:
V 4.23.0000, V 4.24.0000, V 4.25.0000, V 4.26.0000, V 4.27.0000

The software version number appears on the display of the measuring device in the
format 2.XX or 3.XX or 4.XX.

Statement issued by TÜV Rheinland Energy & Environment GmbH
dated 28 September 2024

TÜV Rheinland Energy & Environment GmbH
Cologne, 20 May 2025


i. V. Dipl.-Ing. Guido Baum


i. A. Dipl.-Ing. Carsten Röllig

www.umwelt-tuv.eu
TRE@umwelt-tuv.eu
Tel. +49 221 806-5200

TÜV Rheinland Energy & Environment GmbH
Am Grauen Stein
51105 Cologne

Test institute accredited to EN ISO/IEC 17025 by DAkkS (German Accreditation Body).
This accreditation is limited to the accreditation scope defined in the enclosure to the certificate D-PL-11120-02-00.