

# CERTIFICATE

## of Product Conformity (QAL1)

Certificate No.: 0000040205\_03

**Certified AMS:** Serinus 50 for SO<sub>2</sub>

**Manufacturer:** ACOEM Australasia (Ecotech Pty Ltd)  
1492 Ferntree Gully Road,  
Knoxfield, VIC, 3180  
Australia

**Test Institute:** TÜV Rheinland Energy & Environment GmbH

**This is to certify that the AMS has been tested  
and found to comply with the standards  
VDI 4202-1 (2018), EN 14212 (2012),  
as well as EN 15267-1 (2009) and EN 15267-2 (2023).**

Certification is awarded in respect of the conditions stated in this certificate  
(this certificate contains 16 pages).  
The present certificate replaces certificate 0000040205\_02 dated 1 July 2020.



Suitability Tested  
Complying with  
2008/50/EC  
EN 15267  
Regular  
Surveillance  
[www.tuv.com](http://www.tuv.com)  
ID 0000040205

Publication in the German Federal Gazette  
(BAnz) of 1 April 2014


German Environment Agency  
Dessau, 27 June 2025

This certificate will expire on:  
30 June 2030

TÜV Rheinland Energy &  
Environment GmbH  
Cologne, 30 June 2025



Dr. Marcel Langner  
Head of Section II 4



ppa. Dr. Peter Wilbring

[www.umwelt-tuv.eu](http://www.umwelt-tuv.eu)  
[qal1-info@tuv.com](mailto:qal1-info@tuv.com)  
Tel. + 49 221 806-5200

TÜV Rheinland Energy & Environment GmbH  
Am Grauen Stein  
51105 Köln

Test institute accredited to EN ISO/IEC 17025 by DAkkS (German Accreditation Body).  
This accreditation is limited to the accreditation scope defined in the enclosure to the certificate D-PL-11120-02-00.

**Test report:** 936/21221977/B dated 8 October 2013  
**Initial certification:** 1 April 2014  
**Expiry date:** 30 June 2030  
**Certificate:** Renewal (of previous certificate 0000040205\_02 of 1 July 2020 valid until 30 June 2025)  
**Publication:** BAnz AT 01.04.2014 B12, chapter IV No. 3.1

### Approved application

The tested AMS is suitable for continuous immission measurement of SO<sub>2</sub> in stationary use.

The suitability of the AMS for this application was assessed on the basis of a laboratory test and a three month field test.

The AMS is approved for an ambient temperature range of 0 °C to 30 °C.

The notification of suitability of the AMS, performance testing and the uncertainty calculation have been effected on the basis of the regulations applicable at the time of testing. As changes in legal provisions are possible, any potential user should ensure that this AMS is suitable for monitoring the measured values relevant to the application.

Any potential user should ensure, in consultation with the manufacturer, that this AMS is suitable for the intended use.

### Basis of the certification

This certification is based on:

- Test report 936/21221977/B dated 8 October 2013 of TÜV Rheinland Energie und Umwelt GmbH
- Suitability announced by the German Federal Environment Agency (UBA) as the relevant body
- The ongoing surveillance of the product and the manufacturing process

Publication in the German Federal Gazette: BAnz AT 01.04.2014 B12, chapter IV No. 3.1,  
Announcement by UBA dated 27 February 2014:

**AMS designation:**

Serinus 50 for SO<sub>2</sub>

**Manufacturer:**

Ecotech Pty Ltd., Knoxfield, Australia

**Field of application:**

Continuous measurement of sulphur dioxide concentration in ambient air (stationary operation)

**Measuring ranges during the performance test:**

Component	Certification range	Unit
Sulphur dioxide	0 – 1,000	µg/m <sup>3</sup>

**Software version:**

Firmware: 2.09.0005

**Restrictions:**

None

**Notes:**

1. The measuring system must be operated inside a lockable measuring cabinet or measurement container.
2. The test report on performance testing is available on the internet at [www.qal1.de](http://www.qal1.de).

**Test institute:** TÜV Rheinland Energie und Umwelt GmbH, Cologne

Report No.: 936/21221977/B dated 8 October 2013



Publication in the German Federal Gazette: BAnz AT 02.04.2015 B5, Chap. IV notification 7,  
Announcement by UBA dated 25 February 2015:

**7 Notification as regards Federal Environment Agency (UBA) notice  
of 27 February 2014 (BAnz AT 01.04.2014 B12, chapter IV number 3.1)**

Hereafter, the Serinus 50 measuring system for SO<sub>2</sub>, manufactured by Ecotech Pty Ltd., will be equipped with a new microprocessor board (C010014). This results in modifications of the power plug as well as software changes.

The current two software versions are designated as follows:

2.20.0009 for systems using the old microprocessor board (C010001)

3.10.001 for systems using the new microprocessor board (C010014).

Statement by TÜV Rheinland Energie und Umwelt GmbH of 12 September 2014

Publication in the German Federal Gazette: BAnz AT 15.03.2017 B6, Chap. V notification 8,  
Announcement by UBA dated 22 February 2017:

**8 Notification as regards Federal Environment Agency notices  
of 27 February 2014 (BAnz AT 01.04.2014 B12, chapter IV number 3.1) and  
of 25 February 2015 (BAnz AT 02.04.2015 B5 chapter IV notification 7)**

The current software version of the Serinus 50 for SO<sub>2</sub> manufactured by Ecotech Pty Ltd. for systems with micro processor board (C010001) is:  
V 2.31.0004.

The following software versions are approved for this instrument version:  
V 2.21.0000, V 2.22.0000, V 2.23.0000, V 2.24.0000, V 2.25.0004, V 2.26.0000,  
V 2.27.0000, V 2.28.0000, V 2.29.0003 and V 2.30.0000.

The current software version of the Serinus 50 for SO<sub>2</sub> manufactured by Ecotech Pty Ltd. for systems with micro processor board (C010014) is: V 3.48.011.

The following software versions are approved for this instrument version:  
V 3.13.000, V 3.14.001, V 3.15.010, V 3.16.001, V 3.18.003, V 3.20.000, V 3.22.000,  
V 3.23.015, V 3.24.000, V 3.26.000, V 3.27.000, V 3.28.000, V 3.29.013, V 3.30.005,  
V 3.31.002, V 3.32.003, V 3.33.004, V 3.34.000, V 3.35.004, V 3.36.000, V 3.37.004,  
V 3.38.006, V 3.39.000, V 3.40.001, V 3.41.004, V 3.42.000, V 3.43.000, V 3.44.004,  
V 3.45.011, V 3.46.002, V 3.47.006.

Statement by TÜV Rheinland Energy GmbH dated 13 October 2016

Publication in the German Federal Gazette: BAnz AT 26.03.2019 B7, Chap. IV  
notification 18, Announcement by UBA dated 27 February 2019:

**18 Notification as regards Federal Environment Agency notices  
of 27 February 2014 (BAnz AT 01.04.2014 B12, chapter IV number 3.1) and  
of 22 February 2017 (BAnz AT 15.03.2017 B6, chapter IV notification 8)**

The current software version of the Serinus 50 for SO<sub>2</sub> manufactured  
by Ecotech Pty Ltd. for systems with micro processor board (C010001) is:  
V 2.35.0001.

In addition, the following software versions have been approved for this instrument  
version: V 2.32.0000, V 2.33.0000, V 2.34.0000

The current software version of the Serinus 50 for SO<sub>2</sub> manufactured  
by Ecotech Pty Ltd. for systems with micro processor board (C010014) is:  
V 3.74.0003.

In addition, the following software versions have been approved for this instrument  
version:

V 3.49.0000, V 3.51.0011, V 3.52.0000, V 3.53.0012, V 3.54.0000, V 3.55.0000,  
V 3.56.0001, V 3.57.0002, V 3.58.0000, V 3.59.0004, V 3.60.0005, V 3.61.0000,  
V 3.62.0000, V 3.63.0001, V 3.64.0000, V 3.65.0001, V 3.66.0000, V 3.67.0003,  
V 3.68.0009, V 3.69.0001, V 3.70.0000, V 3.71.0000

The display of the measuring system shows the software version in the following  
format: 2.XX or 3.XX.

Statement by TÜV Rheinland Energy GmbH dated 10 October 2018

Publication in the German Federal Gazette: BAnz AT 24.03.2020 B7, Chap. IV  
notification 22, Announcement by UBA dated 24 February 2020:

**22 Notification as regards Federal Environment Agency (UBA) notices  
of 27 February 2014 (BAnz AT 01.04.2014 B12, chapter IV number 3.1) and  
of 27 February 2019 (BAnz AT 26.03.2019 B7, chapter IV notification 18)**

The latest software version of the Serinus 50 measuring system for SO<sub>2</sub>  
with microprocessor C010001 manufactured by Ecotech Pty Ltd. remains:  
V 2.35.0001.

The latest software version of the Serinus 50 measuring system for SO<sub>2</sub>  
with microprocessor C010014 manufactured by Ecotech Pty Ltd. is:  
V 3.87.0000.

Moreover, the following software version are approved for this instrument version:  
V 3.75.0003, V 3.76.0004, V 3.77.0009, V 3.78.0000, V 3.79.0001, V 3.81.0000,  
V 3.83.0000, V 3.84.0000, V 3.85.0001, V 3.86.0000.

The instrument's display shows the software version in the following format: 2.XX or  
3.XX.

Statement by TÜV Rheinland Energy GmbH dated 20 September 2019



Publication in the German Federal Gazette: BAnz AT 03.05.2021 B9, Chap. III  
notification 12, Announcement by UBA dated 31 March 2021:

**12 Notification as regards Federal Environment Agency (UBA) notices  
of 27 February 2014 (BAnz AT 01.04.2014 B12, chapter IV number 3.1) and  
of 24 February 2020 (BAnz AT 24.03.2020 B7, chapter IV notification 22)**

The latest software version of the Serinus 50 measuring system for SO<sub>2</sub> with  
microprocessor C010001 manufactured by Ecotech Pty Ltd. is:  
V 2.35.0001.

The latest software version of the Serinus 50 measuring system for SO<sub>2</sub> with  
microprocessor C010014 manufactured by Ecotech Pty Ltd. is:  
V 4.02.0000.

Furthermore, the following software versions are approved for this instrument  
version:

V 3.88.0000, V 3.89.0000, V 3.90.0002, V 4.00.0000, V 4.01.0000

The instrument's display shows the software version in the following format: 2.XX or  
3.XX or 4.XX.

The Serinus Main Controller Board (PCB) received an update from Rev. N to Rev. P.

Statement by TÜV Rheinland Energy GmbH dated 14 July 2020

Publication in the German Federal Gazette: BAnz AT 11.04.2022 B10, Chap. VI  
notification 6, Announcement by UBA dated 9 March 2022:

**6 Notification as regards Federal Environment Agency (UBA) notices  
of 27 February 2014 (BAnz AT 01.04.2014 B12, chapter IV number 3.1) and  
of 31 March 2021 (BAnz AT 03.05.2021 B9, chapter III notification 12)**

The company name of Ecotech Pty. Ltd. changes to ACOEM Australasia.

The current software version of the measuring device Serinus 50 for SO<sub>2</sub> of the  
company ACOEM Australasia for devices with the microprocessor board (C010001)  
is unchanged: V 2.35.0001.

The current software version of the measuring device Serinus 50 for SO<sub>2</sub> of the  
company ACOEM Australasia for devices with the microprocessor board (C010014)  
is: V 4.13.0000.

Furthermore, the following software versions are approved for this instrument  
version: V 4.04.0000, V 4.06.0000, V 4.07.0000, V 4.08.0000, V 4.09.0000,  
V 4.10.0000, V 4.11.0000.

The software version number appears in the display of the measuring device in the  
format 2.XX or 3.XX or 4.XX.

Statement by TÜV Rheinland Energy GmbH dated 20 August 2021

Publication in the German Federal Gazette: BAnz AT 20.03.2023 B6, Chap. IV  
notification 60, Announcement by UBA dated 21 February 2023:

**60 Notification as regards Federal Environment Agency (UBA) notices  
of 27 February 2014 (BAnz AT 01.04.2014 B12, chapter IV number 3.1) and  
of 9 March 2022 (BAnz AT 11.04.2022 B10, chapter VI notification 6)**

The current software version of the Serinus 50 measuring system for SO<sub>2</sub> from the  
company ACOEM Australasia for devices with the microprocessor board (C010001)  
remains: V 2.35.0001

The current software version of the Serinus 50 measuring system for SO<sub>2</sub> from the  
company ACOEM Australasia for devices with the microprocessor board (C010014)  
is: V 4.18.0000.

Furthermore, the following software versions are approved for this device version:  
V 4.14.0000, V 4.15.0000, V 4.16.0000, V 4.17.0000

The software version number appears in the display of the measuring system in the  
format 2.XX or 3.XX or 4.XX.

Statement by TÜV Rheinland Energy GmbH dated 5 September 2022

Publication in the German Federal Gazette: BAnz AT 10.05.2024 B7, Chap. V notification 46,  
Announcement by UBA dated 19 March 2024:

**46 Notification as regards Federal Environment Agency (UBA) notices  
of 27 February 2014 (BAnz AT 01.04.2014 B12, chapter IV number 3.1) and  
of 21 February 2023 (BAnz AT 20.03.2023 B6, chapter IV notification 60)**

The current software version for the Serinus 50 measuring system for SO<sub>2</sub> from  
ACOEM Australasia is unchanged for systems with the microprocessor board  
(C010001): V 2.35.0001

The current software version for the Serinus 50 measuring system for SO<sub>2</sub> from  
ACOEM Australasia for systems with the microprocessor board (C010014) is  
V 4.22.0000.

The following software versions are also authorised for this device version:  
V 4.19.0000, V 4.20.0000, V 4.21.0000

The software version number appears on the display of the measuring system in the  
format 2.XX or 3.XX or 4.XX.

Statement by TÜV Rheinland Energy GmbH dated 10 August 2023



Publication in the German Federal Gazette: BAnz AT 19.05.2025 B3, Chap. IV  
notification 97, Announcement by UBA dated 2 April 2025:

**97 Notification as regards Federal Environment Agency (UBA) notices of 27 February 2014 (BAnz AT 01.04.2014 B12, chapter IV number 3.1) and of 19 March 2024 (BAnz AT 10.05.2024 B7, chapter V notification 46)**

The current software version of the Serinus 50 measuring system for SO<sub>2</sub> from ACOEM Australasia is unchanged for devices with the microprocessor board (C010001):  
V 2.35.0001

The current software version of the Serinus 50 measuring system for SO<sub>2</sub> from ACOEM Australasia for devices with the microprocessor board is (C010014):  
V 4.28.0000

The following software versions are also authorised for this device version: V 4.23.0000, V 4.24.0000, V 4.25.0000, V 4.26.0000, V 4.27.0000

The software version number appears on the display of the measuring device in the format 2.XX or 3.XX or 4.XX.

Statement issued by TÜV Rheinland Energy & Environment GmbH dated 28 September 2024



**Certified product**

This certificate applies to automated measurement systems conforming to the following description:

The Serinus 50 measuring system is a continuous sulphur dioxide monitor which uses the method of ultraviolet fluorescence. The instrument is designed for the continuous measuring of sulphur dioxide in ambient air.

Measurements are performed by means of the following components:

- Hydrocarbon kicker
- UV lamp
- fluorescence cell
- optical band-pass filter
- photomultiplier tube (PMT)

The SO<sub>2</sub> concentration is automatically corrected for gas temperature and pressure changes and referenced to 0°C, 20°C or 25°C at 1 atmosphere. This allows the Serinus 50 to sample in the most useful range of SO<sub>2</sub> ambient measurement (25–500 ppb SO<sub>2</sub> in the air).

The measurement of sulphur dioxide is based on classical fluorescence spectroscopy principles. Sulphur dioxide (SO<sub>2</sub>) exhibits a strong ultraviolet (UV) absorption spectrum between 200 and 240 nm. When SO<sub>2</sub> absorbs UV from this wavelength, photon emission occurs (300–420 nm). The amount of fluorescence emitted is directly proportional to the SO<sub>2</sub> concentration.

The Serinus 50 follows these principles and measurement techniques:

- Sample air passes through a hydrocarbon kicker which removes hydrocarbons.
- UV energy from the zinc discharge lamp passes through a UV band-pass filter are used to produce radiation at 214 nm.
- The radiation is focused into the fluorescence cell where it is absorbed by the SO<sub>2</sub> molecules.
- The SO<sub>2</sub> molecules then emit photons (fluorescent light) uniformly in all directions.
- Wavelengths between 310–350 nm, which are specific to SO<sub>2</sub>, pass through a band pass filter where they reach the photomultiplier and record a signal. The signal is recorded accordingly.
- A reference detector monitors the emission from the zinc lamp and is used to correct for fluctuations in lamp intensity.

Exhaust air is scrubbed with a charcoal scrubber to eliminate hydrocarbons and SO<sub>2</sub>. This air is then clean enough for use in the hydrocarbon kicker to remove hydrocarbons from the incoming sample air.

The Serinus 50 sulphur dioxide analyser consists of five main assemblies:

- The pneumatics to transfer sample and exhaust gas,
- The sensors for the measurement of SO<sub>2</sub> (optical cell) and other relevant parameters,
- The control system which encompasses all circuit boards controlling sensors and pneumatic,
- The power supply which supplies power for all the instrument processors,
- The communication module to access data.

**Particle filter:**

The particulate filter is a Teflon 5 micron (µm) filter with a diameter of 47 mm. This filter eliminates all particles larger than 5 µm that could interfere with sample measurements.

**Hydrocarbon scrubber**

The hydrocarbon scrubber removes interfering hydrocarbons from the sample air. To this effect a counter current exchange is used, where an air with a lower concentration of hydrocarbons moves in an opposite direction to air with a higher concentration. The high concentrations of hydrocarbons diffuse through a selective permeation membrane to the low concentration exhaust air and are removed. Increasing the flow of the low concentration air also increases the rate of diffusion.

**Sample gas pump**

Manufacturer: Thomas, Type: 617CD22-194 C

During performance testing, the sample gas pump mentioned above was used for the laboratory as well as in the field test. As far as the models Serinus 10 (ozone), Serinus 30 (CO) and Serinus 50 (SO<sub>2</sub>) are concerned, one pump can be operated with up to two analysers. However, operation of the Serinus 40 (NO<sub>x</sub>) requires one sample gas pump per analyser.

**General notes**

This certificate is based upon the equipment tested. The manufacturer is responsible for ensuring that on-going production complies with the requirements of the EN 15267. The manufacturer is required to maintain an approved quality management system controlling the manufacture of the certified product. Both the product and the quality management systems shall be subject to regular surveillance.

If a product of the current production does not conform to the certified product, TÜV Rheinland Energy & Environment GmbH must be notified at the address given on page 1.

A certification mark with an ID-Number that is specific to the certified product is presented on page 1 of this certificate. This certification mark may be applied to the product or used in advertising materials for the certified product.

This document as well as the certification mark remains property of TÜV Rheinland Energy & Environment GmbH. With revocation of the publication the certificate loses its validity. After the expiration of the certificate and on requests of the TÜV Rheinland Energy & Environment GmbH this document shall be returned and the certificate mark must not be employed anymore.

The relevant version of this certificate and its expiration is also accessible on the internet: [gal1.de](http://gal1.de).



### History of documents

Certification of Serinus 50 is based on the documents listed below and the regular, continuous monitoring of the Quality Management System of the manufacturer:

### Initial certification according to EN 15267

Certificate No. 0000040205\_00: 29 April 2014  
Expiry date of the certificate: 31 March 2019  
Test report: 936/21221977/B dated 8 October 2013  
TÜV Rheinland Energie und Umwelt GmbH  
Publication: BAnz AT 01.04.2014 B12, chapter IV number 3.1  
UBA announcement dated 27 February 2014

### Notifications

Statement issued by TÜV Rheinland Energie und Umwelt GmbH dated 12 September 2014  
Publication: BAnz AT 02.04.2015 B5, chapter IV notification 7  
UBA announcement dated 25 February 2015  
(Soft- and hardware changes)

Statement issued by TÜV Rheinland Energy GmbH dated 13 October 2016  
Publication: BAnz AT 15.03.2017 B6, chapter V notification 8  
UBA announcement dated 22 February 2017  
(Software changes)

### Renewal of certificates

Certificate No. 0000040205\_01: 1 April 2019  
Expiry date of the certificate: 30 June 2020

### Notifications

Statement issued by TÜV Rheinland Energy GmbH dated 10 October 2018  
Publication: BAnz AT 26.03.2019 B7, chapter IV notification 18  
UBA announcement dated 27 February 2019  
(Software changes)

Statement issued by TÜV Rheinland Energy GmbH dated 20 September 2019  
Publication: BAnz AT 24.03.2020 B7, chapter IV notification 22  
UBA announcement dated 24 February 2020  
(Software changes)

### Renewal of certificates

Certificate No. 0000040205\_02: 1 July 2020  
Expiry date of the certificate: 30 June 2025

### Notifications

Statement issued by TÜV Rheinland Energy GmbH dated 14 July 2020  
Publication: BAnz AT 03.05.2021 B9, chapter III notification 12  
UBA announcement dated 31 March 2021  
(Software changes)

Statement issued by TÜV Rheinland Energy GmbH dated 20 August 2021  
Publication: BAnz AT 11.04.2022 B10, chapter VI notification 6  
UBA announcement dated 9 March 2022  
(Software changes and new producer name formerly Ecotech Pty. Ltd.)

Statement issued by TÜV Rheinland Energy GmbH dated 5 September 2022  
Publication: BAnz AT 20.03.2023 B6, chapter IV notification 60  
UBA announcement dated 21 February 2023  
(Software changes)

Statement issued by TÜV Rheinland Energy GmbH dated 10 August 2023  
Publication: BAnz AT 10.05.2024 B7, chapter V notification 46  
UBA announcement dated 19 March 2024  
(Software changes)

**Renewal of certificates**

Certificate No. 0000040205\_03: 27 June 2025  
Expiry date of the certificate: 30 June 2030



Expanded uncertainty laboratory, system 1

Measuring device: Ecotech Serinus 50		Serial No.: 13-0086 (Device 1)		132		nmol/mol	
Measured component: SO <sub>2</sub>		1h-limit value:					
No.	Performance characteristic	Performance criterion	Result	Partial uncertainty	Square of partial uncertainty		
1	Repeatability standard deviation at zero	≤ 1.0 nmol/mol	0.050	u <sub>z</sub>	0.002		
2	Repeatability standard deviation at 1h-limit value	≤ 3.0 nmol/mol	0.240	u <sub>1h</sub>	0.0050		
3	"lack of fit" at 1h-limit value	≤ 4.0% of measured value	2.620	u <sub>1h</sub>	3.9868		
4	Sensitivity coefficient of sample gas pressure at 1h-limit value	≤ 2.0 nmol/mol/kPa	0.340	u <sub>gp</sub>	7.2852		
5	Sensitivity coefficient of sample gas temperature at 1h-limit value	≤ 1.0 nmol/mol/K	0.050	u <sub>gt</sub>	0.1609		
6	Sensitivity coefficient of surrounding temperature at 1h-limit value	≤ 1.0 nmol/mol/K	0.305	u <sub>st</sub>	6.1146		
7	Sensitivity coefficient of electrical voltage at 1h-limit value	≤ 0.30 nmol/mol/V	0.027	u <sub>v</sub>	0.0608		
8a	Interferent H <sub>2</sub> O with 21 nmol/mol	≤ 10 nmol/mol (Zero)	0.010	u <sub>go</sub>	5.0688		
8b	Interferent H <sub>2</sub> S with 200 nmol/mol	≤ 10 nmol/mol (Span)	3.040	U <sub>int,pos</sub>			
8c	Interferent NH <sub>3</sub> with 200 nmol/mol	≤ 5.0 nmol/mol (Zero)	1.600				
		≤ 5.0 nmol/mol (Span)	2.390				
		≤ 5.0 nmol/mol (Zero)	-0.290				
		≤ 5.0 nmol/mol (Span)	1.080				
8d	Interferent NO with 500 nmol/mol	≤ 5.0 nmol/mol (Zero)	3.420	or	5.83	34.0086	
8e	Interferent NO <sub>2</sub> with 200 nmol/mol	≤ 5.0 nmol/mol (Span)	2.850				
		≤ 5.0 nmol/mol (Zero)	0.100				
		≤ 5.0 nmol/mol (Span)	0.740				
		≤ 10 nmol/mol (Zero)	1.250				
8f	Interferent m-Xylene with 1 µmol/mol	≤ 10 nmol/mol (Span)	3.050	U <sub>int,neg</sub>			
9	Averaging effect	≤ 7.0% of measured value	-2.930	u <sub>av</sub>	4.9861		
18	Difference sample/calibration port	≤ 1.0%	0.220	u <sub>sc</sub>	0.0843		
21	Uncertainty of test gas	≤ 3.0%	2.000	u <sub>sg</sub>	1.7424		
				Combined standard uncertainty		u <sub>c</sub>	
				Expanded uncertainty		U	
				Relative expanded uncertainty		W	
				Maximum allowed expanded uncertainty		W <sub>req</sub>	
						nmol/mol	
						nmol/mol	
						%	
						%	

Expanded uncertainty laboratory, system 2

Measuring device: Ecotech Serinus 50		Serial-No.: 13-0097 (Device 2)		132		nmol/mol	
Measured component: SO <sub>2</sub>		1h-limit value:					
No.	Performance characteristic	Performance criterion	Result	Partial uncertainty	Square of partial uncertainty		
1	Repeatability standard deviation at zero	≤ 1.0 nmol/mol	0.000	u <sub>r,z</sub>	0.0000		
2	Repeatability standard deviation at 1h-limit value	≤ 3.0 nmol/mol	0.230	u <sub>r,h</sub>	0.0048		
3	"lack of fit" at 1h-limit value	≤ 4.0% of measured value	1.590	u <sub>lf,h</sub>	1.4683		
4	Sensitivity coefficient of sample gas pressure at 1h-limit value	≤ 2.0 nmol/mol/kPa	0.270	u <sub>sp</sub>	4.5625		
5	Sensitivity coefficient of sample gas temperature at 1h-limit value	≤ 1.0 nmol/mol/K	0.030	u <sub>st</sub>	0.0587		
6	Sensitivity coefficient of surrounding temperature at 1h-limit value	≤ 1.0 nmol/mol/K	0.152	u <sub>st</sub>	1.5295		
7	Sensitivity coefficient of electrical voltage at 1h-limit value	≤ 0.30 nmol/mol/V	0.028	u <sub>v</sub>	0.0701		
8a	Interferent H <sub>2</sub> O with 21 nmol/mol	≤ 10 nmol/mol (Zero) ≤ 10 nmol/mol (Span)	-0.510 3.060	u <sub>go</sub>	4.4660		
8b	Interferent H <sub>2</sub> S with 200 nmol/mol	≤ 5.0 nmol/mol (Zero) ≤ 5.0 nmol/mol (Span)	1.410 2.210	u <sub>int,pos</sub> or u <sub>int,neg</sub>	5.48		
8c	Interferent NH <sub>3</sub> with 200 nmol/mol	≤ 5.0 nmol/mol (Zero) ≤ 5.0 nmol/mol (Span)	-0.310 0.230				
8d	Interferent NO with 500 nmol/mol	≤ 5.0 nmol/mol (Zero) ≤ 5.0 nmol/mol (Span)	3.670 4.160			30.0628	
8e	Interferent NO <sub>2</sub> with 200 nmol/mol	≤ 5.0 nmol/mol (Zero) ≤ 5.0 nmol/mol (Span)	1.000 0.310				
8f	Interferent m-Xylene with 1 µmol/mol	≤ 10 nmol/mol (Zero) ≤ 10 nmol/mol (Span)	0.860 2.660				
9	Averaging effect	≤ 7.0% of measured value	-2.620	u <sub>av</sub>	3.9868		
18	Difference sample/calibration port	≤ 1.0%	0.280	u <sub>sc</sub>	0.1366		
21	Uncertainty of test gas	≤ 3.0%	2.000	u <sub>cg</sub>	1.7424		
Combined standard uncertainty				u <sub>c</sub>	6.9346	nmol/mol	
Expanded uncertainty				U	13.8692	nmol/mol	
Relative expanded uncertainty				W	10.51	%	
Maximum allowed expanded uncertainty				W <sub>req</sub>	15	%	



Combined uncertainty, laboratory and field, system 1

Measuring device: Ecotec h Serinus 50		Serial-No.: 13-0095 (Device 1)		132		nmol/mol	
Measured component: SO <sub>2</sub>		1h-limit value:					
No.	Performance characteristic	Performance criterion	Result	Partial uncertainty	Square of partial uncertainty		
1	Repeatability standard deviation at zero	≤ 1.0 nmol/mol	0.050	U <sub>rz</sub>	0.02	0.0002	
2	Repeatability standard deviation at 1h-limit value	≤ 3.0 nmol/mol	0.240	U <sub>rh</sub>	not considered, as ur.lh = 0.07 < ur.f	-	
3	"lack of fit" at 1h-limit value	≤ 4.0% of measured value	2.620	U <sub>rh</sub>	2.00	3.9668	
4	Sensitivity coefficient of sample gas pressure at 1h-limit value	≤ 2.0 nmol/mol/kPa	0.340	U <sub>gp</sub>	2.70	7.2852	
5	Sensitivity coefficient of sample gas temperature at 1h-limit value	≤ 1.0 nmol/mol/K	0.050	U <sub>gt</sub>	0.40	0.1609	
6	Sensitivity coefficient of surrounding temperature at 1h-limit value	≤ 1.0 nmol/mol/K	0.305	U <sub>st</sub>	2.47	6.1146	
7	Sensitivity coefficient of electrical voltage at 1h-limit value	≤ 0.30 nmol/mol/V	0.027	U <sub>v</sub>	0.25	0.0608	
8a	Interferent H <sub>2</sub> O with 21 nmol/mol	≤ 10 nmol/mol (Zero)	0.010				
8b	Interferent H <sub>2</sub> S with 200 nmol/mol	≤ 10 nmol/mol (Span)	3.040	U <sub>H2O</sub>	2.25	5.0688	
8c	Interferent NH <sub>3</sub> with 200 nmol/mol	≤ 5.0 nmol/mol (Span)	1.600	U <sub>NH3,pos</sub>			
8d	Interferent NO with 500 nmol/mol	≤ 5.0 nmol/mol (Zero)	-0.290				
8e	Interferent NO <sub>2</sub> with 200 nmol/mol	≤ 5.0 nmol/mol (Span)	1.080				
8f	Interferent m-Xylene with 1 µmol/mol	≤ 5.0 nmol/mol (Zero)	3.420				
9	Averaging effect	≤ 5.0 nmol/mol (Zero)	2.850	or	5.83	34.0086	
10	Reproducibility standard deviation under field conditions	≤ 7.0% of measured value	0.100	U <sub>rel,neg</sub>			
11	Long term drift at zero level	≤ 5.0% of average over 3 months	0.740	U <sub>av</sub>	-2.23	4.9861	
12	Long term drift at span level	≤ 5.0% of max. of certification range	3.740	U <sub>rl</sub>	4.94	24.3720	
18	Difference sample/calibration port	≤ 1.0%	-0.940	U <sub>li,z</sub>	-0.54	0.2945	
21	Uncertainty of test gas	≤ 3.0%	3.810	U <sub>li,h</sub>	2.90	8.4310	
			0.220	U <sub>ass</sub>	0.29	0.0843	
			2.000	U <sub>cg</sub>	1.32	1.7424	
		Combined standard uncertainty		U <sub>c</sub>		9.8283	nmol/mol
		Expanded uncertainty		U		19.6567	nmol/mol
		Relative expanded uncertainty		W		14.89	%
		Maximum allowed expanded uncertainty		W <sub>req</sub>		15	%

Combined uncertainty, laboratory and field, system 2

Measuring device: Ecotech Sennius 50		Serial-No.: 13-0097 (Device 2)		132		nmol/mol	
Measured component: SO <sub>2</sub>		1h-limit value:					
No.	Performance characteristic	Performance criterion	Result	Partial uncertainty	Square of partial uncertainty		
1	Repeatability standard deviation at zero	≤ 1.0 nmol/mol	0.000	U <sub>r,z</sub>	0.00	0.0000	
2	Repeatability standard deviation at 1h-limit value	≤ 3.0 nmol/mol	0.230	U <sub>r,1h</sub>	not considered, as U <sub>r,1h</sub> = 0.06 < U <sub>r,f</sub>	-	
3	"lack of fit" at 1h-limit value	≤ 4.0% of measured value	1.590	U <sub>f,1h</sub>	1.21	1.4683	
4	Sensitivity coefficient of sample gas pressure at 1h-limit value	≤ 2.0 nmol/mol/kPa	0.270	U <sub>sp</sub>	2.14	4.5625	
5	Sensitivity coefficient of sample gas temperature at 1h-limit value	≤ 1.0 nmol/mol/K	0.030	U <sub>gt</sub>	0.24	0.0587	
6	Sensitivity coefficient of surrounding temperature at 1h-limit value	≤ 1.0 nmol/mol/K	0.152	U <sub>st</sub>	1.24	1.5295	
7	Sensitivity coefficient of electrical voltage at 1h-limit value	≤ 0.30 nmol/mol/V	0.028	U <sub>v</sub>	0.26	0.0701	
8a	Interferent H <sub>2</sub> O with 21 nmol/mol	≤ 10 nmol/mol (Zero)	-0.510				
8b	Interferent H <sub>2</sub> S with 200 nmol/mol	≤ 10 nmol/mol (Span)	3.060	U <sub>go</sub>	2.11	4.4660	
8c	Interferent NH <sub>3</sub> with 200 nmol/mol	≤ 5.0 nmol/mol (Zero)	1.410	U <sub>nt,pos</sub>			
8d	Interferent NO with 500 nmol/mol	≤ 5.0 nmol/mol (Span)	2.210				
8e	Interferent NO <sub>2</sub> with 200 nmol/mol	≤ 5.0 nmol/mol (Zero)	-0.310				
8f	Interferent m-Xylene with 1 µmol/mol	≤ 5.0 nmol/mol (Span)	0.230				
9	Averaging effect	≤ 5.0 nmol/mol (Zero)	3.670	or	5.48	30.0628	
10	Reproducibility standard deviation under field conditions	≤ 5.0% of average over 3 months	4.160				
11	Long term drift at zero level	≤ 5.0 nmol/mol (Zero)	1.000				
12	Long term drift at span level	≤ 5.0 nmol/mol (Span)	0.310				
18	Difference sample/calibration port	≤ 10 nmol/mol (Zero)	0.860	U <sub>nt,neg</sub>			
21	Uncertainty of test gas	≤ 7.0% of measured value	2.660	U <sub>nt,neg</sub>			
		≤ 5.0% of average over 3 months	-2.620	U <sub>av</sub>	-2.00	3.9868	
		≤ 4.0 nmol/mol	3.740	U <sub>f</sub>	4.94	24.3720	
		≤ 5.0% of max. of certification range	1.470	U <sub>cl,z</sub>	0.85	0.7203	
		≤ 1.0%	3.540	U <sub>cl,1h</sub>	2.70	7.2784	
		≤ 3.0%	0.280	U <sub>acc</sub>	0.37	0.1366	
		Combined standard uncertainty	2.000	U <sub>c</sub>	1.32	1.7424	nmol/mol
		Expanded uncertainty		U		8.9696	nmol/mol
		Relative expanded uncertainty		W		17.9393	%
		Maximum allowed expanded uncertainty		W <sub>req</sub>		13.59	%
						15	%