

CERTIFICATE

of Product Conformity (QAL1)

Certificate No.: 0000038507_02

AMS designation: AS32M for nitrogen dioxide

Manufacturer: Environnement S.A.
111, Boulevard Robespierre
78304 Poissy Cedex
France

Test Laboratory: TÜV Rheinland Energy GmbH

This is to certify that the AMS has been tested and certified according to the standards
VDI 4202-1 (2010), VDI 4203-3 (2010), EN 14211 (2012),
Guide to the demonstration of equivalence of ambient air monitoring methods (2010),
EN 15267-1 (2009) and EN 15267-2 (2009).

Certification is awarded in respect of the conditions stated in this certificate
(this certificate contains 10 pages).




Suitability Tested
Complying with
2008/50/EC
EN 15267
Regular
Surveillance
www.tuv.com
ID 0000038507


Publication in the German Federal Gazette
(BAnz) of 01 April 2014

German Federal Environment Agency
Dessau, 22 July 2018

This certificate will expire on:
22 July 2023

TÜV Rheinland Energy GmbH
Cologne, 21 July 2018


Dr Marcel Langner
Head of Section II 4.1


ppa. Dr Peter Wilbring

www.umwelt-tuv.eu
tre@umwelt-tuv.eu
Phone: + 49 221 806-5200

TÜV Rheinland Energy GmbH
Am Grauen Stein
51105 Köln

Test institute accredited to EN ISO/IEC 17025:2005 by DAkkS (German Accreditation Body).
This accreditation is limited to the accreditation scope defined in the enclosure to the certificate D-PL-11120-02-00.
qal1.de info@qal1.de

Test Report:	936/21219819/B dated 9 September 2013
Initial certification:	23 July 2013
Expiry date:	22 July 2023
Certificate:	Renewal (of previous certificate 0000038507_01 dated 29 April 2014 valid until 22 July 2018)
Publication:	BAz AT 01.04.2014 B12, chapter IV no. 4.2

Approved application

The certified AMS is suitable for continuous ambient air monitoring of nitrogen dioxide (stationary operation).

The suitability of the AMS for this application was assessed on the basis of a laboratory test and a three-months field test.

The AMS is approved for an ambient temperature range of 0 °C to +30 °C.

The notification of suitability of the AMS, performance testing and the uncertainty calculation have been effected on the basis of the regulations applicable at the time of testing. As changes in legal provisions are possible, any potential user should ensure, in consultation with the manufacturer, that this AMS is suitable for monitoring the limit values relevant to the application.

Any potential user should ensure, in consultation with the manufacturer, that this AMS is suitable for the intended purpose.

Basis of the certification

This certification is based on:

- Test report no. 936/21219819/B dated 9 September 2013 issued by TÜV Rheinland Energie und Umwelt GmbH
- Suitability announced by the German Federal Environment Agency (UBA) as the relevant body
- The ongoing surveillance of the product and the manufacturing process

Publication in the German Federal Gazette: BAnz AT 01.04.2014 B12, chapter IV no. 4.2,
UBA announcement dated 27 February 2014:

AMS designation:

AS32M for nitrogen dioxide

Manufacturer:

Environnement S.A., Poissy, France

Field of application:

For continuous ambient air monitoring of nitrogen dioxide (stationary operation)

Measuring range during performance testing:

Component	Certification range	Unit
Nitrogen dioxide	0–500	µg/m ³

Software version:

3.6.a

Restrictions:

None

Notes:

1. The measuring system must be operated inside a lockable measurement container.
2. The test report on performance testing is available on the internet at www.qal1.de.
3. Equivalence with the reference method was demonstrated for the component NO₂ in accordance with the requirements of the "Demonstration of Equivalence of Ambient Air Monitoring Methods".
4. Supplementary testing (demonstration of equivalence with the reference method) as regards Federal Environment Agency (UBA) notice of 03 July 2013 (BAnz AT 23.07.2013, chapter III number 1.1).

Test Laboratory:

TÜV Rheinland Energie und Umwelt GmbH, Cologne

Report no.: 936/21219819/B dated 9 September 2013

Publication in the German Federal Gazette: BAnz AT 26.08.2015 B4, chapter V
notification 49, UBA announcement dated 22 July 2015:

**49 Notification as regards Federal Environment Agency (UBA) notice
of 27 February 2014 (BAnz AT 01.04.2014 B12, chapter IV number 4.2)**

The current software version of the AS32M measuring system for NO₂ manufactured by Environnement S.A. is:

v1.05 (calculation process)

v3.6.h (display process)

To increase leak tightness, the diameter of the orifice was increased from 0.35 mm to 0.36 mm.

Statement issued by TÜV Rheinland Energie und Umwelt GmbH
dated 14 March 2015

Certified product

This certification applies to automated measurement systems conforming to the following description:

The AS32M air quality monitor is a continuous nitrogen dioxide analyser. The measuring principle is based on direct UV absorption. The instrument was designed for continuous ambient air monitoring of nitrogen dioxide. The measuring principle of the AS32M is based on the CAPS technique (Cavity Attenuated Phase Shift Spectroscopy).

The sampling is performed by a pump mounted at the end of the fluid circuit through a Teflon tube connected to the rear panel of the analyser. An assembly of two 3-way solenoid valves enables the selection one of the three inlets of the analyser: "sample", "zero air" or "span gas". Protection against dust is ensured by a Teflon filter (PTFE) connected to the "sample" inlet.

A PERMA-PURE dryer is used to dry the sample gas. The dryer is made of two concentric tubes, the internal tube is made of a special water-permeable polymer. The molecules are transferred, through this tube, from the side where the water content is the highest to the side where it is the lowest. To ensure a lower partial pressure of water on the outside of the polymer tube, the tube periphery is placed under vacuum condition and rinsed by a portion of the flow rate leaving the tube.

After drying, the sampled gas passes through a dust filter (made of microfibers of borosilicate glass bound in PTFE) which holds 99.5% of the particles that have an aerodynamic diameter higher than 10 nm. This enables to avoid the optical interference generated by the light scattering induced by the particles whose diameter is higher than the emission wavelength (450 nm).

From here, the sampled gas is led into the optical cavity. The optical cavity is a stainless steel hollow cylinder, closed at each end by a semi-transparent mirror of high reflectivity. The luminous source placed upstream of the M1 inlet mirror of the cavity is an LED which emits light at 450 nm. A convergent lens placed between the LED and the M1 mirror focuses the beam of light. A phototube placed behind the M2 outlet mirror of the cavity detects the photons transmitted by this mirror. Between the M2 mirror and the detector, a convergent lens focuses the beam on the detector and a band-pass interference filter centred around 450 ± 10 nm enables to select the photons that have a wavelength between 440 and 460 nm.

The current software versions are: v1.05 (calculation process)
v3.6.h (display process)

General remarks

This certificate is based upon the equipment tested. The manufacturer is responsible for ensuring that on-going production complies with the requirements of the EN 15267. The manufacturer is required to maintain an approved quality management system controlling the manufacturing process for the certified product. Both the product and the quality management systems shall be subject to regular surveillance.

If a product of the current production does not conform to the certified product, TÜV Rheinland Energy GmbH must be notified at the address given on page 1.

A certification mark with an ID number that is specific to the certified product is presented on page 1 of this certificate. This certification mark may be applied to the product or used in advertising materials for the certified product.

This document as well as the certification mark remains property of TÜV Rheinland Energy GmbH. Upon revocation of the publication the certificate loses its validity. After the expiration of the certificate and on request of TÜV Rheinland Energy GmbH this document shall be returned and the certificate mark must no longer be used.

The relevant version of this certificate and its expiration date are also accessible on the internet at qal1.de.

Certification of the AS32M measuring system is based on the documents listed below and the regular, continuous surveillance of the manufacturer's quality management system:

Initial certification according to EN 15267

Certificate no. 0000038507: 20 August 2013
Expiry date of the certificate: 22 July 2018

Test report: 936/21219819/A dated 11 March 2013
TÜV Rheinland Energie und Umwelt GmbH, Cologne
Publication: BAuz AT 23.07.2013 B4, chapter III no. 1.1
UBA announcement dated 03 July 2013

Supplementary testing according to EN 15267

Certificate no. 0000038507_01: 29 April 2014
Expiry date of the certificate: 22 July 2018

Test report: 936/21219819/B dated 9 September 2013
TÜV Rheinland Energie und Umwelt GmbH, Cologne
Publication: BAuz AT 01.04.2014 B12, chapter IV no. 4.2
UBA announcement dated 27 February 2014

Notifications in accordance with EN 15267

Statement issued by TÜV Rheinland Energie und Umwelt GmbH dated 14 March 2015
Publication: BAuz AT 26.08.2015 B4, chapter V notification 49
UBA announcement dated 22 July 2015
(new software version, modification of the blind's diameter)

Renewal of the certificate

Certificate no. 0000038507_02: 22 July 2018
Expiry date of the certificate: 22 July 2023

Expanded uncertainty based on the results of the laboratory testing of system 1

Instrument: Component:	Environnement AS32M NO2	Serial.-No.: SN 1 (001)	200 µg/m³
		1h-limit value: 11h-limit value:	
No.	Performance characteristic	Performance criteria	Result
1	Repeatability at zero	≤ 1.92 µg/m³	0.200
2	Repeatability at concentration ct "lack of fit"	≤ 5.76 µg/m³ ≤ 4.0% of measured value	1.300 1.100
3	Sensitivity coefficient of sample gas pressure	≤ 8.0 µg/m³/kPa	0.137
4	Sensitivity coefficient of sample gas temperature	≤ 5.76 µg/m³/K	0.072
5	Sensitivity coefficient of surrounding temperature	≤ 5.76 µg/m³/K	0.200
6	Sensitivity coefficient of electrical voltage	≤ 0.57 µg/m³/V	0.034
7	Interference of H2O at 21 mmol/mol	≤ 9.6 µg/m³ (zero)	0.200
8a	Interference of CO2 at 500 µmol/mol	≤ 9.6 µg/m³ (span)	-1.800
8b	Interference of NH3 at 200 nmol/mol	≤ 9.6 µg/m³ (span)	0.200
8c	Averaging effect	≤ 7.0% of measured value	-0.600
9	Difference sample/calibration port	≤ 1%	0.200
18	Converter efficiency	≥ 98 %	---
21	Uncertainty calibration gas	≤ 3%	2.000
23			ucg
		combined standard uncertainty	uc
		expanded uncertainty	Uc
		expanded uncertainty actual	Uc,rel
		expanded uncertainty required	Ureq,rel.
		µg/m³	µg/m³
		%	%
		15	15

Expanded uncertainty based on the results of the laboratory testing of Device 2

Instrument: Component:	Environnement AS32M NO2	Serial.-No. SN 2 (002)	200 $\mu\text{g}/\text{m}^3$
1h-limit value:			
No.	Performance characteristic	Result	Partial uncertainty
1	Repeatability at zero	$\leq 1.92 \mu\text{g}/\text{m}^3$	$0.100 \text{ U}_{r,Z}$
2	Repeatability at concentration ct "lack of fit"	$\leq 5.76 \mu\text{g}/\text{m}^3$ $\leq 4.0\% \text{ of measured value}$	$1.600 \text{ U}_{r,IV}$ $1.500 \text{ U}_{l,IV}$
3	Sensitivity coefficient of sample gas pressure	$\leq 8.0 \mu\text{g}/\text{m}^3/\text{kPa}$	0.119 U_{gp}
4	Sensitivity coefficient of sample gas temperature	$\leq 5.76 \mu\text{g}/\text{m}^3/\text{K}$	0.021 U_{gt}
5	Sensitivity coefficient of surrounding temperature	$\leq 5.76 \mu\text{g}/\text{m}^3/\text{K}$	0.170 U_{st}
6	Sensitivity coefficient of electrical voltage	$\leq 0.57 \mu\text{g}/\text{m}^3/\text{V}$	0.011 U_V
7	Interference of H2O at 21 mmol/mol	$\leq 9.6 \mu\text{g}/\text{m}^3/\text{zero}$	0.000 U_{H2O}
8a	Interference of CO2 at 500 $\mu\text{mol}/\text{mol}$	$\leq 9.6 \mu\text{g}/\text{m}^3/\text{(span)}$	-1.44 U_{CO2}
8b	Interference of NH3 at 200 nmol/mol	$\leq 9.6 \mu\text{g}/\text{m}^3/\text{(span)}$	2.0833 U_{NH3}
8c	Averaging effect	$\leq 9.6 \mu\text{g}/\text{m}^3/\text{(span)}$	$7.3633 \text{ U}_{int, pos}$
9	Difference sample/calibration port	$\leq 1\% \text{ of measured value}$	$2.700 \text{ U}_{int,neg}$
18	Converter efficiency	$\geq 98\%$	0.040 U_{Dsc}
21	Uncertainty calibration gas	$\leq 3\%$	$---$ 2.000 U_{cg}
23			3.12 U_{av} 0.08 U_{Dsc} 0.0064 U_{Dsc} 0.0000 U_{cg} 4.0000 U_{cg} 5.4724 U_c 10.9449 U_c $5.47 \text{ U}_{c, rel}$ $15 \text{ U}_{req, rel}$
$\mu\text{g}/\text{m}^3$			

Expanded uncertainty based on the results of the laboratory and field testing of system 1

Instrument: Component:	Environnement AS32M NO2	Serial.-No. SN 1 (001)	1h-limit value: 200 µg/m³
Performance characteristic			
No.	Performance criteria	Result	Partial uncertainty
1	Repeatability at zero	≤ 1.92 µg/m³	0.200 $U_{r,z}$ 0.04
2	Repeatability at concentration ct	≤ 5.76 µg/m³	1.300 $U_{r,lv}$
3	"lack of fit"	≤ 4.0% of measured value	1.100 $U_{l,lv}$ 1.27
4	Sensitivity coefficient of sample gas pressure	≤ 8.0 µg/m³/kPa	0.137 U_{gp} 0.95
5	Sensitivity coefficient of sample gas temperature	≤ 5.76 µg/m³/K	0.072 U_{gt} 0.71
6	Sensitivity coefficient of surrounding temperature	≤ 5.76 µg/m³/K	0.200 U_{st} 1.98
7	Sensitivity coefficient of electrical voltage	≤ 0.57 µg/m³/V	0.034 U_{v} 0.67
8a	Interference of H2O at 21 mmol/mol	≤ 9.6 µg/m³ (zero)	0.200 U_{H2O} -1.04
8b	Interference of CO2 at 500 µmol/mol	≤ 9.6 µg/m³ (zero)	0.200 $U_{int, pos}$
8c	Interference of NH3 at 200 nmol/mol	≤ 9.6 µg/m³ (zero)	0.200 $U_{int, neg}$ 3.64
9	Averaging effect	≤ 7.0% of measured value	-0.600 U_{av} -0.69
10	Reproducibility under field conditions	≤ 5.0% of the average of 3 Mon.	1.770 U_{rf} 3.54
11	Long term drift at zero level	≤ 9.36 µg/m³	1.160 $U_{dl,z}$ 0.67
12	Long term drift at span level	≤ 5.0% of certification range	1.810 $U_{dl,lv}$ 2.09
18	Difference sample/calibration port	≤ 1%	0.200 U_{bsc} 0.40
21	Converter efficiency	≤ 98	--- U_{EC} 0.00
23	Uncertainty calibration gas	≤ 3%	2.000 U_{cg} 2.00
combined standard uncertainty		U_c	7.4975 µg/m³
expanded uncertainty		U_c	14.9950 µg/m³
expanded uncertainty actual		$U_{c,rel}$	7.50 %
expanded uncertainty required		$U_{req,rel}$	15 %

Expanded uncertainty based on the results of the laboratory and field testing of Device 2

Instrument:	Environnement AS32M	Serial.-No.	SN 2 (002)		
Component:	NO2	1h-limit value:	200 µg/m³		
No.	Performance characteristic	Performance criteria	Result	Partial uncertainty	Square of uncertainty
1	Repeatability at zero	≤ 1.92 µg/m³	0.100	U _{r,z}	0.02
2	Repeatability at concentration ct	≤ 5.76 µg/m³	1.600	U _{r,IV}	-
3	"lack of fit"	≤ 4.0% of measured value	1.500	U _{l,IV}	1.73
4	Sensitivity coefficient of sample gas pressure	≤ 8.0 µg/(m³/kPa)	0.119	U _{gp}	0.82
5	Sensitivity coefficient of sample gas temperature	≤ 5.76 µg/(m³/K)	0.021	U _{gt}	0.21
6	Sensitivity coefficient of surrounding temperature	≤ 5.76 µg/(m³/K)	0.170	U _{st}	1.68
7	Sensitivity coefficient of electrical voltage	≤ 0.57 µg/(m³/V)	0.011	U _V	0.22
8a	Interference of H2O at 21 mmol/mol	≤ 9.6 µg/m³ (zero)	0.200	U _{H2O}	-1.44
8b	Interference of CO2 at 500 µmol/mol	≤ 9.6 µg/m³ (zero)	-2.500	U _{int, pos}	0.300
8c	Interference of NH3 at 200 nmol/mol	≤ 9.6 µg/m³ (zero)	2.000	U _{int,neg}	0.100
9	Averaging effect	≤ 7.0% of measured value	2.700	U _{av}	3.12
10	Reproducibility under field conditions	≤ 5.0% of the average of 3 Mon.	1.770	U _f	3.54
11	Long term drift at zero level	≤ 9.36 µg/m³	1.170	U _{d,l,z}	0.68
12	Long term drift at span level	≤ 5.0% of certification range	1.730	U _{d,l,IV}	2.00
18	Difference sample/calibration port	≤ 1 %	0.040	U _{psc}	0.08
21	Converter efficiency	≥ 98	---	U _{EC}	0.0000
23	Uncertainty calibration gas	≤ 3%	2.000	U _{CQ}	2.00
combined standard uncertainty					
expanded uncertainty					
expanded uncertainty actual					
expanded uncertainty required					