

# CERTIFICATE

## of Product Conformity (QAL1)

Certificate No.: 0000043106\_04

**Certified AMS:** O3 42e / O3 42e\* for O<sub>3</sub>

**Manufacturer:** ENVEA  
111, Boulevard Robespierre  
78304 Poissy Cedex  
France

**Test Institute:** TÜV Rheinland Energy & Environment GmbH

**This is to certify that the AMS has been tested  
and found to comply with the standards  
VDI 4202-1 (2018), EN 14625 (2012), EN 14625 (2024)  
as well as EN 15267-1 (2009) and EN 15267-2 (2023).**

Certification is awarded in respect of the conditions stated in this certificate  
(this certificate contains 11 pages).  
The present certificate replaces certificate 0000043106\_03 dated 28 March 2025.



Suitability Tested  
Complying with  
2008/50/EC  
EN 15267  
Regular  
Surveillance  
[www.tuv.com](http://www.tuv.com)  
ID 0000043106


Publication in the German Federal Gazette  
(BAnz) of 1 August 2016

German Environment Agency


Dessau, 23 March 2026

This certificate will expire on:  
22 March 2031

TÜV Rheinland  
Energy & Environment GmbH  
Cologne, 20 March 2026



Dr. Marcel Langner  
Head of Section II 4



I. V. Guido Baum

[www.umwelt-tuv.eu](http://www.umwelt-tuv.eu)  
[qal1-info@tuv.com](mailto:qal1-info@tuv.com)  
Tel. + 49 221 806-5200

TÜV Rheinland Energy & Environment GmbH  
Am Grauen Stein  
51105 Köln

Test institute accredited to EN ISO/IEC 17025 by DAkkS (German Accreditation Body).  
This accreditation is limited to the accreditation scope defined in the enclosure to the certificate D-PL-11120-02-00.

<b>Test report:</b>	936/21225396/B dated 26 February 2016 and Addendum EuL/21264142/C_V1 dated 7 February 2025
<b>Initial certification:</b>	2 April 2015
<b>Expiry date:</b>	22 March 2031
<b>Publication:</b>	BAnz AT 01.08.2016 B11, chapter III No. 1.1 and BAnz AT 31.10.2025 B5, chapter IV notification 38

### **Approved application**

The tested AMS is suitable for continuous immission measurement of O<sub>3</sub> in stationary use.

The suitability of the AMS for this application was assessed on the basis of a laboratory test and a three-month field test at a location close to transport links.

The AMS is approved for an ambient temperature range of +0 °C to +40 °C.

The notification of suitability of the AMS, performance testing and the uncertainty calculation have been effected on the basis of the regulations applicable at the time of testing. As changes in legal provisions are possible, any potential user should ensure that this AMS is suitable for monitoring the measured values relevant to the application.

Any potential user should ensure, in consultation with the manufacturer, that this AMS is suitable for the intended use.

### **Basis of the certification**

This certification is based on:

- Test report 936/21225396/B dated 26 February 2016 of TÜV Rheinland Energie und Umwelt GmbH and Addendum EuL/21264142/C\_V1 dated 7 February 2025 of TÜV Rheinland Energy & Environment GmbH
- Suitability announced by the German Federal Environment Agency (UBA) as the relevant body
- The ongoing surveillance of the product and the manufacturing process

Publication in the German Federal Gazette: BAnz AT 01.08.2016 B11, chapter III No. 1.1,  
Announcement by UBA dated 14 July 2016:

**AMS designation:**

O3 42e\* / O3 42e for Ozone

**Manufacturer:**

Environnement S.A., Poissy, France

**Field of application:**

For continuous ambient air monitoring of ozone (stationary operation).

**Measuring ranges during the performance test:**

Component	Certification range	Unit
Ozone	0 – 500	µg/m <sup>3</sup>

**Software version:**

O3 42e Version: 1.0.4

O3 42e\* Version: 1.0.3

**Restrictions:**

None

**Notes:**

1. Measured values are displayed by means of a connected PC or Laptop.
2. The performance test also includes the O3 42e instrument version with integrated display.
3. The report on the performance test is available online at [www.qal1.de](http://www.qal1.de).
4. Supplementary testing (optimization of the LED's wavelength range as well as pressure compensation) as regards Federal Environment Agency (UBA) notices of 25 February 2015 (BAnz AT 02.04.2015 B5, chapter III number 1.1) and of 22 July 2015 (BAnz AT 26.08.2015 B4, chapter IV notification 47).

**Test Report:**

TÜV Rheinland Energie und Umwelt GmbH, Cologne  
Report No.: 936/21225396/B dated 26 February 2016

Publication in the German Federal Gazette: BAnz AT 19.05.2025 B3, chap. IV notification 86,  
Announcement by UBA dated 2 April 2025

**86 Notification as regards Federal Environment Agency (UBA) notices of 14 July 2016 (BAnz AT 01.08.2016 B11, chapter III number 1.1) and of 21 February 2023 (BAnz AT 20.03.2023 B6, chapter IV notification 68)**

The current software version for the O342e\*/O342e measuring system for ozone from ENVEA is:  
v1.2.b

Software version v1.2.a can also be used.  
The measuring system can be equipped with a revised version of the ARM20 board.

Statement issued by TÜV Rheinland Energy & Environment GmbH dated 4 October 2024

The basis for this certificate is the publication of the following announcement:

Publication in the German Federal Gazette: BAnz AT 31.10.2025 B5, chap. IV notification 38,  
Announcement by UBA dated 27 August 2025

**38 Notification as regards Federal Environment Agency (UBA) notices of 14 July 2016 (BAnz AT 01.08.2016 B11, chapter III number 1.1) and of 02 April 2025 (BAnz AT 19.05.2025 B3, chapter IV, notification 86)**

The current software version for the O342e\*/O342e measuring system for ozone from ENVEA is  
v1.3.a

From software version v1.3.a, the measuring system fulfils the requirements of EN 14625 (2024 version).

An addendum to the test report with the report number EuL/21264142/C\_V1 can be viewed on the internet at [www.qal1.de](http://www.qal1.de).

Statement issued by TÜV Rheinland Energy & Environment GmbH dated 20 May 2025

## Certified product

This certificate applies to automated measurement systems conforming to the following description:

The ambient air monitor O342e\* is a continuous ozone monitor. The measurement principle is based on ultraviolet absorption. The instrument was developed for the continuous measurement of ozone concentrations in ambient air.

The measurement principle of the O342e\* is based on UV photometry according to the Beer-Lambert law. The absorption spectrum of ozone has its maximum in the wavelength range of 250 to 270 nanometres. The monochromatic UV-LED light source of the O342e\* is adjusted to a wavelength of 255 nm and therefore within the maximum absorption range of ozone.

The O342e\* analyser uses non-dispersive ultraviolet (UV) absorption technology to measure ozone concentrations. The sample to be analysed is led to the measurement module via a dust filter. The measurement module consists of the following parts:

- LED for monochromatic UV light with a wavelength of 255 nm, placed under a protective cover, which is fastened with 4 screws. The LED card is directly connected to the card of the reference photodetector.
- Two photodetector cards: the reference photodetector card for measuring the energy of the incoming LED light (UV<sub>0</sub>) and the photodetector card for measuring UV absorption, which enables detection of signals  $i$  and  $i_0$ . Both cards are mounted beneath a protective cover to protect them against interfering light.
- The optical chamber consists of a beam splitter and a convex, flat lens for concentrating the light on the reference photodetector. In the optical chamber, the LED light can be distributed to reference photodetector and measuring chamber.
- A measurement chamber consisting of a glass tube and two mechanical parts at the inlet and outlet where the LED light is absorbed. The optical path length for the sample gas is 400 mm.
- Cycle solenoid valve by means of which the sample gas can either cyclically or alternately be changed over to cycle channel  $i$  or cycle channel  $i_0$ .
- A flow restrictor which regulates the sample gas flow to 55 litres/hour. The excess flow valve is mounted at the fluid outlet of the measurement chamber.
- Ozone filter which can filter out any trace of ozone from the sample gas
- Connection for the pressure sensor
- Type PT1000 temperature sensor
- Gas inlet

The AMS is available in two versions:

- The version O342e is fitted with a TFT LCD coloured display with backlight and a touch screen function. Signal output as well as operation can also be carried out via the web browser using an external PC connected via Ethernet.
- The version O342e\* is not fitted with a display. Signal output as well as operation can only be operated via the web browser on an external PC connected via Ethernet.

Apart from that, both versions of the AMS are of identical design.

### **General notes**

This certificate is based upon the equipment tested. The manufacturer is responsible for ensuring that on-going production complies with the requirements of the EN 15267. The manufacturer is required to maintain an approved quality management system controlling the manufacture of the certified product. Both the product and the quality management systems shall be subject to regular surveillance.

If a product of the current production does not conform to the certified product, TÜV Rheinland Energy & Environment GmbH must be notified at the address given on page 1.

A certification mark with an ID-Number that is specific to the certified product is presented on page 1 of this certificate. This certification mark may be applied to the product or used in advertising materials for the certified product.

This document as well as the certification mark remains property of TÜV Rheinland Energy & Environment GmbH. With revocation of the publication the certificate loses its validity. After the expiration of the certificate and on request of TÜV Rheinland Energy & Environment GmbH this document shall be returned and the certificate mark must not be employed anymore.

The relevant version of this certificate and its expiration is also accessible on the internet: [qal1.de](http://qal1.de).

### **History of documents**

Certification of O342e / O342e\* is based on the documents listed below and the regular, continuous monitoring of the Quality Management System of the manufacturer:

### **Initial certification according to EN 15267**

Certificate No. 0000043106\_00: 30 April 2015  
Expiry date of the certificate: 1 April 2020  
Test report: 936/21225396/A dated 1 October 2014  
TÜV Rheinland Energie und Umwelt GmbH  
Publication: BAnz AT 02.04.2015 B5, chapter III number 1.1  
UBA announcement dated 25 February 2015

### **Notifications**

Statement issued by TÜV Rheinland Energie und Umwelt GmbH dated 16 March 2015  
Publication: BAnz AT 26.08.2015 B4, chapter V notification 47  
UBA announcement dated 22 July 2015  
(Software changes)

### **Supplementary testing according to EN 15267**

Certificate No. 0000043106\_01: 19 August 2016  
Expiry date of the certificate: 1 April 2020  
Test report: 936/21225396/B dated 26 February 2016  
TÜV Rheinland Energie und Umwelt GmbH  
Publication: BAnz AT 01.08.2016 B11, chapter III number 1.1  
UBA announcement dated 14 July 2016

### **Notifications**

Statement issued by TÜV Rheinland Energy GmbH dated 27 September 2018  
Publication: BAnz AT 26.03.2019 B7, chapter IV notification 30  
UBA announcement dated 27 February 2019  
(Software changes)

Statement issued by TÜV Rheinland Energy GmbH dated 1 October 2019  
Publication: BAnz AT 24.03.2020 B7, chapter IV notification 35  
UBA announcement dated 24 February 2020  
(Software changes and new producer name formerly Environnement S.A.)

### **Renewal of certificates**

Certificate No. 0000043106\_02: 2 April 2020  
Expiry date of the certificate: 1 April 2025

### **Notifications**

Statement issued by TÜV Rheinland Energy GmbH dated 9 September 2020  
Publication: BAnz AT 03.05.2021 B9, chapter III notification 22  
UBA announcement dated 31 March 2021  
(Software and hardware changes)

Statement issued by TÜV Rheinland Energy GmbH dated 16 September 2021  
Publication: BAnz AT 11.04.2022 B10, chapter VI notification 12  
UBA announcement dated 9 March 2022  
(Software changes)

Statement issued by TÜV Rheinland Energy GmbH dated 9 September 2022  
Publication: BAnz AT 20.03.2023 B6, chapter IV notification 68  
UBA announcement dated 21 February 2023  
(Software changes)

**Renewal of certificates**

Certificate No. 0000043106\_03: 28 March 2025  
Expiry date of the certificate: 1 April 2030

**Notifications**

Statement issued by TÜV Rheinland Energy & Environment GmbH dated 4 October 2024  
Publication: BAnz AT 19.05.2025 B3, chapter IV notification 86  
UBA announcement dated 2 April 2025  
(Software and hardware changes)

**Certificate based on a notification**

Certificate No. 0000043106\_04: 23 March 2026  
Expiry date of the certificate: 22 March 2031  
Statement issued by TÜV Rheinland Energy & Environment GmbH dated 20 May 2025  
Addendum EuL/21264142/C\_V1 dated 7 February 2025  
Publication: BAnz AT 31.10.2025 B5, chapter IV notification 38  
UBA announcement dated 27 August 2025

### Expanded uncertainty laboratory (DIN EN 14625:2012), system 1

Measuring device:		Environnement O3 42e*		Serial-No.:		SN 12 / SN 23	
Measured component:		O <sub>3</sub>		1h-alert threshold:		120 nmol/mol	
No.	Performance characteristic	Performance criterion	Result	Partial uncertainty		Square of partial uncertainty	
1	Repeatability standard deviation at zero	≤ 1.0 nmol/mol	0,020	u <sub>r,z</sub>	0,00	0,0000	
2	Repeatability standard deviation at 1h-alert threshold	≤ 3.0 nmol/mol	0,070	u <sub>r,h</sub>	0,01	0,0001	
3	"lack of fit" at 1h-alert threshold	≤ 4.0% of measured value	0,880	u <sub>if</sub>	0,61	0,3717	
4	Sensitivity coefficient of sample gas pressure at 1h-alert threshold	≤ 2.0 nmol/mol/kPa	0,130	u <sub>sp</sub>	1,44	2,0656	
5	Sensitivity coefficient of sample gas temperature at 1h-alert threshold	≤ 1.0 nmol/mol/K	0,010	u <sub>st</sub>	0,11	0,0122	
6	Sensitivity coefficient of surrounding temperature at 1h-alert threshold	≤ 1.0 nmol/mol/K	0,392	u <sub>st</sub>	2,92	8,5280	
7	Sensitivity coefficient of electrical voltage at 1h-alert threshold	≤ 0.30 nmol/mol/V	0,010	u <sub>v</sub>	0,13	0,0166	
8a	Interferent H <sub>2</sub> O with 21 mmol/mol	≤ 10 nmol/mol (Zero)	0,300	u <sub>ISO</sub>	-2,14	4,5862	
		≤ 10 nmol/mol (Span)	-2,870				
8b	Interferent Toluene with 0,5 µmol/mol	≤ 5.0 nmol/mol (Zero)	0,870	u <sub>int,pos</sub>	0,81	0,6533	
		≤ 5.0 nmol/mol (Span)	0,400				
8c	Interferent Xylene with 0,5 µmol/mol	≤ 5.0 nmol/mol (Zero)	1,760	or			
		≤ 5.0 nmol/mol (Span)	1,000				
9	Averaging effect	≤ 7.0% of measured value	-4,280	u <sub>av</sub>	-2,97	8,7928	
18	Difference sample/calibration port	≤ 1.0%	-0,350	u <sub>acc</sub>	-0,42	0,1764	
21	Uncertainty of test gas	≤ 3.0%	2,000	u <sub>cg</sub>	1,20	1,4400	
Combined standard uncertainty				u <sub>c</sub>		5,1617	nmol/mol
Expanded uncertainty				U		10,3234	nmol/mol
Relative expanded uncertainty				W		8,60	%
Maximum allowed expanded uncertainty				W <sub>req</sub>		15	%

### Expanded uncertainty laboratory (DIN EN 14625:2012), system 2

Measuring device:		Environnement O3 42e*		Serial-No.:		SN 14 / SN 24	
Measured component:		O <sub>3</sub>		1h-alert threshold:		120 nmol/mol	
No.	Performance characteristic	Performance criterion	Result	Partial uncertainty		Square of partial uncertainty	
1	Repeatability standard deviation at zero	≤ 1.0 nmol/mol	0,020	u <sub>r,z</sub>	0,00	0,0000	
2	Repeatability standard deviation at 1h-alert threshold	≤ 3.0 nmol/mol	0,090	u <sub>r,h</sub>	0,01	0,0002	
3	"lack of fit" at 1h-alert threshold	≤ 4.0% of measured value	2,370	u <sub>if</sub>	1,64	2,6961	
4	Sensitivity coefficient of sample gas pressure at 1h-alert threshold	≤ 2.0 nmol/mol/kPa	0,030	u <sub>sp</sub>	0,34	0,1124	
5	Sensitivity coefficient of sample gas temperature at 1h-alert threshold	≤ 1.0 nmol/mol/K	0,020	u <sub>st</sub>	0,22	0,0489	
6	Sensitivity coefficient of surrounding temperature at 1h-alert threshold	≤ 1.0 nmol/mol/K	0,231	u <sub>st</sub>	1,72	2,9614	
7	Sensitivity coefficient of electrical voltage at 1h-alert threshold	≤ 0.30 nmol/mol/V	0,020	u <sub>v</sub>	0,26	0,0665	
8a	Interferent H <sub>2</sub> O with 21 mmol/mol	≤ 10 nmol/mol (Zero)	0,530	u <sub>ISO</sub>	-2,01	4,0590	
		≤ 10 nmol/mol (Span)	-2,700				
8b	Interferent Toluene with 0,5 µmol/mol	≤ 5.0 nmol/mol (Zero)	0,930	u <sub>int,pos</sub>	0,75	0,5633	
		≤ 5.0 nmol/mol (Span)	0,400				
8c	Interferent Xylene with 0,5 µmol/mol	≤ 5.0 nmol/mol (Zero)	1,100	or			
		≤ 5.0 nmol/mol (Span)	0,900				
9	Averaging effect	≤ 7.0% of measured value	-4,770	u <sub>av</sub>	-3,30	10,9214	
18	Difference sample/calibration port	≤ 1.0%	-0,360	u <sub>acc</sub>	-0,43	0,1866	
21	Uncertainty of test gas	≤ 3.0%	2,000	u <sub>cg</sub>	1,20	1,4400	
Combined standard uncertainty				u <sub>c</sub>		4,8017	nmol/mol
Expanded uncertainty				U		9,6033	nmol/mol
Relative expanded uncertainty				W		8,00	%
Maximum allowed expanded uncertainty				W <sub>req</sub>		15	%

### Combined uncertainty, laboratory and field (EN 14625:2012), system 1

Measuring device:		Environnement O3 42e*		Serial-No.:		SN 12 / SN 23	
Measured component:		O <sub>3</sub>		1h-alert threshold:		120 nmol/mol	
No.	Performance characteristic	Performance criterion	Result	Partial uncertainty		Square of partial uncertainty	
1	Repeatability standard deviation at zero	≤ 1.0 nmol/mol	0,020	u <sub>r,z</sub>	0,00	0,0000	
2	Repeatability standard deviation at 1h-alert threshold	≤ 3.0 nmol/mol	0,070	u <sub>r,h</sub>	not considered, as u <sub>r,h</sub> = 0,01 < u <sub>r,f</sub>	-	
3	"lack of fit" at 1h-alert threshold	≤ 4.0% of measured value	0,880	u <sub>if</sub>	0,61	0,3717	
4	Sensitivity coefficient of sample gas pressure at 1h-alert threshold	≤ 2.0 nmol/mol/kPa	0,130	u <sub>sp</sub>	1,44	2,0656	
5	Sensitivity coefficient of sample gas temperature at 1h-alert threshold	≤ 1.0 nmol/mol/K	0,010	u <sub>st</sub>	0,11	0,0122	
6	Sensitivity coefficient of surrounding temperature at 1h-alert threshold	≤ 1.0 nmol/mol/K	0,392	u <sub>st</sub>	2,92	8,5280	
7	Sensitivity coefficient of electrical voltage at 1h-alert threshold	≤ 0.30 nmol/mol/V	0,010	u <sub>v</sub>	0,13	0,0166	
8a	Interferent H <sub>2</sub> O with 21 mmol/mol	≤ 10 nmol/mol (Zero)	0,300	u <sub>ISO</sub>	-2,14	4,5862	
		≤ 10 nmol/mol (Span)	-2,870				
8b	Interferent Toluene with 0,5 µmol/mol	≤ 5.0 nmol/mol (Zero)	0,870	u <sub>int,pos</sub>	0,81	0,6533	
		≤ 5.0 nmol/mol (Span)	0,400				
8c	Interferent Xylene with 0,5 µmol/mol	≤ 5.0 nmol/mol (Zero)	1,760	or			
		≤ 5.0 nmol/mol (Span)	1,000				
9	Averaging effect	≤ 7.0% of measured value	-4,280	u <sub>av</sub>	-2,97	8,7928	
10	Reproducibility standard deviation under field conditions	≤ 5.0% of average over 3 months	2,590	u <sub>f</sub>	3,11	9,6597	
11	Long term drift at zero level	≤ 5.0 nmol/mol	0,590	u <sub>LT</sub>	0,34	0,1160	
12	Long term drift at span level	≤ 5.0% of max. of certification range	1,190	u <sub>LT</sub>	0,82	0,6797	
18	Difference sample/calibration port	≤ 1.0%	-0,350	u <sub>acc</sub>	-0,42	0,1764	
21	Uncertainty of test gas	≤ 3.0%	2,000	u <sub>cg</sub>	1,20	1,4400	
Combined standard uncertainty				u <sub>c</sub>		6,0908	nmol/mol
Expanded uncertainty				U		12,1817	nmol/mol
Relative expanded uncertainty				W		10,15	%
Maximum allowed expanded uncertainty				W <sub>req</sub>		15	%

### Combined uncertainty, laboratory and field (EN 14625:2012), system 2

Measuring device:		Environment O3 42e*		Serial-No.:		SN 14 / SN 24	
Measured component:		O <sub>3</sub>		1h-alert threshold:		120 nmol/mol	
No.	Performance characteristic	Performance criterion	Result	Partial uncertainty		Square of partial uncertainty	
1	Repeatability standard deviation at zero	≤ 1.0 nmol/mol	0,020	u <sub>r,z</sub>	0,00	0,0000	
2	Repeatability standard deviation at 1h-alert threshold	≤ 3.0 nmol/mol	0,090	u <sub>r,h</sub>	not considered, as ur,h = 0,01 < ur,f	-	
3	"lack of fit" at 1h-alert threshold	≤ 4.0% of measured value	2,370	u <sub>if</sub>	1,64	2,6961	
4	Sensitivity coefficient of sample gas pressure at 1h-alert threshold	≤ 2.0 nmol/mol/kPa	0,030	u <sub>sp</sub>	0,34	0,1124	
5	Sensitivity coefficient of sample gas temperature at 1h-alert threshold	≤ 1.0 nmol/mol/K	0,020	u <sub>st</sub>	0,22	0,0489	
6	Sensitivity coefficient of surrounding temperature at 1h-alert threshold	≤ 1.0 nmol/mol/K	0,231	u <sub>st</sub>	1,72	2,9614	
7	Sensitivity coefficient of electrical voltage at 1h-alert threshold	≤ 0.30 nmol/mol/V	0,020	u <sub>v</sub>	0,26	0,0665	
8a	Interferent H <sub>2</sub> O with 21 mmol/mol	≤ 10 nmol/mol (Zero)	0,530	u <sub>int,zo</sub>	-2,01	4,0590	
		≤ 10 nmol/mol (Span)	-2,700				
8b	Interferent Toluene with 0,5 μmol/mol	≤ 5.0 nmol/mol (Zero)	0,930	u <sub>int,pos</sub>	0,75	0,5633	
		≤ 5.0 nmol/mol (Span)	0,400				
8c	Interferent Xylene with 0,5 μmol/mol	≤ 5.0 nmol/mol (Zero)	1,100	u <sub>int,neg</sub>	0,900	10,9214	
		≤ 5.0 nmol/mol (Span)	0,900				
9	Averaging effect	≤ 7.0% of measured value	-4,770	u <sub>av</sub>	-3,30	10,9214	
10	Reproducibility standard deviation under field conditions	≤ 5.0% of average over 3 months	2,990	u <sub>r,f</sub>	3,11	9,6597	
11	Long term drift at zero level	≤ 5.0 nmol/mol	0,850	u <sub>l,z</sub>	0,49	0,2408	
12	Long term drift at span level	≤ 5.0% of max. of certification range	1,450	u <sub>l,h</sub>	1,00	1,0092	
18	Difference sample/calibration port	≤ 1.0%	-0,360	u <sub>dsc</sub>	-0,43	0,1866	
21	Uncertainty of test gas	≤ 3.0%	2,000	u <sub>cg</sub>	1,20	1,4400	
Combined standard uncertainty				u <sub>c</sub>	5,8280	nmol/mol	
Expanded uncertainty				U	11,6560	nmol/mol	
Relative expanded uncertainty				W	9,71	%	
Maximum allowed expanded uncertainty				W <sub>req</sub>	15	%	

### Expanded uncertainty laboratory (EN 14625:2024), system 1

Measuring device:		O342e / O342e*		Serial-No.:		SN 12 / 23 / 1688	
Measured component:		O <sub>3</sub>		1h-alert threshold:		120 nmol/mol	
No.	Performance characteristic	Performance criterion	Result	Partial uncertainty		Square of partial uncertainty	
1	Repeatability standard deviation at zero	≤ 1.0 nmol/mol	0,020	u <sub>r,z</sub>	0,00	0,00	
2	Repeatability standard deviation at 1h-alert threshold	≤ 3.0 nmol/mol	0,070	u <sub>r,h</sub>	0,01	0,00	
3	"lack of fit" at 1h-alert threshold	≤ 4.0% of measured value	0,880	u <sub>if</sub>	0,61	0,37	
4	Sensitivity coefficient of sample gas pressure at 1h-alert threshold	≤ 2.0 nmol/mol/kPa	0,160	u <sub>sp</sub>	1,18	1,39	
5	Sensitivity coefficient of sample gas temperature at 1h-alert threshold	≤ 1.0 nmol/mol/K	0,030	u <sub>st</sub>	0,39	0,15	
6	Sensitivity coefficient of surrounding temperature at 1h-alert threshold	≤ 1.0 nmol/mol/K	0,183	u <sub>st</sub>	2,39	5,69	
7	Sensitivity coefficient of electrical voltage at 1h-alert threshold	≤ 0.30 nmol/mol/V	0,010	u <sub>v</sub>	0,13	0,02	
8a	Interferent H <sub>2</sub> O with 21 mmol/mol	≤ 10 nmol/mol (Zero)	0,300	u <sub>int,zo</sub>	-2,14	4,59	
		≤ 10 nmol/mol (Span)	-2,870				
8b	Interferent Toluene with 0,5 μmol/mol	≤ 5.0 nmol/mol (Zero)	0,870	u <sub>int,pos</sub>	0,81	0,65	
		≤ 5.0 nmol/mol (Span)	0,400				
8c	Interferent Xylene with 0,5 μmol/mol	≤ 5.0 nmol/mol (Zero)	1,760	u <sub>int,neg</sub>	1,000	8,79	
		≤ 5.0 nmol/mol (Span)	1,000				
9	Averaging effect	≤ 7.0% of measured value	-4,280	u <sub>av</sub>	-2,97	8,79	
18	Difference sample/calibration port	≤ 1.0%	-0,350	u <sub>dsc</sub>	-0,42	0,18	
21	Uncertainty of test gas	≤ 3.0%	2,000	u <sub>cg</sub>	1,20	1,44	
Combined standard uncertainty				u <sub>c</sub>	4,82	nmol/mol	
Expanded uncertainty				U	9,65	nmol/mol	
Relative expanded uncertainty				W	8,04	%	
Maximum allowed expanded uncertainty				W <sub>req</sub>	15	%	

### Expanded uncertainty laboratory (EN 14625:2024), system 2

Measuring device:		O342e / O342e*		Serial-No.:		SN 14 / 24 / 1735	
Measured component:		O <sub>3</sub>		1h-alert threshold:		120 nmol/mol	
No.	Performance characteristic	Performance criterion	Result	Partial uncertainty		Square of partial uncertainty	
1	Repeatability standard deviation at zero	≤ 1.0 nmol/mol	0,020	u <sub>r,z</sub>	0,00	0,00	
2	Repeatability standard deviation at 1h-alert threshold	≤ 3.0 nmol/mol	0,090	u <sub>r,h</sub>	0,01	0,00	
3	"lack of fit" at 1h-alert threshold	≤ 4.0% of measured value	2,370	u <sub>if</sub>	1,64	2,70	
4	Sensitivity coefficient of sample gas pressure at 1h-alert threshold	≤ 2.0 nmol/mol/kPa	0,210	u <sub>sp</sub>	1,56	2,45	
5	Sensitivity coefficient of sample gas temperature at 1h-alert threshold	≤ 1.0 nmol/mol/K	0,040	u <sub>st</sub>	0,52	0,27	
6	Sensitivity coefficient of surrounding temperature at 1h-alert threshold	≤ 1.0 nmol/mol/K	0,049	u <sub>st</sub>	0,64	0,41	
7	Sensitivity coefficient of electrical voltage at 1h-alert threshold	≤ 0.30 nmol/mol/V	0,020	u <sub>v</sub>	0,26	0,07	
8a	Interferent H <sub>2</sub> O with 21 mmol/mol	≤ 10 nmol/mol (Zero)	0,530	u <sub>int,zo</sub>	-2,01	4,06	
		≤ 10 nmol/mol (Span)	-2,700				
8b	Interferent Toluene with 0,5 μmol/mol	≤ 5.0 nmol/mol (Zero)	0,930	u <sub>int,pos</sub>	0,75	0,56	
		≤ 5.0 nmol/mol (Span)	0,400				
8c	Interferent Xylene with 0,5 μmol/mol	≤ 5.0 nmol/mol (Zero)	1,100	u <sub>int,neg</sub>	0,900	10,92	
		≤ 5.0 nmol/mol (Span)	0,900				
9	Averaging effect	≤ 7.0% of measured value	-4,770	u <sub>av</sub>	-3,30	10,92	
18	Difference sample/calibration port	≤ 1.0%	-0,360	u <sub>dsc</sub>	-0,43	0,19	
21	Uncertainty of test gas	≤ 3.0%	2,000	u <sub>cg</sub>	1,20	1,44	
Combined standard uncertainty				u <sub>c</sub>	4,80	nmol/mol	
Expanded uncertainty				U	9,60	nmol/mol	
Relative expanded uncertainty				W	8,00	%	
Maximum allowed expanded uncertainty				W <sub>req</sub>	15	%	

### Combined uncertainty, laboratory and field (EN 14625:2024), system 1

Measuring device:		O342e / O342e*		Serial-No.:		SN 12 / 23 / 1688	
Measured component:		O <sub>3</sub>		1h-alert threshold:		120 nmol/mol	
No.	Performance characteristic	Performance criterion	Result	Partial uncertainty		Square of partial uncertainty	
1	Repeatability standard deviation at zero	≤ 1.0 nmol/mol	0,020	U <sub>r,z</sub>	0,00	0,00	
2	Repeatability standard deviation at 1h-alert threshold	≤ 3.0 nmol/mol	0,070	U <sub>r,th</sub>	not considered, as ur,th = 0,01 < ur,f	-	
3	"lack of fit" at 1h-alert threshold	≤ 4.0% of measured value	0,880	U <sub>l,th</sub>	0,61	0,37	
4	Sensitivity coefficient of sample gas pressure at 1h-alert threshold	≤ 2.0 nmol/mol/kPa	0,160	U <sub>sp</sub>	1,18	1,39	
5	Sensitivity coefficient of sample gas temperature at 1h-alert threshold	≤ 1.0 nmol/mol/K	0,030	U <sub>gt</sub>	0,39	0,15	
6	Sensitivity coefficient of surrounding temperature at 1h-alert threshold	≤ 1.0 nmol/mol/K	0,183	U <sub>st</sub>	2,39	5,69	
7	Sensitivity coefficient of electrical voltage at 1h-alert threshold	≤ 0.30 nmol/mol/V	0,010	U <sub>v</sub>	0,13	0,02	
8a	Interferent H <sub>2</sub> O with 21 mmol/mol	≤ 10 nmol/mol (Zero) ≤ 10 nmol/mol (Span)	0,300 -2,870	U <sub>int,0</sub>	-2,14	4,59	
8b	Interferent Toluene with 0,5 μmol/mol	≤ 5.0 nmol/mol (Zero) ≤ 5.0 nmol/mol (Span)	0,870 0,400	U <sub>int,pos</sub> or	0,81	0,65	
8c	Interferent Xylene with 0,5 μmol/mol	≤ 5.0 nmol/mol (Zero) ≤ 5.0 nmol/mol (Span)	1,760 1,000	U <sub>int,neg</sub>			
9	Averaging effect	≤ 7.0% of measured value	-4,280	U <sub>av</sub>	-2,97	8,79	
10	Reproducibility standard deviation under field conditions	≤ 5.0% of average over 3 months	2,590	U <sub>r,f</sub>	3,11	9,66	
11	Long term drift at zero level	≤ 5.0 nmol/mol	0,590	U <sub>d,z</sub>	0,34	0,12	
12	Long term drift at span level	≤ 5.0% of max. of certification range	1,190	U <sub>d,1,th</sub>	0,82	0,68	
18	Difference sample/calibration port	≤ 1.0%	-0,350	U <sub>pac</sub>	-0,42	0,18	
21	Uncertainty of test gas	≤ 3.0%	2,000	U <sub>cg</sub>	1,20	1,44	
Combined standard uncertainty				U <sub>c</sub>		5,81	nmol/mol
Expanded uncertainty				U		11,61	nmol/mol
Relative expanded uncertainty				W		9,68	%
Maximum allowed expanded uncertainty				W <sub>req</sub>		15	%

### Combined uncertainty, laboratory and field (EN 14625:2024), system 2

Measuring device:		O342e / O342e*		Serial-No.:		SN 14 / 24 / 1735	
Measured component:		O <sub>3</sub>		1h-alert threshold:		120 nmol/mol	
No.	Performance characteristic	Performance criterion	Result	Partial uncertainty		Square of partial uncertainty	
1	Repeatability standard deviation at zero	≤ 1.0 nmol/mol	0,020	U <sub>r,z</sub>	0,00	0,00	
2	Repeatability standard deviation at 1h-alert threshold	≤ 3.0 nmol/mol	0,090	U <sub>r,th</sub>	not considered, as ur,th = 0,01 < ur,f	-	
3	"lack of fit" at 1h-alert threshold	≤ 4.0% of measured value	2,370	U <sub>l,th</sub>	1,64	2,70	
4	Sensitivity coefficient of sample gas pressure at 1h-alert threshold	≤ 2.0 nmol/mol/kPa	0,210	U <sub>sp</sub>	1,56	2,45	
5	Sensitivity coefficient of sample gas temperature at 1h-alert threshold	≤ 1.0 nmol/mol/K	0,040	U <sub>gt</sub>	0,52	0,27	
6	Sensitivity coefficient of surrounding temperature at 1h-alert threshold	≤ 1.0 nmol/mol/K	0,049	U <sub>st</sub>	0,64	0,41	
7	Sensitivity coefficient of electrical voltage at 1h-alert threshold	≤ 0.30 nmol/mol/V	0,020	U <sub>v</sub>	0,26	0,07	
8a	Interferent H <sub>2</sub> O with 21 mmol/mol	≤ 10 nmol/mol (Zero) ≤ 10 nmol/mol (Span)	0,530 -2,700	U <sub>int,0</sub>	-2,01	4,06	
8b	Interferent Toluene with 0,5 μmol/mol	≤ 5.0 nmol/mol (Zero) ≤ 5.0 nmol/mol (Span)	0,930 0,400	U <sub>int,pos</sub> or	0,75	0,56	
8c	Interferent Xylene with 0,5 μmol/mol	≤ 5.0 nmol/mol (Zero) ≤ 5.0 nmol/mol (Span)	1,100 0,900	U <sub>int,neg</sub>			
9	Averaging effect	≤ 7.0% of measured value	-4,770	U <sub>av</sub>	-3,30	10,92	
10	Reproducibility standard deviation under field conditions	≤ 5.0% of average over 3 months	2,590	U <sub>r,f</sub>	3,11	9,66	
11	Long term drift at zero level	≤ 5.0 nmol/mol	0,810	U <sub>d,z</sub>	0,47	0,22	
12	Long term drift at span level	≤ 5.0% of max. of certification range	1,450	U <sub>d,1,th</sub>	1,00	1,01	
18	Difference sample/calibration port	≤ 1.0%	-0,360	U <sub>pac</sub>	-0,43	0,19	
21	Uncertainty of test gas	≤ 3.0%	2,000	U <sub>cg</sub>	1,20	1,44	
Combined standard uncertainty				U <sub>c</sub>		5,83	nmol/mol
Expanded uncertainty				U		11,65	nmol/mol
Relative expanded uncertainty				W		9,71	%
Maximum allowed expanded uncertainty				W <sub>req</sub>		15	%