

CONFIRMATION

Notification: 0000074620_00_03
on changes according to EN 15267 regarding certificate 0000074620_00 dated 17 June 2020

AMS designation: MP101M PM₁₀ for PM₁₀

Manufacturer: ENVEA
111, Boulevard Robespierre
78304 Poissy Cedex
France

German Federal Environmental Agency (UBA)

**Announcement regarding the uniform practice in
monitoring emissions and ambient air
dated 19 March 2024
Publication: BAnz AT 10.05.2024 B7**

- V. Notifications as regards the uniform practice in monitoring emissions and ambient air:**
- 49 Notification as regards Federal Environment Agency (UBA) notices of 31 March 2020 (BAnz AT 07.05.2020 B8, chapter II number 2.1) and of 28 June 2022 (BAnz AT 28.07.2022 B4, chapter III notification 40)**

The MP101M measuring system for PM₁₀ particulate matter from ENVEA can also be equipped with the RCT type C14 source from RC TRITEC AG. Until now, a C14 source of type CFCB18760 from the manufacturer Eckert & Ziegler Nuclitec GmbH has been used.


The measuring system can also be equipped with the Geiger-Müller detector type LND 72423 from the manufacturer LND, INC. Previously, a Geiger-Müller detector of type LND 72412 from the same manufacturer was used.


The current software version for the measuring system is v.4.0.n

In addition to this version, the following intermediate version is also valid: v.4.0.m

Statement issued by TÜV Rheinland Energy GmbH dated 10 August 2023

TÜV Rheinland Energy & Environment GmbH
Cologne, 24 May 2024


i. V. Dipl.-Ing. Guido Baum


i. A. Dipl.-Ing. Carsten Röllig

www.umwelt-tuv.eu
TRE@umwelt-tuv.eu
Tel. +49 221 806-5200

TÜV Rheinland Energy & Environment GmbH
Am Grauen Stein
51105 Cologne

Test institute accredited to EN ISO/IEC 17025 by DAkkS (German Accreditation Body).
This accreditation is limited to the accreditation scope defined in the enclosure to the certificate D-PL-11120-02-00.