

Confirmation

Notification: 4149451_03
on design changes in accordance with DIN EN 15267 regarding certificate no. 4061336-ts from 24 June 2025

AMS	IR-Pyrometer KT 15.67 IIP
Manufacturer	Heitronics Infrarot Messtechnik GmbH Kreuzberger Ring 40 65205 Wiesbaden Germany

German Federal Environmental Agency (UBA)

**Announcement regarding the uniform practice in monitoring emissions and immissions
dated 27 August 2025
Publication: AT 31.10.2025 B5**

IV. Notification on uniform practise

32 Notification on the UBA announcement from 2 April 2025 (BAnz AT 19.05.2025 B3, Chapter III No. 1.1)


The measuring system KT 15.67 IIP for monitoring the minimum temperature in plants requiring authorisation and plants in compliance with the 27th BImSchV, manufactured by Heitronics Infrarot Messtechnik GmbH, Wiesbaden, can alternatively be operated with the following 24 V DC power supplies for power supply:

1. Siemens LOGO!Power® power supply unit 24 V/0.6 A (Article number 6EP3330-6SB00-0AY0) for parallel power supply of up to 2 measuring systems,
2. Siemens SITOP® power supply unit 24 V/5 A (Article number 6EP3333-7SB00-0AX0) for parallel power supply of up to 16 measuring systems.

The power supply units are intended for installation in control cabinets.

Comment of TÜV Süd Industrie Service GmbH dated 19 February 2025

TÜV SÜD Industrie Service GmbH
Testing laboratory Emission measurement/ calibration
Munich, 14 November 2025



Dipl.-Ing. Hans-Jörg Eisenberger



M.Sc. Ulrich Peschke