

CERTIFICATE

of Product Conformity (QAL1)

Certificate No: 0000072203_01

Certified AMS: EM-D5200 for dust

Manufacturer: Horiba GmbH
Kaplanstrasse 5
A-3430 Tulln
Austria

Test Institute: TÜV Rheinland Energy & Environment GmbH

**This is to certify that the AMS has been tested
and found to comply with the standards
EN 15267-1 (2009), EN 15267-2 (2023), EN 15267-3 (2007),
as well as EN 14181 (2004).**

Certification is awarded in respect of the conditions stated in this certificate
(this certificate contains 9 pages).
The present certificate replaces certificate 0000072203_00 dated 7 September 2020.



Suitability Tested
EN 15267
QAL1 Certified
Regular
Surveillance

www.tuv.com
ID 0000072203


Publication in the German Federal Gazette
(BAnz) of 31 July 2020

German Environment Agency


Dessau, 27 June 2025

This certificate will expire on:
30 July 2030

TÜV Rheinland Energy &
Environment GmbH
Cologne, 26 June 2025



Dr. Marcel Langner
Head of Section II 4



ppa. Dr. Peter Wilbring

www.umwelt-tuv.eu
qal1-info@tuv.com
Tel. + 49 221 806-5200

TÜV Rheinland Energy & Environment GmbH
Am Grauen Stein
51105 Köln

Test institute accredited to EN ISO/IEC 17025 by DAkkS (German Accreditation Body).
This accreditation is limited to the accreditation scope defined in the enclosure to the certificate D-PL-11120-02-00.

Test report: 936/21242768/B dated 11 March 2020
Initial certification: 31 July 2020
Expiry date: 30 July 2030
Certificate: Renewal (of previous certificate 0000072203_00 of 7 September 2020 valid until 30 July 2025)
Publication: BAnz AT 31.07.2020 B10, chapter I No. 1.1

Approved application

The tested AMS is suitable for use at plants according to Directive 2010/75/EC, chapter III (combustion plants / 13th BImSchV:2020), chapter IV (waste incineration plants / 17th BImSchV:2013), Directive 2015/2193/EC (44th BImSchV:2022), TA Luft:2002, 30th BImSchV:2019 and 27th BImSchV:2013. The measured ranges have been selected so as to ensure as broad a field of application as possible.

The suitability of the AMS for this application was assessed on the basis of a laboratory test and a twelve month field test at a waste incineration.

The AMS is approved for an ambient temperature range of -40 °C to +60 °C.

The notification of suitability of the AMS, performance testing and the uncertainty calculation have been effected on the basis of the regulations applicable at the time of testing. As changes in legal provisions are possible, any potential user should ensure that this AMS is suitable for monitoring the emission limit values relevant to the application.

Any potential user should ensure, in consultation with the manufacturer, that this AMS is suitable for the installation at which it will be installed.

Note

The legal regulations mentioned correspond to the current state of legislation during certification. Each user should, if necessary, in consultation with the competent authority, ensure that this AMS meets the legal requirements for the intended use. In addition, it cannot be ruled out that legal regulations governing the use of a measuring device for emission monitoring may change during the lifetime of the certificate.

Basis of the certification

This certification is based on:

- Test report 936/21242768/B dated 11 March 2020 of TÜV Rheinland Energy GmbH
- Suitability announced by the German Federal Environment Agency (UBA) as the relevant body
- The ongoing surveillance of the product and the manufacturing process

Publication in the German Federal Gazette: BAnz AT 31.07.2020 B10, chapter I No. 1.1,
Announcement by UBA dated 27 May 2020:

AMS designation:

EM-D5200 for dust

Manufacturer:

HORIBA GmbH, Tulln, Austria

Field of application:

For plants requiring official approval and for plants according to the 27th BImSchV

Measuring ranges during the performance test:

Component	Certification range	Unit
Dust	0 – 7.5*	mg/m ³

* corresponds to 0 to 500 SL

Component	supplementary ranges				Unit
Dust	0 - 100	0 – 1,000	0 – 4,000	0 – 20,000	SL

Software versions:

EM-D5200: 01.10R0001
EM-5800CU: 02.02R0066
D-ESI 100: 01.11R0018

Restrictions:

None

Notes:

1. The maintenance interval is six months.
2. The AMS can be operated with the EM5800CU control unit, an EM-CB-L supply unit or the EM-CB-S connection box.
3. The AMS may be supplied with purge air either by way of the EM-CB-L supply unit or an external purge air supply.
4. The EM5800CU control unit has a digital Modbus RTU interface and a Modbus TCP in accordance with VDI 4201 parts 1 and 3 (EIA-485, serial and TCP/IP, Ethernet).
5. The EM-D5200 measuring system has a digital Modbus RTU interface in accordance with VDI 4201 parts 1 and 3 (EIA-485, serial).
6. When the EM-D5200 measuring system is combined with the EM5800CU control unit, the instrument's Modbus interface cannot be used. Instead the Modbus interface provided by the EM5800CU control unit is used.
7. When using the AMS without the EM5800CU control unit, the AMS shall be operated by means of the D-ESI 100 software on a customary PC/notebook/tablet.
8. During performance testing in accordance with EN 15267-3, the requirement for the determination coefficient R^2 of the calibration function was not fulfilled.

9. The EM5800CU control unit is available in the following versions:

- EM5800CU M (standard)
- EM5800CU C (compact housing)
- EM5800CU P (c/w purge air blower)
- EM5800CU R (housing for 19" rack mounting)

Test Institute: TÜV Rheinland Energy GmbH, Cologne
Report No.: 936/21242768/B dated 11 March 2020

Publication in the German Federal Gazette: BAnz AT 05.08.2021 B5, Chap. IV
notification 56, Announcement by UBA dated 29 June 2021:

**56 Notification as regards Federal Environment Agency (UBA) notice
of 27 May 2020 (BAnz AT 31.07.2020 B10, chapter I number 1.1)**

The above-mentioned announcement requires amendment.

The note 8 is formulated as a restriction as follows:

Restrictions:

The requirement for the determination coefficient R^2 of the calibration function in the performance test according to EN 15267-3 was not fulfilled.

The note 9, concerning the design of the control unit, becomes note 8.

Resolution of the expert committee on test reports from 3-4 November, 2020

Publication in the German Federal Gazette: BAnz AT 02.08.2023 B7, Chap. III
notification 11, Announcement by UBA dated 5 July 2023:

**11 Notification as regards Federal Environment Agency (UBA) notice
of 27 May 2020 (BAnz AT 31.07.2020 B10, chapter I number 1.1)**

The current software versions of the EM-D5200 measuring system
for dust from HORIBA GmbH are:

EM-D5200: 01.10R0001
EM5800CU: 02.02R0073
D-ESI 100: 01.11R0018

Statement issued by TÜV Rheinland Energy GmbH dated 2 February 2023

Publication in the German Federal Gazette: BAnz AT 31.10.2024 B9, Chap. IV
notification 24, Announcement by UBA dated 31 August 2024:

**24 Notification as regards Federal Environment Agency (UBA) notices
of 27 May 2020 (BAnz AT 31.07.2020 B10, chapter I number 1.1) and
of 5 July 2023 (BAnz AT 02.08.2023 B7, chapter III notification 11)**

The current software versions for the EM-D5200 measuring system
for dust from the company Horiba GmbH are:

EM-D5200: 01.10R0002
EM-5800CU: 02.02R0073
D-ESI 100: 01.11R0018

Statement issued by TÜV Rheinland Energy GmbH dated 19 February 2024

Certified product

This certificate applies to automated measurement systems conforming to the following description:

The EM-D5200 measuring system is a dust monitor that relies on scattered light (backwards scattering) as its measuring principle. Measurements are made contact-free, continuous and without sampling in the flue gas flow above dew point. The red light from a laser diode is sent into the flue gas duct and illuminates the dust particles in the measuring volume. The particles present in the measurement volume scatter this light. A photodiode then detects the backscattered light. The proportion of the measured intensity of the scattered light to the intensity of the emitted light corresponds to the particle density in the measuring volume.

The measuring system comprises the following main components:

- EM-D5200MH measuring head
and
- electric EM-CB-S connection box for voltage supply
or
- EM-CB-L supply unit with built-in purge air blower
or
- Electronic control unit EM5800CU

When using the EM-CB-S connection box or the EM-CB-L supply unit, the EM-D5200 measuring system is controlled via the D-ESI 100 software installed on a PC. The EM5800CU control unit allows for operation of the AMS without a PC and may also provide additional data outputs. The EM5800CU control unit is available in the following versions:

- EM5800CU M (standard)
- EM5800CU C (compact housing)
- EM5800CU P (c/w purge air blower)
- EM5800CU R (housing for 19" rack mounting)

When using the EM-CB-S connection box and the EM5800CU control unit, the measuring system must be equipped with an external purge air supply, e.g. pressurised air class 1 as defined in standard ISO 8573-1-1:2010.

The EM-CB-S connection box and the EM-CB-L supply unit are merely used for mains supply, signal transmission (without affecting the actual processing of measured values), and purge air supply (EM-CB-L only). The generating of measured values as well as all calculation processes relevant to measuring (incl. the analogue and digital generating of measurements) occur directly within the measuring head.

The measuring system is available in two different versions for narrower and wider measurement channels (variants "narrow" and "wide"). With respect to the variant for narrow measurement channels, the measuring volume is situated at a distance ranging from 70 to 450 mm from the aperture. As far as the variant for wider measurement channels is concerned, the measuring volume is situated at a distance ranging from 240 to 1200 mm distance from the aperture / duct wall. The performance test was carried out with the variant for larger measurement channels.

Control measurements (control functions, zero point, contamination, span point) are made by use of an automatic swing-in "shuttle" (internal reference standard). A linearity test is effected by pluggable opacity filters, that are placed in a measuring device which can be inserted in the measuring head. By swinging-in the internal reference standard device and dimming the light source, every settable measuring range (min. 0 to 100 SL) can be checked by means of this filter set. For this purpose it is not necessary to remove the instrument from the measuring location as it only needs to be opened up.

General notes

This certificate is based upon the equipment tested. The manufacturer is responsible for ensuring that on-going production complies with the requirements of the EN 15267. The manufacturer is required to maintain an approved quality management system controlling the manufacture of the certified product. Both the product and the quality management systems shall be subject to regular surveillance.

If a product of the current production does not conform to the certified product, TÜV Rheinland Energy & Environment GmbH must be notified at the address given on page 1.

A certification mark with an ID-Number that is specific to the certified product is presented on page 1 of this certificate. This certification mark may be applied to the product or used in advertising materials for the certified product.

This document as well as the certification mark remains property of TÜV Rheinland Energy & Environment GmbH. With revocation of the publication the certificate loses its validity. After the expiration of the certificate and on requests of the TÜV Rheinland Energy & Environment GmbH this document shall be returned and the certificate mark must not be employed anymore.

The relevant version of this certificate and its expiration is also accessible on the internet: **qal1.de**.

History of documents

Certification of EM-D5200 is based on the documents listed below and the regular, continuous monitoring of the Quality Management System of the manufacturer:

Initial certification according to EN 15267

Certificate No. 0000072203_00: 7 September 2020
Expiry date of the certificate: 30 July 2025
Test report: 936/21242768/B dated 11 March 2020
TÜV Rheinland Energy GmbH
Publication: BAnz AT 31.07.2020 B10, chapter I number 1.1
UBA announcement dated 27 May 2020

Notifications

Statement issued by TÜV Rheinland Energy GmbH dated 4 May 2021
Publication: BAnz AT 05.08.2021 B5, chapter IV notification 56
UBA announcement dated 29 June 2021
(Correction of the announcement)

Statement issued by TÜV Rheinland Energy GmbH dated 2 February 2023
Publication: BAnz AT 02.08.2023 B7, chapter III notification 11
UBA announcement dated 5 July 2023
(Software changes)

Statement issued by TÜV Rheinland Energy & Environment GmbH dated 19 February 2024
Publication: BAnz AT 31.10.2024 B9, chapter IV notification 24
UBA announcement dated 31 August 2024
(Software changes)

Renewal of certificates

Certificate No. 0000072203_01: 27 June 2025
Expiry date of the certificate: 30 July 2030

Calculation of overall uncertainty according to EN 14181 and EN 15267-3

Measuring system

Manufacturer	HORIBA GmbH
AMS designation	EM-D5200
Serial number of units under test	1235301 / 1235302 / 1236093 / 1236094
Measuring principle	Scattered light analysis (back scattering)

Test report

Test laboratory	TÜV Rheinland
Date of report	2020-03-11

Measured component

Certification range	Dust 0 - 7.5 mg/m ³ (0 - 500 SU)
---------------------	--

Calculation of the combined standard uncertainty

Tested parameter

			u^2
Standard deviation from paired measurements under field conditions *	u_D	0.090 mg/m ³	0.008 (mg/m ³) ²
Lack of fit	u_{lof}	-0.030 mg/m ³	0.001 (mg/m ³) ²
Zero drift from field test	$u_{d,z}$	-0.078 mg/m ³	0.006 (mg/m ³) ²
Span drift from field test	$u_{d,s}$	-0.095 mg/m ³	0.009 (mg/m ³) ²
Influence of ambient temperature at span	u_t	0.020 mg/m ³	0.000 (mg/m ³) ²
Influence of supply voltage	u_v	0.060 mg/m ³	0.004 (mg/m ³) ²
Influence of sample gas pressure	u_p	0.000 mg/m ³	0.000 (mg/m ³) ²
Uncertainty of reference material at 70% of certification range	u_{rm}	0.061 mg/m ³	0.004 (mg/m ³) ²

* The larger value is used :

"Repeatability standard deviation at set point" or

"Standard deviation from paired measurements under field conditions"

Combined standard uncertainty (u_c)

$$u_c = \sqrt{\sum (u_{max,j})^2} \quad 0.18 \text{ mg/m}^3$$

Total expanded uncertainty

$$U = u_c \cdot k = u_c \cdot 1.96 \quad 0.35 \text{ mg/m}^3$$

Relative total expanded uncertainty

Requirement of 2010/75/EU

Requirement of EN 15267-3

U in % of the ELV 5 mg/m³ **7.0**

U in % of the ELV 5 mg/m³ **30.0**

U in % of the ELV 5 mg/m³ **22.5**