

CERTIFICATE

of Product Conformity (QAL1)

Certificate No.: 0000028754_03

Certified AMS: APMA 370 for CO

Manufacturer: HORIBA, Ltd.
2 Miyanohigashi
Kisshoin Minami-ku
Kyoto 610-8510
Japan

Test Institute: TÜV Rheinland Energie und Umwelt GmbH

**This is to certify that the AMS has been tested and certified
according to the standards**

**VDI 4202-1 (2002), VDI 4203-3 (2004), EN 14626 (2012),
EN 15267-1 (2009) and EN 15267-2 (2009)**

Certification is awarded in respect of the conditions stated in this certificate
(This certificate contains 10 pages).



Publication in the German Federal Gazette
(BAnz.) of 08 April 2006

This certificate will expire on:
25 January 2021

German Federal Environment Agency
Dessau, 21 January 2016

TÜV Rheinland Energie und Umwelt GmbH
Cologne, 20 January 2016

i. A. Dr. Marcel Langner



ppa. Dr. Peter Wilbring



www.umwelt-tuv.de / www.eco-tuv.com
teu@umwelt-tuv.de
Tel. +49 221 806-5200

TÜV Rheinland Energie und Umwelt GmbH
Am Grauen Stein
51105 Cologne

Test institute accredited to EN ISO/IEC 17025:2005 by DAkkS (German Accreditation Body).
This accreditation is limited to the accreditation scope defined in the enclosure to the certificate D-PL-11120-02-00.

Test report: 936/21204643/B of 05 January 2006

Initial certification: 26 January 2011

Certificate: renewal (previous certificate 0000028754_02 of 29 April 2014
valid until 25 January 2016)

Date of expiry: 25 January 2021

Publication: BAnz. 08 April 2006, No. 70, p. 2653, chapter IV, No. 2.1

Approved application

The certified AMS is suitable for continuous ambient air monitoring (stationary operation) of CO.

The suitability of the AMS for this application was assessed on the basis of a laboratory test and a three-month field test.

The AMS is approved for the temperature range of 0 °C to +40 °C.

The notification of suitability of the AMS, performance testing, and the uncertainty calculation have been effected on the basis of the regulations valid at the time of performance testing. As changes in legal regulations are possible, any potential user should ensure that this AMS is suitable for monitoring the limit value relevant to the application.

Any potential user should ensure, in consultation with the manufacturer, that this AMS is suitable for ambient air applications at which it will be installed.

Basis of the certification

This certification is based on:

- test report 936/21204643/B dated 05 January 2006 of TÜV Rheinland Immissionsschutz und Energiesysteme GmbH
- addenda 936/21204643/B1 of 27 July 2011 and 936/21222689/B of 05 October 2013
- suitability announced by the German Environmental Agency (UBA) as the relevant body
- the ongoing surveillance of the product and the manufacturing process

Publication in the German Federal Gazette: BAnz. 08 April 2006, No. 70, p. 2653, chapter IV No. 2.1,
UBA announcement from 21 February 2006:

AMS name:

APMA 370

Manufacturer:

HORIBA, Ltd., Kyoto 610 - 8510, Japan

Distributor:

HORIBA Europe GmbH, 42799 Leichlingen

Approval:

For continuous monitoring of CO (stationary operation).

Measuring ranges during the performance test:

CO 0 - 60 mg/m³
 0 - 100 mg/m³

Software version:

Version P1000878001C

Test report:

TÜV Rheinland Immissionsschutz und Energiesysteme GmbH, Cologne
Report-No: 936/21204643/B of 05 January 2006

Publication in the German Federal Gazette: BAnz. 25 August 2009, No. 125, p. 2929, chapter III,
notification 1, UBA announcement from 03 August 2009:

1 Notification on the announcement of the Federal Environment Agency of 21 February 2006 (BAnz. p. 2655)

The current software version of the ambient air measuring system APMA 370 of the company Horiba Europe GmbH is:

P1000878001J

As an option, the pump of the type GD-6 EH of the company Horiba can be used alongside the so far used measured gas pump type N 86 KNE of the company KNF.

Statement of TÜV Rheinland Immissionsschutz und Energiesysteme GmbH
of 31 March 2009

Publication in the German Federal Gazette: BAnz. 26 January 2011, No. 14, p. 294, chapter IV,
notification 5, UBA announcement from 10 January 2011:

5 Notification on the announcement of the Federal Environment Agency of 21 February 2006 (BAnz. p. 2653, chapter IV No 2.1) and of 03 August 2009 (BAnz. p. 2929, chapter III, 1st notification)

The APMA 370 measuring system by Horiba Ltd., Japan and Horiba Europe GmbH for component CO fulfills the requirements of EN 14626. Moreover, the production and quality management of the APMA 370 measuring system for component CO complies with the requirements of EN 15267.

The report of the suitability test is available on the internet at www.qal1.de.

Statement of TÜV Rheinland Energie und Umwelt GmbH of 06 October 2010

Publication in the German Federal Gazette: BAnz. 02 March 2012, No. 36, p. 920, chapter V, notification 16, UBA announcement from 23 February 2012:

16 Notification on the announcement of the Federal Environment Agency of 21 February 2006 (BAnz. p. 2653, chapter IV No 2.1) and of 10 January 2011 (BAnz. p. 294, chapter IV 5th notification)

There is an addendum to the test report 936/21204643/B for the APMA 370 measuring system by Horiba Ltd., Japan and Horiba Europe GmbH for the component CO. The addendum has the report number 936/21204643/B1 and is an integral part of the test report 936/21204643/B after its publication and is also published on www.qal1.de.

Statement of TÜV Rheinland Energie und Umwelt GmbH of 03 November 2011

Publication in the German Federal Gazette: BAnz AT 05.03.2013 B10, chapter V, notification 7, UBA announcement from 12 February 2013:

7 Notification on the announcement of the Federal Environment Agency of 21 February 2006 (BAnz. p. 2653, chapter IV, No. 2.1) and of 23 February 2012 (BAnz. p. 920, chapter V 16th notification)

The APMA 370 measuring system for CO manufactured by Horiba Ltd., Japan and Horiba Europe GmbH can be optionally equipped with an additional calibration gas inlet. Calibration gas can be fed either before or after the sample gas filter by means of an additional three-way valve.

Statement of TÜV Rheinland Energie und Umwelt GmbH of 11 October 2012

Publication in the German Federal Gazette: BAnz AT 01.04.2014 B12, Chapter VI, notification 26, UBA announcement from 27 February 2014:

26 Notification on the announcement of the Federal Environment Agency of 21 February 2006 (BAnz. p. 2653, chapter IV No 2.1) and of 12 February 2013 (BAnz AT 05.03.2013 B10, chapter V 7th notification)

The measuring system APMA 370 for CO by Horiba Ltd., Japan and Horiba Europe GmbH fulfils the requirements of Standard EN 14626 (dated December 2012). An addendum that is an integral part of the test report 936/21222689/B can be viewed on the internet at www.qal1.de.

Statement of TÜV Rheinland Energie und Umwelt GmbH of 5 October 2013

Certified product

This certificate applies to automated measurement systems conforming to the following description:

The APMA 370 CO-Analyser operates based on the principle of non-dispersal infrared absorption. This measurement principle complies with the reference measuring method described in section 5.2 of Standard EN 14626 (2012). The attenuation of infrared light is measured after passing through the test cuvette. According to Lambert-Beer law, attenuation is a measure of the concentration of CO within the cuvette.

The APMA-370 measuring system uses the modulation effect of infrared absorption within the test gas itself if zero and test gas are led to the measuring cell at specific flow rates. The switch is made via a solenoid valve which is clocked at a frequency of 1 Hz. As long as the concentration of the measured component within the cell does not change, the output of the detector is practically zero. Thus, no zero drift occurs.

General notes

This certificate is based upon the equipment tested. The manufacturer is responsible for ensuring that on-going production complies with the requirements of the EN 15267. The manufacturer is required to maintain an approved quality management system controlling the manufacture of the certified product. Both the product and the quality management systems shall be subject to regular surveillance.

If a product of the current production does not conform to the certified product, TÜV Rheinland Energie und Umwelt GmbH must be notified at the address given on page 1.

A certification mark with an ID-Number that is specific to the certified product is presented on page 1 of this certificate. This can be applied to the product or used in publicity material for the certified product is presented on page 1 of this certificate.

This document as well as the certification mark remains property of TÜV Rheinland Energie und Umwelt GmbH. With revocation of the publication the certificate loses its validity. After the expiration of the certificate and on requests of the TÜV Rheinland Energie und Umwelt GmbH this document shall be returned and the certificate mark must not be employed anymore.

The relevant version of this certificate and the validity is also accessible on the internet: qal1.de.

Certification of APMA 370 for CO is based on the documents listed below and the regular, continuous monitoring of the Quality Management System of the manufacturer:

First suitability test:

Test report: 936/21204643/B of 05 January 2006
TÜV Rheinland Immissionsschutz und Energiesysteme GmbH, cologne
Publication: BAnz. 08 April 2006, No. 70, p. 2653, chapter IV, No. 2.1
Announcement by UBA from 21 February 2006

Notifications:

Statement of TÜV Rheinland Immissionsschutz und Energiesysteme GmbH of 31 March 2009
Publication: BAnz. 25 August 2009, No. 125, p. 2929, chapter III, notification 1
Announcement by UBA from 03 August 2009 (Software changes and hardware extension)

Initial certification according to EN 15267:

Certificate No. 0000028754: 09 February 2011
Validity of the certificate: 25 January 2016
Test report: 936/21204643/B dated 05 January 2006
TÜV Rheinland Immissionsschutz und Energiesysteme GmbH, cologne
Statement of TÜV Rheinland Energie und Umwelt GmbH of 6 October 2010
Publication: BAnz. 26 January 2011, No. 14, p. 294, chapter IV, notification 5
Announcement by UBA from 10 January 2011

Notifications according to EN 15267:

Statement of TÜV Rheinland Energie und Umwelt GmbH of 3 November 2011
with Addendum 936/21204643/B1 from 27 July 2011
TÜV Rheinland Energie und Umwelt GmbH, cologne
Publication: BAnz. 02 March 2012, No. 36, p. 920, chapter V, notification 16
UBA announcement from 23 February 2012, (addition of an Addendum)
Certificate No. 0000028754_01: 16 March 2012
Validity of the certificate until: 25 January 2016

Statement of TÜV Rheinland Energie und Umwelt GmbH of 11 October 2012
Publication: BAnz AT 05.03.2013 B10, chapter V, notification 7
Announcement by UBA from 12 February 2013, (hardware extension)

Statement of TÜV Rheinland Energie und Umwelt GmbH of 5 October 2013
with Addendum 936/21222689/B from 05 October 2013
TÜV Rheinland Energie und Umwelt GmbH, cologne
Publication: BAnz AT 01.04.2014 B12, chapter VI, notification 26
Announcement by UBA from 27 February 2014, (EN14626 (2012))
Certificate No. 0000028754_02: 29 April 2014
Validity of the certificate until: 25 January 2016

Renewal of the certificate:

Certificate No 0000028754_03: 21 January 2016
Validity of the certificate: 25 January 2021

Expanded uncertainty based on the results of the laboratory testing of System 1

Measuring device: Measured component:	Horiba APMA 370 CO	Serial-No.:	SN 10031		
		8h-limit value:	8.62 µmol/mol		
No.	Performance characteristic	Performance criterion	Result	Partial uncertainty	Square of partial uncertainty
1	Repeatability standard deviation at 8h-limit value	≤ 0.3 µmol/mol	0.035	$u_{r,z}$	0.01
2	"lack of fit" at 8h-limit value	≤ 0.4 µmol/mol	0.064	u_t	0.00
3	Sensitivity coefficient of sample gas pressure at 8h-limit value	≤ 4.0% of measured value	0.700	u_i	0.03
4	Sensitivity coefficient of sample gas temperature at 8h-limit value	≤ 0.7 µmol/mol/KPa	0.006	u_{gp}	0.02
5	Sensitivity coefficient of sample gas temperature at 8h-limit value	≤ 0.3 µmol/mol/K	0.005	u_{gt}	0.01
6	Sensitivity coefficient of surrounding temperature at 8h-limit value	≤ 0.3 µmol/mol/K	0.077	u_{st}	0.20
7	Sensitivity coefficient of electrical voltage at 8h-limit value	≤ 0.3 µmol/mol/V	0.006	u_V	0.06
8a	Interferent H ₂ O with 21 nmol/mol	≤ 1.0 µmol/mol (Zero)	0.233	u_{H2O}	0.02
8b	Interferent CO ₂ with 500 µmol/mol	≤ 1.0 µmol/mol (Span)	-0.009	$u_{CO2, pos}$	0.009
8c	Interferent NO with 1 µmol/mol	≤ 0.5 µmol/mol (Span)	0.017	$u_{NO, pos}$	0.017
8d	Interferent N ₂ O with 50 nmol/mol	≤ 0.5 µmol/mol (Zero)	-0.164	$u_{N2O, neg}$	0.164
9	Averaging effect	≤ 7.0% of measured value	1.250	u_{av}	0.06
18	Difference sample/calibration port	≤ 1.0%	0.000	u_{asc}	0.000
21	Uncertainty of test gas	≤ 3.0%	2.000	u_{cg}	0.09
		Combined standard uncertainty	u_c	0.2582 µmol/mol	
		Expanded uncertainty	U	0.5165 µmol/mol	
		Relative expanded uncertainty	W	5.99 %	
		Maximum allowed expanded uncertainty	W _{req}	15	%

Expanded uncertainty based on the results of the laboratory testing of System 2

Measuring device:	Horiba APIMA 370	Measured component:	CO	Serial-No.:	SN 10032	8h-limit value:	8.62	µmol/mol
Performance characteristic								
No.		Performance criterion	Result	Partial uncertainty	Square of partial uncertainty			
1	Repeatability standard deviation at zero	≤ 0.3 µmol/mol	0.028	$u_{r,z}$	0.01	0.0000		
2	Repeatability standard deviation at 8h-limit value	≤ 0.4 µmol/mol	0.070	u_r	0.00	0.0000		
3	"lack of fit" at 8h-limit value	≤ 4.0% of measured value	0.800	u_i	0.04	0.0016		
4	Sensitivity coefficient of sample gas pressure at 8h-limit value	≤ 0.7 µmol/mol/kPa	0.009	u_{gp}	0.02	0.0005		
5	Sensitivity coefficient of sample gas temperature at 8h-limit value	≤ 0.3 µmol/mol/K	0.004	u_{gt}	0.01	0.0001		
6	Sensitivity coefficient of surrounding temperature at 8h-limit value	≤ 0.3 µmol/mol/K	0.067	u_{st}	0.17	0.0304		
7	Sensitivity coefficient of electrical voltage at 8h-limit value	≤ 0.3 µmol/mol/V	-0.007	u_V	-0.06	0.0041		
8a	Interferent H ₂ O with 21 mmol/mol	≤ 1.0 µmol/mol (Zero)	0.181	u_{H2O}	0.01	0.0001		
8b	Interferent CO ₂ with 500 µmol/mol	≤ 1.0 µmol/mol (Span)	-0.009	$u_{int, pos}$	0.017	0.0000		
8c	Interferent NO with 1 µmol/mol	≤ 0.5 µmol/mol (Span)	-0.086	$u_{int, neg}$	-0.009	0.0000		
8d	Interferent N ₂ O with 50 nmol/mol	≤ 0.5 µmol/mol (Zero)	-0.009	$u_{int, neg}$	0.017	0.0000		
9	Averaging effect	≤ 7.0% of measured value	-1.100	u_{av}	-0.05	0.0030		
18	Difference sample/calibration port	≤ 1.0%	0.000	u_{asc}	0.00	0.0000		
21	Uncertainty of test gas	≤ 3.0%	2.000	u_{cg}	0.09	0.0074		
	Combined standard uncertainty		u_c		0.2176	µmol/mol		
	Expanded uncertainty		U		0.4353	µmol/mol		
	Relative expanded uncertainty		W		5.05	%		
	Maximum allowed expanded uncertainty		W_{eq}		15	%		

Expanded uncertainty based on the results of the laboratory and field testing of System 1

Measured component:	Measuring device: Horiba APMA 370 CO	Serial-No.:	SN 10031		
		8h-limit value: 8.62 µmol/mol			
No.	Performance characteristic	Performance criterion	Result	Partial uncertainty	Square of partial uncertainty
1	Repeatability standard deviation at zero	≤ 0.3 µmol/mol	0.035	$u_{t,z}$ 0.01	0.0001
2	Repeatability standard deviation at 8h-limit value	≤ 0.4 µmol/mol	0.064	u_r not considered, as $u_r = 0 < u_{t,f}$	-
3	"lack of fit" at 8h-limit value	≤ 4.0% of measured value	0.700	u_t 0.03	0.0012
4	Sensitivity coefficient of sample gas pressure at 8h-limit value	≤ 0.7 µmol/mol/kPa	0.006	u_{gp} 0.02	0.0002
5	Sensitivity coefficient of sample gas temperature at 8h-limit value	≤ 0.3 µmol/mol/K	0.005	u_{gt} 0.01	0.0002
6	Sensitivity coefficient of surrounding temperature at 8h-limit value	≤ 0.3 µmol/mol/K	0.077	u_{st} 0.20	0.0388
7	Sensitivity coefficient of electrical voltage at 8h-limit value	≤ 0.3 µmol/mol/V	0.006	u_v 0.06	0.0030
8a	Interferent H ₂ O with 21 nmol/mol	1.0 µmol/mol (Zero) 1.0 µmol/mol (Span)	-0.009 0.233	u_{H2O} 0.02	0.0003
8b	Interferent CO ₂ with 500 µmol/mol	0.5 µmol/mol (Zero) 0.5 µmol/mol (Span)	0.017 -0.164	$u_{CO2, pos}$ $u_{CO2, neg}$	
8c	Interferent NO with 1 µmol/mol	0.5 µmol/mol (Zero) 0.5 µmol/mol (Span)	-0.026 -0.353	or 0.10	0.0105
8d	Interferent N ₂ O with 50 nmol/mol	0.5 µmol/mol (Span)	0.009 -0.164		
9	Averaging effect	≤ 7.0% of measured value	1.250	u_{av} 0.06	0.0039
10	Reproducibility standard deviation under field conditions	≤ 5.0% of average over 3 months	3.420	u_{rf} 0.29	0.0889
11	Long term drift at zero level	≤ 0.5 µmol/mol	-0.172	$u_{dl,z}$ -0.10	0.0099
12	Long term drift at span level	≤ 5.0% of max. of certification range	-1.750	$u_{dl,8h}$ -0.09	0.0076
18	Difference sample/calibration port	≤ 1.0%	0.000	u_{sc} 0.00	0.0000
21	Uncertainty of test gas	≤ 3.0%	2.000	u_{tg} 0.09	0.0074
		Combined standard uncertainty Expanded uncertainty Relative expanded uncertainty Maximum allowed expanded uncertainty	u_c u W W_{eq}	$u_{c, rel}$ u_{rel} W_{rel} $W_{eq, max}$	0.4136 0.8271 9.60 15
			%	%	%

Expanded uncertainty based on the results of the laboratory and field testing of System 2

Measured device:	Horiba APIMA 370	Measured component:	CO	Serial-No.:	SN 10032	8h-limit value:	8.62
Performance characteristic							
No.		Performance criterion	Result	Partial uncertainty	Square of partial uncertainty		
1	Repeatability standard deviation at zero	≤ 0.3 µmol/mol	0.028	U _{t,z}	0.01	0.0000	
2	Repeatability standard deviation at 8h-limit value	≤ 0.4 µmol/mol	0.070	U _r	not considered, as ur = 0 < ur,f	-	
3	"lack of fit" at 8h-limit value	≤ 4.0% of measured value	0.800	U _t	0.04	0.0016	
4	Sensitivity coefficient of sample gas pressure at 8h-limit value	≤ 0.7 µmol/mol/kPa	0.009	U _{gp}	0.02	0.0005	
5	Sensitivity coefficient of sample gas temperature at 8h-limit value	≤ 0.3 µmol/mol/K	0.004	U _{gt}	0.01	0.0001	
6	Sensitivity coefficient of surrounding temperature at 8h-limit value	≤ 0.3 µmol/mol/K	0.067	U _{st}	0.17	0.0304	
7	Sensitivity coefficient of electrical voltage at 8h-limit value	≤ 0.3 µmol/mol/V	-0.007	U _v	-0.06	0.0041	
8a	Interferent H ₂ O with 21 nmol/mol	≤ 1.0 µmol/mol (Zero)	-0.009	U _{H2O}	0.01	0.0001	
8b	Interferent CO ₂ with 500 µmol/mol	≤ 0.5 µmol/mol (Zero)	0.017	U _{int, pos}			
8c	Interferent NO with 1 µmol/mol	≤ 0.5 µmol/mol (Span)	-0.009	U _{NO}	0.01	0.0000	
8d	Interferent N ₂ O with 50 nmol/mol	≤ 0.5 µmol/mol (Zero)	0.017	U _{N2O}			
9	Averaging effect	≤ 7.0% of measured value	-1.100	U _{av}	-0.05	0.0030	
10	Reproducibility standard deviation under field conditions	≤ 5.0% of average over 3 months	3.420	U _{t,f}	0.29	0.0869	
11	Long term drift at zero level	≤ 0.5 µmol/mol	-0.164	U _{t,z}	-0.09	0.0090	
12	Long term drift at span level	≤ 5.0% of max. of certification range	-1.610	U _{t,8h}	-0.08	0.0064	
18	Difference sample/calibration port	≤ 1.0%	0.000	U _{sc}	0.00	0.0000	
21	Uncertainty of test gas	≤ 3.0%	2.000	U _{cg}	0.09	0.0074	
Combined standard uncertainty							
Expanded uncertainty							
Relative expanded uncertainty							
Maximum allowed expanded uncertainty							
				W _{eq}	15	%	