

CONFIRMATION

Notification: 0000043107_03_03
on changes according to EN 15267 regarding certificate 0000043107_03 dated 20 March 2024

AMS designation: APDA-372 / APDA-372 E for PM₁₀ and PM_{2.5}

Manufacturer: HORIBA Europe GmbH
Hans-Mess-Str. 6
61440 Oberursel /Ts.
Germany

German Federal Environmental Agency (UBA)

**Announcement regarding the uniform practice in
monitoring emissions and ambient air
dated 27 August 2025
Publication: BAnz AT 31.10.2025 B5**

**IV. Notifications as regards the uniform practice in monitoring
emissions and ambient air:**

**41 Notification as regards Federal Environment Agency (UBA) notices
of 25 February 2015 (BAnz AT 02.04.2015 B5, chapter III number 3.1) and
of 2 April 2025 (BAnz AT 19.05.2025 B3, chapter IV notification 83)**

The current software version of the APDA-372 and APDA-372 E measuring devices
for PM₁₀ and PM_{2.5} from HORIBA Europe GmbH is:


100544.0014.0001.0001.0011


This includes version 100543.0014.0001.0001.0011.

The flow sensor (Siargo FS4008-10-O6-CV-A) for the APDA-372 and APDA-372 E
measuring devices for PM₁₀ and PM_{2.5} is being replaced by the successor model (Siargo
FS4308-10-O6-BV-A-0.5/4.5V).

Statement issued by TÜV Rheinland Energy & Environment GmbH dated 20 May 2025

TÜV Rheinland Energy & Environment GmbH
Cologne, 21 November 2025


i. V. Dipl.-Ing. Guido Baum


i. A. Dipl.-Ing. Carsten Röllig

www.umwelt-tuv.eu
qal1-info@tuv.com
Tel. +49 221 806-5200

TÜV Rheinland Energy & Environment GmbH
Am Grauen Stein
51105 Cologne

Test institute accredited to EN ISO/IEC 17025 by DAkkS (German Accreditation Body).
This accreditation is limited to the accreditation scope defined in the enclosure to the certificate D-PL-11120-02-00.