

CERTIFICATE

of Product Conformity (QAL1)

Certificate No.: 0000040211_02

AMS designation: K-BAR 2000B for velocity

Manufacturer: Kurz Instruments, Inc.
2411 Garden Road
Monterey
CA 93940
USA

Test Laboratory: TÜV Rheinland Energy GmbH

**This is to certify that the AMS has been tested
and found to comply with the standards
EN 15267-1 (2009), EN 15267-2 (2009), EN 15267-3 (2007),
EN 16911-2 (2013) and EN 14181 (2004).**

Certification is awarded in respect of the conditions stated in this certificate
(this certificate contains 7 pages).
The present certificate replaces certificate 0000040211_01 of 01 April 2019.



Suitability Tested
EN 15267
QAL1 Certified
Regular Surveillance

www.tuv.com
ID 0000040211

Publication in the German Federal Gazette
(BAnz) of 01 April 2014


This certificate will expire on:
30 June 2025

German Federal Environment Agency
Dessau, 01 July 2020

TÜV Rheinland Energy GmbH
Cologne, 30 June 2020



Dr. Marcel Langner
Head of Section II 4.1



ppa. Dr. Peter Wilbring

www.umwelt-tuv.eu
tre@umwelt-tuv.eu
Phone: + 49 221 806-5200

TÜV Rheinland Energy GmbH
Am Grauen Stein
51105 Köln

Test institute accredited to EN ISO/IEC 17025 by DAkkS (German Accreditation Body).
This accreditation is limited to the accreditation scope defined in the enclosure to certificate D-PL-11120-02-00.

Test Report:	936/21219690/A dated 10 October 2013
Initial certification:	01 April 2014
Expiry date:	30 June 2025
Certificate:	Renewal (of previous certificate 0000040211_01 dated 01 April 2019 valid until 30 June 2020)
Publication:	BAnz AT 01.04.2014 B12, chapter II number 2.2

Approved application

The tested AMS is suitable for use at combustion plants according to Directive 2010/75/EU, chapter III (13th BImSchV), chapter IV (17th BImSchV), 30th BImSchV, plants in compliance with TA Luft and plants according to the 27th BImSchV. The measured ranges have been selected so as to ensure as broad a field of application as possible.

The suitability of the AMS for this application was assessed on the basis of a laboratory test and a three-months field test at a municipal waste incineration plant.

The AMS is approved for an ambient temperature range of -20 °C to +50 °C.

The notification of suitability of the AMS, performance testing and the uncertainty calculation have been effected on the basis of the regulations applicable at the time of testing. As changes in legal provisions are possible, any potential user should ensure that this AMS is suitable for monitoring the velocities relevant to the application.

Any potential user should ensure, in consultation with the manufacturer, that this AMS is suitable for the intended purpose.

Basis of the certification

This certification is based on:

- Test report no. 936/21219690/A dated 10 October 2013 issued by TÜV Rheinland Energie und Umwelt GmbH
- Suitability announced by the German Federal Environment Agency (UBA) as the relevant body
- The ongoing surveillance of the product and the manufacturing process

Publication in the German Federal Gazette: BAnz AT 01.04.2014 B12, chapter II number 2.2,
UBA announcement dated 27 February 2014:

AMS designation:

K-BAR 2000B for velocity

Manufacturer:

Kurz Instruments, Inc., Monterey, USA

Field of application:

For plants requiring official approval and for plants according to the 27th BImSchV

Measuring range during performance testing:

Component	Certification range	Unit
Velocity	0–30	m/s

Software version:

MFT-B VER 2.08

Restriction:

The measuring system may only be employed if the temperature does not fall below dew point.

Notes:

1. The maintenance interval is four weeks.
2. The measuring system may be used at exhaust gas temperatures of up to 500 °C.

Test Report:

TÜV Rheinland Energie und Umwelt GmbH, Cologne
Report no.: 936/21219690/A dated 10 October 2013

Publication in the German Federal Gazette: BAnz AT 24.03.2020 B7, chapter IV
notification 57, UBA announcement dated 24 February 2020:

**57 Notification as regards Federal Environment Agency (UBA) notices
of 27 February 2014 (BAnz AT 01.04.2014 B12, chapter II number 2.2)**

In the context of continuous product development, Kurz Instruments introduced the following changes to their K-BAR 2000B measuring system for velocity:

The latest software version of the measuring system now is:

MFT-B VER 2.15

Moreover, the following software version are approved for this instrument version:

MFT-B VER 2.08, MFT-B VER 2.09, MFT-B VER 2.10, MFT-B VER 2.11,
MFT-B VER 2.12, MFT-B VER 2.13 and MFT-B VER 2.14.

Statement issued by TÜV Rheinland Energy GmbH dated 17 September 2019

Certified product

This certification applies to automated measurement systems conforming to the following description:

The measuring system K-BAR 2000B for monitoring exhaust gas velocity consists of one or more sensor probe rods in which one or more sensor elements are fitted (the tested measuring system is equipped with 2 built-in sensor elements) that measure velocity according to the principle of thermal anemometry. To do so, an electrically heated resistance temperature detector (RTD) is used which maintains a constant temperature difference to the surrounding sample gas (temperature is measured with a second RTD). The measurement signal produced is the electricity required to maintain a constant temperature difference between the heated RTD and the sample gas.

An electronic analysis component is fitted directly on the probe rod and is connected to the external analysis and control electronics Adam 155B. The Adam 155B component calculates and provides the mean value of the individual elements. The parameters of the entire measuring system can also be controlled using the keyboard and display.

A control cycle for zero and span point control can be initiated via an external Siemens Logo PC. No proper reference point checks were carried out, but the evaluation electronics of the sensor elements were subjected to testing.

General remarks

This certificate is based upon the equipment tested. The manufacturer is responsible for ensuring that on-going production complies with the requirements of the EN 15267. The manufacturer is required to maintain an approved quality management system controlling the manufacturing process for the certified product. Both the product and the quality management systems shall be subject to regular surveillance.

If a product of the current production does not conform to the certified product, TÜV Rheinland Energy GmbH must be notified at the address given on page 1.

A certification mark with an ID-Number that is specific to the certified product is presented on page 1 of this certificate.

This document as well as the certification mark remains property of TÜV Rheinland Energy GmbH. Upon revocation of the publication the certificate loses its validity. After the expiration of the certificate and on request of TÜV Rheinland Energy GmbH this document shall be returned and the certificate mark must no longer be used.

The relevant version of this certificate and its expiration date are also accessible on the internet at qal1.de.

Document history

Certification of the K-BAR 2000B measuring system is based on the documents listed below and the regular, continuous surveillance of the manufacturer's quality management system.

Initial certification according to EN 15267

Certificate no. 0000040211: 29 April 2014
Expiry date of the certificate: 31 March 2019
Test report no.: 936/21219690/A dated 10 October 2013
TÜV Rheinland Energie und Umwelt GmbH, Cologne
Publication: BAnz AT 01.04.2014 B12, chapter II number 2.2
UBA announcement dated 27 February 2014

Renewal of the certificate in accordance with EN 15267

Certificate no. 0000040211_01: 01 April 2019
Expiry date of the certificate: 30 June 2020

Notifications in accordance with EN 15267

Statement issued by TÜV Rheinland Energie und Umwelt GmbH dated 17 September 2019
Publication: BAnz AT 24.03.2020 B7, chapter IV notification 57,
UBA announcement dated 24 February 2020
(software updates)

Renewal of the certificate

Certificate no. 0000040211_02: 01 July 2020
Expiry date of the certificate: 30 June 2025

Calculation of overall uncertainty according to EN 14181 and EN 15267-3

Measuring system

Manufacturer	Kurz Instruments Inc.
AMS designation	K-Bar 2000B
Serial number of units under test	1294A / 1294B
Measuring principle	Thermal convection

Test report

Test laboratory	TÜV Rheinland
Date of report	2013-10-10

Measured component

Certification range	Velocity 0 - 30 m/s
---------------------	------------------------

Calculation of the combined standard uncertainty

Tested parameter

			u^2
Standard deviation from paired measurements under field conditions *	u_D	0.215 m/s	0.046 (m/s) ²
Lack of fit	u_{lof}	-0.230 m/s	0.053 (m/s) ²
Zero drift from field test	u_{dz}	0.035 m/s	0.001 (m/s) ²
Span drift from field test	u_{ds}	0.052 m/s	0.003 (m/s) ²
Influence of ambient temperature at span	u_t	0.115 m/s	0.013 (m/s) ²
Influence of supply voltage	u_v	0.012 m/s	0.000 (m/s) ²
Uncertainty of reference material at 70% of certification range	u_{rm}	0.242 m/s	0.059 (m/s) ²

* The larger value is used :
"Repeatability standard deviation at span" or
"Standard deviation from paired measurements under field conditions"

Combined standard uncertainty (u_c)	$u_c = \sqrt{\sum (u_{max,j})^2}$	0.42 m/s
Total expanded uncertainty	$U = u_c * k = u_c * 1.96$	0.82 m/s

Relative total expanded uncertainty

Requirement of 2010/75/EU	U in % of the range 30 m/s	2.7
Requirement of EN 15267-3	U in % of the range 30 m/s	10.0 **
	U in % of the range 30 m/s	7.5

** For this component no requirements in the EC-directives 2010/75/EU on industrial emissions are given.
The chosen value is recommended by the certification body.