

CONFIRMATION

Notification: 0000026912_05_02
on changes according to EN 15267 regarding certificate 0000026912_05 dated 20 March 2024

AMS designation: BAM 1020 with PM_{2,5}-pre selector for PM_{2,5}

Manufacturer: Met One Instruments Inc.
1600, Washington Blvd.
Grants Pass, Oregon 97526,
USA

German Federal Environmental Agency (UBA)

**Announcement regarding the uniform practice in
monitoring emissions and ambient air
dated 2 April 2025
Publication: BAnz AT 19.05.2025 B3**

**IV. Notifications as regards the uniform practice in monitoring
emissions and ambient air:**

**76 Notification as regards Federal Environment Agency (UBA) notices
of 12 July 2010 (BAnz. p. 2597, chapter II number 1.1) and
of 19 March 2024 (BAnz AT 10.05.2024 B7, chapter V notification 56)**

The current software versions of the BAM-1020 measuring system
for particulate matter PM_{2,5} from Met One Instruments, Inc. are as follows:

Version without touch-sensitive screen: 3236-05 3.14.4


Version with touch-sensitive screen (old screen version): 3236-77 V5.2.0


Version with touch-sensitive screen (new screen version): R9.4.5

This includes the versions: R9.2.3, R9.3.0, R9.3.1, R9.3.2, R9.4.3 and R9.4.4

Statement issued by TÜV Rheinland Energy & Environment GmbH
dated 26 September 2024

TÜV Rheinland Energy & Environment GmbH
Cologne, 20 May 2025


i. V. Dipl.-Ing. Guido Baum


i. A. Dipl.-Ing. Carsten Röllig

www.umwelt-tuv.eu
TRE@umwelt-tuv.eu
Tel. +49 221 806-5200

TÜV Rheinland Energy & Environment GmbH
Am Grauen Stein
51105 Cologne

Test institute accredited to EN ISO/IEC 17025 by DAkkS (German Accreditation Body).
This accreditation is limited to the accreditation scope defined in the enclosure to the certificate D-PL-11120-02-00.