

CONFIRMATION

Notification: 0000081155_00_01
on changes according to EN 15267 regarding certificate 0000081155_00 dated 25 April 2023

AMS designation: Fidas Smart 100 / 100E for PM₁₀ and PM_{2,5}

Manufacturer: PALAS GmbH
Greschbachstrasse 3b
76229 Karlsruhe
Germany

German Federal Environmental Agency (UBA)

**Announcement regarding the uniform practice in
monitoring emissions and ambient air
dated 05 July 2023
Publication: BAnz AT 02.08.2023 B7**

**III. Notifications as regards the uniform practice in monitoring
emissions and ambient air:**

**37 Notification as regards Federal Environment Agency (UBA) notice
of 21 February 2023 (BAnz AT 20.03.2023 B6, chapter III number 1.1)**

The current software version for the Fidas Smart 100 / Fidas Smart 100 E measuring system for PM₁₀ and PM_{2,5} from Palas GmbH is:

1.0.12


In addition to this version number, the following intermediate version is also valid:


1.0.11

The memory capacity of the integrated PC has been extended from 32 GB to 128 GB. The option of power supply via the Ethernet socket has been removed and the mounting position of the Ethernet socket has been rotated by 180°. The HDMI socket has been replaced by an AU6 socket. The transfer speed of the integrated USB hub has been increased from 12 MBit to 480 Mbit.

Statement issued by TÜV Rheinland Energy GmbH dated 31 March 2023

TÜV Rheinland Energy GmbH
Cologne, 02 August 2023


i. V. Dipl.-Ing. Guido Baum


i. A. Dipl.-Ing. Carsten Röllig

www.umwelt-tuv.eu
TRE@umwelt-tuv.eu
Tel. +49 221 806-5200

TÜV Rheinland Energy GmbH
Am Grauen Stein
51105 Cologne

Test institute accredited to EN ISO/IEC 17025 by DAkkS (German Accreditation Body).
This accreditation is limited to the accreditation scope defined in the enclosure to the certificate D-PL-11120-02-00.