

CONFIRMATION

Notification: 0000081155_00_02
on changes according to EN 15267 regarding certificate 0000081155_00 dated 25 April 2023

AMS designation: Fidas Smart 100 / 100E for PM₁₀ and PM_{2,5}

Manufacturer: PALAS GmbH
Greschbachstrasse 3b
76229 Karlsruhe
Germany

German Federal Environmental Agency (UBA)

**Announcement regarding the uniform practice in
monitoring emissions and ambient air
dated 19 March 2024
Publication: BAnz AT 10.05.2024 B7**

**V. Notifications as regards the uniform practice in monitoring
emissions and ambient air:**

**58 Notification as regards Federal Environment Agency (UBA) notices
of 21 February 2023 (BAnz AT 20.03.2023 B6, chapter III number 1.1) and
of 5 July 2023 (BAnz AT 02.08.2023 B7, chapter III notification 37)**

The current software version for the Fidas® Smart 100 and Fidas® Smart 100 E
measuring systems for PM₁₀ and PM_{2,5} from Palas GmbH is: 1.0.14


In addition to this version number, the following interim version is also valid: 1.0.13


The data stick previously installed in the measuring system is no longer required with
immediate effect.

In the Fidas® Smart 100 E variant, the cable routing in the transition from the IADS to
the casing tube has been adapted.

Statement issued by TÜV Rheinland Energy GmbH dated 01 September 2023

TÜV Rheinland Energy & Environment GmbH
Cologne, 24 May 2024


i. V. Dipl.-Ing. Guido Baum


i. A. Dipl.-Ing. Carsten Röllig

www.umwelt-tuv.eu
TRE@umwelt-tuv.eu
Tel. +49 221 806-5200

TÜV Rheinland Energy & Environment GmbH
Am Grauen Stein
51105 Cologne

Test institute accredited to EN ISO/IEC 17025 by DAkkS (German Accreditation Body).
This accreditation is limited to the accreditation scope defined in the enclosure to the certificate D-PL-11120-02-00.