

CONFIRMATION

Notification: 0000035013_03_04
on changes according to EN 15267 regarding certificate 0000035013_03 dated 16 February 2022

AMS designation: QAL 182 WS for dust

Manufacturer: ENVEA UK Ltd. (PCME Ltd.)
ENVEA House, Rose & Crown Road
Swavesey / Cambridge CB24 4RB
United Kingdom

German Federal Environmental Agency (UBA)

**Announcement regarding the uniform practice in
monitoring emissions and ambient air
dated 2 April 2025
Publication: BAnz AT 19.05.2025 B3**

**IV. Notifications as regards the uniform practice in monitoring
emissions and ambient air:**

**50 Notification as regards Federal Environment Agency (UBA) notices
of 23 February 2012 (BAnz. p. 920, chapter I number 1.2) and
of 19 March 2024 (BAnz AT 10.05.2024 B7, chapter V notification 13)**

The PCME QAL 182 WS measuring system for dust from ENVEA UK Ltd. now has the device name QAL 182 WS.


The NetController control unit has a revised main circuit board. The new software version for the NetController control unit of the QAL 182 WS measuring system is as follows:

Controller Software 1.09

For the external reference filters (audit units) for the measuring system, which can be used for QAL3 activities, measuring system settings and linearity tests, Schott NEXTREMA glass-ceramic can now be used in addition to the OPALIKA material originally used.

Statement issued by TÜV Rheinland Energy & Environment GmbH
dated 20 December 2024

TÜV Rheinland Energy & Environment GmbH
Cologne, 20 May 2025



i. V. Dipl.-Ing. Guido Baum



i. A. Dipl.-Ing. Carsten Röllig

www.umwelt-tuv.eu
TRE@umwelt-tuv.eu
Tel. +49 221 806-5200

TÜV Rheinland Energy & Environment GmbH
Am Grauen Stein
51105 Cologne

Test institute accredited to EN ISO/IEC 17025 by DAkkS (German Accreditation Body).
This accreditation is limited to the accreditation scope defined in the enclosure to the certificate D-PL-11120-02-00.