

CERTIFICATE

of Product Conformity (QAL1)

Certificate No: 0000028729_04

Certified AMS: GM 700-2 for HF

Manufacturer: Endress+Hauser SICK GmbH + Co. KG
Bergener Ring 27
01458 Ottendorf-Okrilla
Germany

Test Institute: TÜV Rheinland Energy & Environment GmbH

**This is to certify that the AMS has been tested
and found to comply with the standards
EN 15267-1 (2009), EN 15267-2 (2023), EN 15267-3 (2007),
as well as EN 14181 (2014).**

Certification is awarded in respect of the conditions stated in this certificate
(this certificate contains 10 pages).

The present certificate replaces certificate 0000028729_03 dated 25 January 2021.




Publication in the German Federal Gazette
(BAnz) of 5 August 2014


This certificate will expire on:
25 January 2031

German Federal Environment Agency

TÜV Rheinland
Energy & Environment GmbH
Cologne, 21 January 2026

Dessau, 22 January 2026


Dr. Marcel Langner
Head of Section II 4


ppa. Dr. Peter Wilbring

www.umwelt-tuv.eu qal1-info@tuv.com Tel. + 49 221 806-5200	TÜV Rheinland Energy & Environment GmbH Am Grauen Stein 51105 Köln
Test institute accredited to EN ISO/IEC 17025 by DAkkS (German Accreditation Body). This accreditation is limited to the accreditation scope defined in the enclosure to the certificate D-PL-11120-02-00.	

Test report: 936/21210058/B dated 2 April 2014**Initial certification:** 26 January 2011**Expiry date:** 25 January 2031**Certificate:** Renewal (of previous certificate 0000028729_03 of
25 January 2021 valid until 25 January 2026)**Publication:** BAnz AT 05.08.2014 B11, chapter I No. 2.1

Approved application

The tested AMS is suitable for use at plants according to Directive 2010/75/EC, chapter III (combustion plants / 13th BImSchV:2013), Directive 2010/75/EC, chapter IV (waste incineration plants / 17th BImSchV:2013), 30th BImSchV:2019, Directive 2015/2193/EC (44th BImSchV:2022), TA Luft:2002 and 27th BImSchV:2013. The measured ranges have been selected so as to ensure as broad a field of application as possible.

The suitability of the AMS for this application was assessed on the basis of a laboratory test and a twelve month field test at a tunnel kiln for ceramics roof tiles.

The AMS is approved for an ambient temperature range of -20 °C to 50 °C.

The notification of suitability of the AMS, performance testing and the uncertainty calculation have been effected on the basis of the regulations applicable at the time of testing. As changes in legal provisions are possible, any potential user should ensure that this AMS is suitable for monitoring the emission limit values relevant to the application.

Any potential user should ensure, in consultation with the manufacturer, that this AMS is suitable for the installation at which it will be installed.

Note

The legal regulations mentioned correspond to the current state of legislation during certification. Each user should, if necessary, in consultation with the competent authority, ensure that this AMS meets the legal requirements for the intended use. In addition, it cannot be ruled out that legal regulations governing the use of a measuring device for emission monitoring may change during the lifetime of the certificate.

Basis of the certification

This certification is based on:

- Test report 936/21210058/B dated 2 April 2014 of TÜV Rheinland Energie und Umwelt GmbH
- Suitability announced by the German Federal Environment Agency (UBA) as the relevant body
- The ongoing surveillance of the product and the manufacturing process

Publication in the German Federal Gazette: BAnz AT 05.08.2014 B11, chapter I No. 2.1,
Announcement by UBA dated 17 July 2014:

AMS designation:

GM700-2 for HF

Manufacturer:

SICK AG, Reute

Field of application:

For plants requiring official approval and for plants according to the 27th BImSchV

Measuring ranges during the performance test:

Component	Certification range	supplementary measuring ranges	Unit
HF	0 – 5	0 – 25	mg/m ³

Software version:

9105060 VA24 (sampling head)

9100821 WN42 (evaluation unit)

9091948 WJ24 (purge air)

Restrictions:

The measuring system did not meet the requirement for total uncertainty as defined in EN 15267-3.

Notes:

1. Wet test gases should be used for testing HF.
2. The maintenance interval is six months.
3. If the ambient temperature range is > 50 °C, the parameterization of the heating element for the sender-/receiver unit must be adjusted.
4. Supplementary test (extension of the maintenance interval) as regards Federal Environment Agency (UBA) notice of 10 January 2011 (BAnz. p. 294, chapter I number 2.1) and of 3 July 2013 (BAnz AT 23.07.2013 B4, chapter V notification 12 [number 1]).

Test Institute: TÜV Rheinland Energie und Umwelt GmbH, Cologne

Report No.: 936/21210058/B dated 2 April 2014

Publication in the German Federal Gazette: BAnz AT 26.08.2015 B4, Chap. V notification 7, Announcement by UBA dated 22 July 2015:

7 Notification as regards Federal Environment Agency (UBA) notice of 17 July 2014 (BAnz AT 05.08.2014 B11, chapter I number 2.1)

The latest software versions of the GM700-2 measuring system for HF manufactured by SICK AG are:

9105060 YEK0 (sampling head)
9100821 WN42 (evaluation unit)
9091948 WJ24 (purge air)

Statement of TÜV Rheinland Energie und Umwelt GmbH dated 27 March 2015

Publication in the German Federal Gazette: BAnz AT 14.03.2016 B7, Chap. V notification 29, Announcement by UBA dated 18 February 2016:

29 Notification as regards Federal Environment Agency (UBA) notices of 17 July 2014 (BAnz AT 05.08.2014 B11, chapter I number 2.1) and of 22 July 2015 (BAnz AT 26.08.2015 B4, chapter V notification 7)

The latest hardware and software versions of the GM700-2 measuring system for HF manufactured by SICK AG are:

9105060 YEK0 (sampling head)
9100821 YKI0 (evaluation unit)
9091948 WJ24 (purge air)

Statement of TÜV Rheinland Energie und Umwelt GmbH dated 21 October 2015

Publication in the German Federal Gazette: BAnz AT 31.07.2017 B12, Chap. II notification 19, Announcement by UBA dated 13 July 2017:

19 Notification as regards Federal Environment Agency (UBA) notices of 17 July 2014 (BAnz AT 05.08.2014 B11, chapter I number 2.1) and of 27 February 2016 (BAnz AT 14.03.2016 B7, chapter V notification 29)

The GM700-2 measuring system for HF manufactured by SICK AG has been equipped with a new housing for the sampling head. The housing is now 10 mm wider and equipped with a handle and quick release fasteners. The service interface is now accessible from the outside and the insulation has been optimised.

The latest software versions of the GM700-2 measuring system for HF manufactured by SICK AG are:

9105060 YVB8 (sampling head)
9100821 YKI0 (evaluation unit)
9091948 WJ24 (purge air)

Statement issued by TÜV Rheinland Energy GmbH dated 8 March 2017

Publication in the German Federal Gazette: BAnz AT 26.03.2018 B8, Chap. IV notification 2, Announcement by UBA dated 21 February 2018:

2 Correction as regards Federal Environment Agency (UBA) notices of 27 July 2016 (BAnz AT 14.03.2016 B7, chapter V 29th notification) and of 13 February 2017 (BAnz AT 31.07.2017 B7, chapter II notification 19)

The Notifications cited above indicate incorrect software versions for the GM700-2 manufactured by SICK AG. The correct software versions of the measuring system are as follows:

9105060 YVB8 (sampling head)
9100821 WN42 (evaluation unit)
9091948 WJ24 (purge air)

The software version 9100821 YKI0 published by mistake is irrelevant.

Statement issued by TÜV Rheinland Energy GmbH dated 8 December 2017

Publication in the German Federal Gazette: BAnz AT 19.05.2025 B3, Chap. IV
notification 16, Announcement by UBA dated 2 April 2025:

**16 Notification as regards Federal Environment Agency (UBA) notices
of 17 July 2014 (BAnz AT 05.08.2014 B11, chapter I number 2.1) and
of 21 February 2018 (BAnz AT 26.03.2018 B8, chapter IV notification 2)**

Change of the company name of the manufacturer of the GM700-2 measuring system for HF of the previous company SICK AG.

The analysis division of Sick AG and Sick Engineering GmbH (flow and dust measurement technology) have entered into a joint venture with Endress+Hauser. The companies will operate under the new company name "Endress+Hauser SICK GmbH + Co. KG" from 9 December 2024. As a result, the QAL1-certified products will be sold under this joint company name in future.

The joint venture company address is
Endress+Hauser SICK GmbH + Co. KG
Bergener Ring 27
01458 Ottendorf-Okrilla, Germany

Statement issued by TÜV Rheinland Energy & Environment GmbH
dated 11 October 2024

Certified product

This certificate applies to automated measurement systems conforming to the following description:

The GM700-2 measuring system is an in-situ laser system for the determination of HF.

A laser which has been developed specifically for gas analysis is operated as light source of the GM700-2 measuring system. Precise stabilisation of the wavelength and temperature is ensured by a Peltier element and a temperature sensor built into the housing of the laser diode.

The laser beam transmitted by the transmitter-receiver unit passes through the active measuring path and hits the detector at the other end of the gas duct, where it is reflected back to the transmitter-receiver unit. There, the light is focused onto a photo diode via a light collector.

The light of the laser diode shines through the sample gas and then detected by a photo diode. The wavelength of the laser diode is tuned to a single absorption line of the test gas component. A signal evaluation unit provides the size of the absorption line which is required for the calculation of the gas concentration. This method is called Tunable Diode Laser Spectroscopy (TDLS) or Tunable Diode Laser Absorption Spectroscopy (TDLAS).

The GM700-2 measuring system is equipped with a closed reference cuvette in order to stabilise the wavelength of the laser.

The tested measuring system consists of the following parts:

- Sender-receiver unit (SR unit) containing the optical and electronical components of the measuring system.
- Triple reflector
- Purge air attachments for SR unit and reflector
- Purge air unit
- Evaluation unit
 - Output of measured values, calculated data and operation states
 - Communication with the peripheral equipment
 - Output of error messages and other status signals
 - Controlling of automatic test functions and access during service (diagnosis)
- Probe for temperature and pressure measurement
- Zero path with GMK10 test cell

The certification range is 5 mg/m³ * m. The length of the measuring path which has been used during the test was 1 m.

General notes

This certificate is based upon the equipment tested. The manufacturer is responsible for ensuring that on-going production complies with the requirements of the EN 15267. The manufacturer is required to maintain an approved quality management system controlling the manufacture of the certified product. Both the product and the quality management systems shall be subject to regular surveillance.

If a product of the current production does not conform to the certified product, TÜV Rheinland Energy & Environment GmbH must be notified at the address given on page 1.

A certification mark with an ID-Number that is specific to the certified product is presented on page 1 of this certificate. This certification mark may be applied to the product or used in advertising materials for the certified product.

This document as well as the certification mark remains property of TÜV Rheinland Energy & Environment GmbH. With revocation of the publication the certificate loses its validity. After the expiration of the certificate and on requests of the TÜV Rheinland Energy & Environment GmbH this document shall be returned and the certificate mark must not be employed anymore.

The relevant version of this certificate and its expiration is also accessible on the internet:
qal1.de

History of documents

Certification of GM 700-2 is based on the documents listed below and the regular, continuous monitoring of the Quality Management System of the manufacturer:

Initial certification according to EN 15267

Certificate No. 0000028729_00: 9 February 2011
Expiry date of the certificate: 25 January 2016
Test report: 936/21210058/A dated 30 September 2010
TÜV Rheinland Energie und Umwelt GmbH
Publication: BAnz. 26 January 2011, No. 14, p. 294, chapter I number 2.1
UBA announcement dated 10 January 2011

Notifications

Statement issued by TÜV Rheinland Energie und Umwelt GmbH dated 14 March 2012
Publication: BAnz AT 20.07.2012 B11, chapter IV notification 15
UBA announcement dated 6 July 2012
(Software changes)

Statement issued by TÜV Rheinland Energie und Umwelt GmbH dated 6 October 2012
Publication: BAnz AT 05.03.2013 B10, chapter V notification 26
UBA announcement dated 12 February 2013
(Software changes)

Statement issued by TÜV Rheinland Energie und Umwelt GmbH dated 25 March 2013

Publication: BAuz AT 23.07.2013 B4, chapter V notification 12

UBA announcement dated 3 July 2013

(new manufacturer name)

Supplementary testing according to EN 15267

Certificate No. 0000028729_01: 9 September 2014

Expiry date of the certificate: 25 January 2016

Test report: 936/21210058/B dated 2 April 2014

TÜV Rheinland Energie und Umwelt GmbH

Publication: BAuz AT 05.08.2014 B11, chapter I number 2.1

UBA announcement dated 17 July 2014

Notifications

Statement issued by TÜV Rheinland Energie und Umwelt GmbH dated 27 March 2015

Publication: BAuz AT 26.08.2015 B4, chapter V notification 7

UBA announcement dated 22 July 2015

(Software changes)

Renewal of certificates

Certificate No. 0000028729_02: 21 January 2016

Expiry date of the certificate: 25 January 2021

Notifications

Statement issued by TÜV Rheinland Energie und Umwelt GmbH dated 21 October 2015

Publication: BAuz AT 14.03.2016 B7, chapter V notification 29

UBA announcement dated 18 February 2016

(Software changes)

Statement issued by TÜV Rheinland Energy GmbH dated 8 March 2017

Publication: BAuz AT 31.07.2017 B12, chapter II notification 19

UBA announcement dated 13 July 2017

(Soft- and hardware changes)

Statement issued by TÜV Rheinland Energy GmbH dated 8 December 2017

Publication: BAuz AT 26.03.2018 B8, chapter IV notification 2

UBA announcement dated 21 February 2018

(Correction of software version)

Renewal of certificates

Certificate No. 0000028729_03: 25 January 2021

Expiry date of the certificate: 25 January 2026

Notifications

Statement issued by TÜV Rheinland Energy & Environment GmbH dated 11 October 2024

Publication: BAuz AT 19.05.2025 B3, chapter IV notification 16

UBA announcement dated 2 April 2025

(new manufacturer name)

Renewal of certificates

Certificate No. 0000028729_04: 22 January 2026

Expiry date of the certificate: 25 January 2031

Calculation of overall uncertainty according to EN 14181 and EN 15267-3

Measuring system

Manufacturer	Sick AG
AMS designation	GM700-2
Serial number of units under test	8308013 / 8308014
Measuring principle	Tunable Diode Laser Spectroscopy

Test report

Test laboratory	936/21210058/B
Date of report	TÜV Rheinland 2014-04-02

Measured component

Certification range	HF
	0 - 5 mg/m³

Evaluation of the cross-sensitivity (CS)

(system with largest CS)

Sum of positive CS at zero point	0.07 mg/m³
Sum of negative CS at zero point	0.00 mg/m³
Sum of positive CS at span point	0.18 mg/m³
Sum of negative CS at span point	-0.11 mg/m³
Maximum sum of cross-sensitivities	0.18 mg/m³
Uncertainty of cross-sensitivity	u _i 0.104 mg/m³

Calculation of the combined standard uncertainty

Tested parameter

		u ²
Standard deviation from paired measurements under field conditions *	u _D 0.065 mg/m³	0.004 (mg/m³) ²
Lack of fit	u _{l,f} -0.029 mg/m³	0.001 (mg/m³) ²
Zero drift from field test	u _{d,z} 0.072 mg/m³	0.005 (mg/m³) ²
Span drift from field test	u _{d,s} 0.084 mg/m³	0.007 (mg/m³) ²
Influence of ambient temperature at span	u _t 0.060 mg/m³	0.004 (mg/m³) ²
Influence of supply voltage	u _v 0.017 mg/m³	0.000 (mg/m³) ²
Cross-sensitivity (interference)	u _i 0.104 mg/m³	0.011 (mg/m³) ²
Influence of sample gas pressure	u _o 0.050 mg/m³	0.003 (mg/m³) ²
Uncertainty of reference material at 70% of certification range	u _{rm} 0.040 mg/m³	0.002 (mg/m³) ²
Excursion of measurement beam	u _{mb} 0.035 mg/m³	0.001 (mg/m³) ²

* The larger value is used :

"Repeatability standard deviation at span" or

"Standard deviation from paired measurements under field conditions"

Combined standard uncertainty (u _c)	$u_c = \sqrt{\sum (u_{max,j})^2}$	0.19 mg/m³
Total expanded uncertainty	$U = u_c * k = u_c * 1.96$	0.38 mg/m³

Relative total expanded uncertainty

Requirement of 2010/75/EU

Requirement of EN 15267-3

U in % of the ELV 1 mg/m³ 37.9

U in % of the ELV 1 mg/m³ 40.0

U in % of the ELV 1 mg/m³ 30.0