

# CERTIFICATE

## of Product Conformity (QAL1)

Certificate No.: 0000081150\_02

**Certified AMS:** Set CEM CERT II 7MB1957 for CO, NO, NO<sub>2</sub>, NO<sub>x</sub>, SO<sub>2</sub>, O<sub>2</sub> and CO<sub>2</sub>

**Manufacturer:** Siemens  
Östliche Rheinbrückenstr. 50  
76187 Karlsruhe  
Germany

**Test Institute:** TÜV Rheinland Energy & Environment GmbH

**This is to certify that the AMS has been tested  
and found to comply with the standards  
EN 15267-1 (2009), EN 15267-2 (2023), EN 15267-3 (2007)  
as well as EN 14181 (2014).**

Certification is awarded in respect of the conditions stated in this certificate  
(this certificate contains 14 pages).  
The present certificate replaces certificate 0000081150\_01 dated 5 September 2023.



Suitability Tested  
EN 15267  
QAL1 Certified  
Regular  
Surveillance

www.tuv.com  
ID 0000081150

Publication in the German Federal Gazette  
(BAnz) of 10 May 2024


German Environment Agency  
Dessau, 12 June 2024

This certificate will expire on:  
9 May 2029

TÜV Rheinland Energy &  
Environment GmbH  
Cologne, 11 June 2024



Dr. Marcel Langner  
Head of Section II 4



ppa. Dr. Peter Wilbring

[www.umwelt-tuv.eu](http://www.umwelt-tuv.eu)  
tre@umwelt-tuv.eu  
Tel. + 49 221 806-5200

TÜV Rheinland Energy & Environment GmbH  
Am Grauen Stein  
51105 Köln

Test institute accredited to EN ISO/IEC 17025 by DAkkS (German Accreditation Body).  
This accreditation is limited to the accreditation scope defined in the enclosure to the certificate D-PL-11120-02-00.

**Test report:** EuL/21258935/A dated 29 September 2023  
**Initial certification:** 20 March 2023  
**Expiry date:** 9 May 2029  
**Publication:** BAnz AT 10.05.2024 B7, chapter I No. 5.2

### Approved application

The tested AMS is suitable for use at plants according to Directive 2010/75/EC, chapter III (combustion plants / 13th BImSchV:2021), chapter IV (waste incineration plants / 17th BImSchV:2021), Directive 2015/2193/EC (44th BImSchV:2022), TA Luft:2021, 30th BImSchV:2019 and 27th BImSchV:2013. The measured ranges have been selected so as to ensure as broad a field of application as possible.

The suitability of the AMS for this application was assessed on the basis of a laboratory test and a twelve month field test at a waste incineration.

The AMS is approved for an ambient temperature range of +5 °C to 40 °C.

The notification of suitability of the AMS, performance testing and the uncertainty calculation have been effected on the basis of the regulations applicable at the time of testing. As changes in legal provisions are possible, any potential user should ensure that this AMS is suitable for monitoring the emission limit values and oxygen concentration relevant to the application.

Any potential user should ensure, in consultation with the manufacturer, that this AMS is suitable for the installation at which it will be installed.

### Note

The legal regulations mentioned correspond to the current state of legislation during certification. Each user should, if necessary, in consultation with the competent authority, ensure that this AMS meets the legal requirements for the intended use. In addition, it cannot be ruled out that legal regulations governing the use of a measuring device for emission monitoring may change during the lifetime of the certificate.

### Basis of the certification

This certification is based on:

- Test report EuL/21258935/A dated 29 September 2023 of TÜV Rheinland Energy GmbH
- Suitability announced by the German Federal Environment Agency (UBA) as the relevant body
- The ongoing surveillance of the product and the manufacturing process

Publication in the German Federal Gazette: BAnz AT 10.05.2024 B7, chapter I No. 5.2,  
Announcement by UBA dated 19 March 2024:

**AMS designation:**

SET CEM CERT II 7MB1957 for CO, NO, NO<sub>2</sub>, NO<sub>x</sub>, SO<sub>2</sub>, CO<sub>2</sub> and O<sub>2</sub>

**Manufacturer:**

Siemens AG, Karlsruhe, Germany

**Field of application:**

Modular measuring system for plants requiring official approval and for plants according to the 27th BImSchV

**Measuring ranges during the performance test:**

Component	Module	Certification range	Supplementary measuring ranges		Unit
	<b>Ultramat23-7MB235a-0bcd6-3efg</b>				
CO	a=5; bc=(AG,AJ) <sup>1</sup>	0 - 50	0 - 1250	0 - 3000	mg/m <sup>3</sup>
	a=7; (bc=(AG,AJ) <sup>1</sup> or ef=AA,(AG,AJ) <sup>1</sup> )				
	a=8; bc=BM,(AK,AS) <sup>1</sup>				
NO <sub>x</sub>	a=7; (bc=PA,(PF,PG,PH,PU,PV,PW) <sup>1</sup> or ef=(PF,PG,PH,PU,PV,PW) <sup>1</sup> )	0 - 50	0 - 2000	-	mg/m <sup>3</sup>
	a=8; bc=AS <sup>1</sup>				
NO	a=5; bc=PA,(PF,PG,PH,PU,PV,PW) <sup>1</sup>	0 - 50	0 - 1000	-	mg/m <sup>3</sup>
	a=7; (bc=PA,(PF,PG,PH,PU,PV,PW) <sup>1</sup> or ef=(PF,PG,PH,PU,PV,PW) <sup>1</sup> )				
	a=8; bc=(AK,AS) <sup>1</sup>				
NO <sub>2</sub>	a=5; bc=NS	0 - 50	0 - 1000	-	mg/m <sup>3</sup>
	a=7,8; ef=NS				
SO <sub>2</sub>	a=5; bc=NS,(NF,NG,NH,NW) <sup>1</sup>	0 - 70	0 - 1250	-	mg/m <sup>3</sup>
	a=7; (bc=(NF,NG,NH,NW) <sup>1</sup> or ef=NS,(NF,NG,NH,NW) <sup>1</sup> )				
	a=8; ef=NS,(NF,NG,NH,NW) <sup>1</sup>				
CO <sub>2</sub>	a=5; bc=CP	0 - 25	-	-	Vol.-%
	a=7; (bc=CP oder ef=CP)				
	a=8; bc=BM				
O <sub>2</sub> electrochemical	a=5,7,8; d=1	0 - 25	-	-	Vol.-%

<sup>1</sup> supplementary measuring ranges

**Software version:**

ULTRAMAT 23-7MB2355 4.02.13

ULTRAMAT 23-7MB2357 4.02.13

ULTRAMAT 23-7MB2358 4.02.13

SIEMENS SIMATIC Set CEM CERT 7MB1957 Rev. 3.0.5

**Restrictions:**

none

**Notes:**

1. The modules of the ULTRAMAT 23 series must be operated with an interval of 24 h for automatic zero point adjustment.
2. The maintenance interval is six months.
3. The modular measuring system set CEM CERT II 7MB1957 includes a system cabinet with housing protection class IP40. The system cabinet can be equipped with a climate control unit or a fan unit.
4. The measuring system has a digital interface for data transmission in accordance with the VDI 4201 standard part 1 (General requirements), part 3 (Modbus TCP/IP) and part 4 (OPC).
5. The measuring system can be operated with the following sample gas cooler models: RC1.2+ and EGK 2-19 (+) from Bühler Technologies GmbH and MAK20-2 from AGT-PSG GmbH.
6. Supplementary test (approval of further sample gas coolers) with regard to the announcement of the Federal Environment Agency of 5 July 2023 (BAnz AT 02.08.2023 B7, Chapter I Number 3.3).

**Test institute:**

TÜV Rheinland Energy GmbH, Cologne

Report No.: EuL/21258935/A dated 29 September 2023

## Certified product

This certificate applies to automated measurement systems conforming to the following description:

The entire tested modular measuring equipment Set CEM CERT II 7MB1957 is composed of a heated sample gas sampling probe, the heated sample gas line, the two-stage sample gas cooler, the sample gas feed pump and a maximum of three multi-component analyzers from the possible analyzers Ultramat 23-7MB2355, Ultramat 23-7MB2357 or Ultramat 23-7MB2358.

Measuring cabinet Set CEM CERT II 7MB1957 System cabinet

### Probe in test

Manufacturer: Bühler Technologies GmbH  
Type: GAS 222.20-Cal-twin incl. ceramic filter (length 100 cm), heated 180 °C

### Heated sample gas line

Temperature: 180 °C  
Length: 50 m in the field, 10 m in the laboratory,  
Diameter (inside): 4 mm  
Material: PTFE

### Compressor cooler in testing

Manufacturer: Bühler Technologies GmbH  
Type: RC1.2+, 2 cooling stages, dew point at 4 °C

### Alternative cooler models

Manufacturer: Bühler Technologies GmbH  
Type: EGK 2-19 (+), 2 cooling stages, dew point at 5 °C  
Manufacturer: AGT-PSG GmbH  
Type: MAK20-2, 2 cooling stages, dew point at 4 °C

### Sample gas pump

Manufacturer: Bühler Technologies GmbH  
Type: P 2.3

### Analytical modules

Manufacturer: Siemens AG  
Type: Ultramat 23-7MB2355  
Ultramat 23-7MB2357  
Ultramat 23-7MB2358

The modular measuring system Set CEM CERT II 7MB1957 includes a system cabinet with housing protection class IP40. The system cabinet can be equipped with an air-conditioning unit or with a fan unit.

The sample gas pump with integrated gas recirculation for adjusting the sample gas flows is located between the first and second cooler stages. A fine filter for fine dust separation is also integrated into the cooler housing. Downstream of the sample gas cooler, the gas path splits into either two or three sections and supplies the analyzer modules arranged in parallel with sample gas. The excess gas flows off via a bypass, if necessary. Immediately upstream of each analyzer module is another condensate filter which closes the gas path in the event

of moisture breakthrough in order to protect the analyzers. To connect zero gas for automatic zero point setting (AutoCal), a three-way valve is installed upstream of the pump, which is switched by the SIMATIC.

For the connection of zero/test gases, a further three-way valve is installed downstream of the pump which, if necessary, can offer corresponding gases for the automatic adjustment of zero and reference point - switched time-controlled by the SIMATIC. Alternatively, the test gases can also be supplied manually via a third three-way valve.

### General notes

This certificate is based upon the equipment tested. The manufacturer is responsible for ensuring that on-going production complies with the requirements of the EN 15267. The manufacturer is required to maintain an approved quality management system controlling the manufacture of the certified product. Both the product and the quality management systems shall be subject to regular surveillance.

If a product of the current production does not conform to the certified product, TÜV Rheinland Energy & Environment GmbH must be notified at the address given on page 1.

A certification mark with an ID-Number that is specific to the certified product is presented on page 1 of this certificate. This certification mark may be applied to the product or used in advertising materials for the certified product.

This document as well as the certification mark remains property of TÜV Rheinland Energy & Environment GmbH. With revocation of the publication the certificate loses its validity. After the expiration of the certificate and on requests of the TÜV Rheinland Energy & Environment GmbH this document shall be returned and the certificate mark must not be employed anymore.

The relevant version of this certificate and its expiration is also accessible on the internet: [qal1.de](http://qal1.de).

### **History of documents**

Certification of Set CEM CERT II 7MB1957 is based on the documents listed below and the regular, continuous monitoring of the Quality Management System of the manufacturer:

#### **Initial certification according to EN 15267**

Certificate No. 0000081150\_00: 25 April 2023  
Expiry date of the certificate: 19 March 2028  
Test report: 936/21253799/A dated 5 August 2022  
TÜV Rheinland Energy GmbH  
Publication: BAnz AT 20.03.2023 B6, chapter I number 3.3  
UBA announcement dated 21 February 2023

#### **Supplementary testing according to EN 15267**

Certificate No. 0000081150\_01: 5 September 2023  
Expiry date of the certificate: 1 August 2028  
Test report: 936/21253799/B dated 3 February 2023  
TÜV Rheinland Energy GmbH  
Publication: BAnz AT 02.08.2023 B7, chapter I number 3.3  
UBA announcement dated 5 July 2023

#### **Supplementary testing according to EN 15267**

Certificate No. 0000081150\_02: 12 June 2024  
Expiry date of the certificate: 9 May 2029  
Test report: EuL/21258935/A dated 29 September 2023  
TÜV Rheinland Energy GmbH  
Publication: BAnz AT 10.05.2024 B7, chapter I number 5.2  
UBA announcement dated 19 March 2024

**Calculation of overall uncertainty according to EN 14181 and EN 15267-3**

**Measuring system**

Manufacturer	SIEMENS AG
AMS designation	Set CEM CERT II 7MB1957
Serial number of units under test	TÜV 1/TÜV 2
Measuring principle	NDIR

**Test report**

Test laboratory	EuL/21253799/B TÜV Rheinland
-----------------	---------------------------------

**Measured component**

Certification range	CO 0 - 50 mg/m³
---------------------	--------------------

**Evaluation of the cross-sensitivity (CS)**

(system with largest CS)

Sum of positive CS at zero point	0.23 mg/m³
Sum of negative CS at zero point	0.00 mg/m³
Sum of positive CS at span point	0.00 mg/m³
Sum of negative CS at span point	-0.30 mg/m³
Maximum sum of cross-sensitivities	-0.30 mg/m³
Uncertainty of cross-sensitivity	$u_i$ -0.173 mg/m³

**Calculation of the combined standard uncertainty**

Tested parameter		$u^2$
Lack of fit	$u_D$ 0.309 mg/m³	0.095 (mg/m³)²
Zero drift from field test	$u_{krf}$ -0.231 mg/m³	0.053 (mg/m³)²
Span drift from field test	$u_{d,z}$ 0.115 mg/m³	0.013 (mg/m³)²
Influence of ambient temperature at span	$u_{d,s}$ 0.462 mg/m³	0.213 (mg/m³)²
Influence of supply voltage	$u_t$ 0.379 mg/m³	0.144 (mg/m³)²
Cross-sensitivity (interference)	$u_v$ 0.107 mg/m³	0.011 (mg/m³)²
Influence of sample gas flow	$u_i$ -0.173 mg/m³	0.030 (mg/m³)²
Uncertainty of reference material at 70% of certification range	$u_n$ 0.196 mg/m³	0.038 (mg/m³)²
	$u_{rm}$ 0.404 mg/m³	0.163 (mg/m³)²

\* The larger value is used :  
"Repeatability standard deviation at set point" or  
"Standard deviation from paired measurements under field conditions"

Combined standard uncertainty ( $u_c$ )	$u_c = \sqrt{\sum (u_{max,i})^2}$	0.87 mg/m³
Total expanded uncertainty	$U = u_c * k = u_c * 1.96$	1.71 mg/m³

<b>Relative total expanded uncertainty</b>	<b>U in % of the ELV 33.3 mg/m³</b>	<b>5.1</b>
<b>Requirement of 2010/75/EU</b>	<b>U in % of the ELV 33.3 mg/m³</b>	<b>10.0</b>
<b>Requirement of EN 15267-3</b>	<b>U in % of the ELV 33.3 mg/m³</b>	<b>7.5</b>



**Calculation of overall uncertainty according to EN 14181 and EN 15267-3**

**Measuring system**

Manufacturer	SIEMENS AG
AMS designation	Set CEM CERT II 7MB1957
Serial number of units under test	TÜV 1/TÜV 2
Measuring principle	NDIR

**Test report**

Test laboratory	EuL/21253799/B TÜV Rheinland
-----------------	---------------------------------

**Measured component**

Certification range	CO <sub>2</sub> 0 - 25 Vol.-%
---------------------	----------------------------------

**Evaluation of the cross-sensitivity (CS)**

(system with largest CS)

Sum of positive CS at zero point	0.00 Vol.-%
Sum of negative CS at zero point	0.00 Vol.-%
Sum of positive CS at span point	0.00 Vol.-%
Sum of negative CS at span point	-0.10 Vol.-%
Maximum sum of cross-sensitivities	0.10 Vol.-%
Uncertainty of cross-sensitivity	$u_i$ 0.058 Vol.-%

**Calculation of the combined standard uncertainty**

**Tested parameter**

			$u^2$
Lack of fit	$u_D$ 0.105 Vol.-%		0.011 (Vol.-%) <sup>2</sup>
Zero drift from field test	$u_{lof}$ 0.058 Vol.-%		0.003 (Vol.-%) <sup>2</sup>
Span drift from field test	$u_{d,z}$ 0.029 Vol.-%		0.001 (Vol.-%) <sup>2</sup>
Influence of ambient temperature at span	$u_{d,s}$ 0.130 Vol.-%		0.017 (Vol.-%) <sup>2</sup>
Influence of supply voltage	$u_f$ 0.115 Vol.-%		0.013 (Vol.-%) <sup>2</sup>
Cross-sensitivity (interference)	$u_v$ 0.000 Vol.-%		0.000 (Vol.-%) <sup>2</sup>
Influence of sample gas flow	$u_i$ 0.058 Vol.-%		0.003 (Vol.-%) <sup>2</sup>
Uncertainty of reference material at 70% of certification range	$u_n$ 0.058 Vol.-%		0.003 (Vol.-%) <sup>2</sup>
	$u_{rm}$ 0.202 Vol.-%		0.041 (Vol.-%) <sup>2</sup>

\* The larger value is used :  
"Repeatability standard deviation at set point" or  
"Standard deviation from paired measurements under field conditions"

Combined standard uncertainty ( $u_c$ )	$u_c = \sqrt{\sum (u_{max,i})^2}$	0.30 Vol.-%
Total expanded uncertainty	$U = u_c * k = u_c * 1.96$	0.60 Vol.-%

**Relative total expanded uncertainty**

Requirement of 2010/75/EU	<b>U in % of the range 25 Vol.-%</b>	<b>2.4</b>
Requirement of EN 15267-3	<b>U in % of the range 25 Vol.-%</b>	<b>10.0 **</b>
	<b>U in % of the range 25 Vol.-%</b>	<b>7.5</b>

\*\* The EU-directive 2010/75/EC on industrial emissions does not define requirements for this component.  
A value of 10.0 % was used instead.

**Calculation of overall uncertainty according to EN 14181 and EN 15267-3**

**Measuring system**

Manufacturer	SIEMENS AG
AMS designation	Set CEM CERT II 7MB1957
Serial number of units under test	TÜV 1/TÜV 2
Measuring principle	NDIR

**Test report**

Test laboratory	EuL/21253799/B TÜV Rheinland
-----------------	---------------------------------

**Measured component**

Certification range	NO 0 - 50 mg/m <sup>3</sup>
---------------------	--------------------------------

**Evaluation of the cross-sensitivity (CS)**

(system with largest CS)

Sum of positive CS at zero point	1.19 mg/m <sup>3</sup>
Sum of negative CS at zero point	-0.97 mg/m <sup>3</sup>
Sum of positive CS at span point	1.10 mg/m <sup>3</sup>
Sum of negative CS at span point	-0.70 mg/m <sup>3</sup>
Maximum sum of cross-sensitivities	1.19 mg/m <sup>3</sup>
Uncertainty of cross-sensitivity	$u_i$ 0.687 mg/m <sup>3</sup>

**Calculation of the combined standard uncertainty**

**Tested parameter**

		$u^2$
Lack of fit	$u_D$ 0.643 mg/m <sup>3</sup>	0.413 (mg/m <sup>3</sup> ) <sup>2</sup>
Zero drift from field test	$u_{krf}$ 0.346 mg/m <sup>3</sup>	0.120 (mg/m <sup>3</sup> ) <sup>2</sup>
Span drift from field test	$u_{d,z}$ 0.173 mg/m <sup>3</sup>	0.030 (mg/m <sup>3</sup> ) <sup>2</sup>
Influence of ambient temperature at span	$u_{d,s}$ 0.635 mg/m <sup>3</sup>	0.403 (mg/m <sup>3</sup> ) <sup>2</sup>
Influence of supply voltage	$u_t$ 0.346 mg/m <sup>3</sup>	0.120 (mg/m <sup>3</sup> ) <sup>2</sup>
Cross-sensitivity (interference)	$u_v$ 0.156 mg/m <sup>3</sup>	0.024 (mg/m <sup>3</sup> ) <sup>2</sup>
Influence of sample gas flow	$u_i$ 0.687 mg/m <sup>3</sup>	0.472 (mg/m <sup>3</sup> ) <sup>2</sup>
Uncertainty of reference material at 70% of certification range	$u_n$ 0.115 mg/m <sup>3</sup>	0.013 (mg/m <sup>3</sup> ) <sup>2</sup>
	$u_{rm}$ 0.404 mg/m <sup>3</sup>	0.163 (mg/m <sup>3</sup> ) <sup>2</sup>

\* The larger value is used :

"Repeatability standard deviation at set point" or

"Standard deviation from paired measurements under field conditions"

Combined standard uncertainty ( $u_c$ )

$$u_c = \sqrt{\sum (u_{max,i})^2} \quad 1.33 \text{ mg/m}^3$$

Total expanded uncertainty

$$U = u_c * k = u_c * 1.96 \quad 2.60 \text{ mg/m}^3$$

**Relative total expanded uncertainty**

**U in % of the ELV 33.3 mg/m<sup>3</sup> 7.8**

Requirement of 2010/75/EU

**U in % of the ELV 33.3 mg/m<sup>3</sup> 20.0**

Requirement of EN 15267-3

**U in % of the ELV 33.3 mg/m<sup>3</sup> 15.0**

**Calculation of overall uncertainty according to EN 14181 and EN 15267-3**

**Measuring system**

Manufacturer	SIEMENS AG
AMS designation	Set CEM CERT II 7MB1957
Serial number of units under test	TÜV 1/TÜV 2
Measuring principle	UV Absorption

**Test report**

Test laboratory	EuL/21253799/B TÜV Rheinland
-----------------	---------------------------------

**Measured component**

Certification range	NO <sub>2</sub> 0 - 50 mg/m <sup>3</sup>
---------------------	---

**Evaluation of the cross-sensitivity (CS)**

(system with largest CS)

Sum of positive CS at zero point	0.68 mg/m <sup>3</sup>
Sum of negative CS at zero point	0.00 mg/m <sup>3</sup>
Sum of positive CS at span point	0.60 mg/m <sup>3</sup>
Sum of negative CS at span point	-0.80 mg/m <sup>3</sup>
Maximum sum of cross-sensitivities	-0.80 mg/m <sup>3</sup>
Uncertainty of cross-sensitivity	$u_i$ -0.462 mg/m <sup>3</sup>

**Calculation of the combined standard uncertainty**

Tested parameter		$u^2$
Lack of fit	$u_D$ 0.468 mg/m <sup>3</sup>	0.219 (mg/m <sup>3</sup> ) <sup>2</sup>
Zero drift from field test	$u_{lof}$ 0.173 mg/m <sup>3</sup>	0.030 (mg/m <sup>3</sup> ) <sup>2</sup>
Span drift from field test	$u_{d,z}$ 0.144 mg/m <sup>3</sup>	0.021 (mg/m <sup>3</sup> ) <sup>2</sup>
Influence of ambient temperature at span	$u_{d,s}$ 0.508 mg/m <sup>3</sup>	0.258 (mg/m <sup>3</sup> ) <sup>2</sup>
Influence of supply voltage	$u_t$ 0.321 mg/m <sup>3</sup>	0.103 (mg/m <sup>3</sup> ) <sup>2</sup>
Cross-sensitivity (interference)	$u_v$ 0.313 mg/m <sup>3</sup>	0.098 (mg/m <sup>3</sup> ) <sup>2</sup>
Influence of sample gas flow	$u_i$ -0.462 mg/m <sup>3</sup>	0.213 (mg/m <sup>3</sup> ) <sup>2</sup>
Uncertainty of reference material at 70% of certification range	$u_n$ 0.115 mg/m <sup>3</sup>	0.013 (mg/m <sup>3</sup> ) <sup>2</sup>
	$u_{rm}$ 0.404 mg/m <sup>3</sup>	0.163 (mg/m <sup>3</sup> ) <sup>2</sup>

\* The larger value is used :  
"Repeatability standard deviation at set point" or  
"Standard deviation from paired measurements under field conditions"

Combined standard uncertainty ( $u_c$ )	$u_c = \sqrt{\sum (u_{max,i})^2}$	1.06 mg/m <sup>3</sup>
Total expanded uncertainty	$U = u_c * k = u_c * 1.96$	2.07 mg/m <sup>3</sup>

<b>Relative total expanded uncertainty</b>	<b>U in % of the ELV 33.3 mg/m<sup>3</sup></b>	<b>6.2</b>
<b>Requirement of 2010/75/EU</b>	<b>U in % of the ELV 33.3 mg/m<sup>3</sup></b>	<b>20.0</b>
<b>Requirement of EN 15267-3</b>	<b>U in % of the ELV 33.3 mg/m<sup>3</sup></b>	<b>15.0</b>

**Calculation of overall uncertainty according to EN 14181 and EN 15267-3**

**Measuring system**

Manufacturer	SIEMENS AG
AMS designation	Set CEM CERT II 7MB1957
Serial number of units under test	TÜV 1/TÜV 2
Measuring principle	calculated

**Test report**

Test laboratory	EuL/21253799/B TÜV Rheinland
-----------------	---------------------------------

**Measured component**

Certification range	NOx 0 - 50 mg/m <sup>3</sup>
---------------------	---------------------------------

**Evaluation of the cross-sensitivity (CS)**

(system with largest CS)

Sum of positive CS at zero point	1.65 mg/m <sup>3</sup>
Sum of negative CS at zero point	-0.86 mg/m <sup>3</sup>
Sum of positive CS at span point	0.00 mg/m <sup>3</sup>
Sum of negative CS at span point	-0.70 mg/m <sup>3</sup>
Maximum sum of cross-sensitivities	1.65 mg/m <sup>3</sup>
Uncertainty of cross-sensitivity	$u_i$ 0.953 mg/m <sup>3</sup>

**Calculation of the combined standard uncertainty**

Tested parameter		$u^2$
Lack of fit	$u_D$ 1.035 mg/m <sup>3</sup>	1.071 (mg/m <sup>3</sup> ) <sup>2</sup>
Zero drift from field test	$u_{lof}$ 0.173 mg/m <sup>3</sup>	0.030 (mg/m <sup>3</sup> ) <sup>2</sup>
Span drift from field test	$u_{d,z}$ 0.177 mg/m <sup>3</sup>	0.031 (mg/m <sup>3</sup> ) <sup>2</sup>
Influence of ambient temperature at span	$u_{d,s}$ 0.574 mg/m <sup>3</sup>	0.329 (mg/m <sup>3</sup> ) <sup>2</sup>
Influence of supply voltage	$u_t$ 0.586 mg/m <sup>3</sup>	0.343 (mg/m <sup>3</sup> ) <sup>2</sup>
Cross-sensitivity (interference)	$u_v$ 0.313 mg/m <sup>3</sup>	0.098 (mg/m <sup>3</sup> ) <sup>2</sup>
Influence of sample gas flow	$u_i$ 0.953 mg/m <sup>3</sup>	0.908 (mg/m <sup>3</sup> ) <sup>2</sup>
Uncertainty of reference material at 70% of certification range	$u_o$ -0.214 mg/m <sup>3</sup>	0.046 (mg/m <sup>3</sup> ) <sup>2</sup>
	$u_{rm}$ 0.404 mg/m <sup>3</sup>	0.163 (mg/m <sup>3</sup> ) <sup>2</sup>

\* The larger value is used :  
"Repeatability standard deviation at set point" or  
"Standard deviation from paired measurements under field conditions"

Combined standard uncertainty ( $u_c$ )	$u_c = \sqrt{\sum (u_{max,i})^2}$	1.74 mg/m <sup>3</sup>
Total expanded uncertainty	$U = u_c * k = u_c * 1.96$	3.41 mg/m <sup>3</sup>

<b>Relative total expanded uncertainty</b>	<b>U in % of the ELV 33.3 mg/m<sup>3</sup></b>	<b>10.2</b>
<b>Requirement of 2010/75/EU</b>	<b>U in % of the ELV 33.3 mg/m<sup>3</sup></b>	<b>20.0</b>
<b>Requirement of EN 15267-3</b>	<b>U in % of the ELV 33.3 mg/m<sup>3</sup></b>	<b>15.0</b>

**Calculation of overall uncertainty according to EN 14181 and EN 15267-3**

**Measuring system**

Manufacturer	SIEMENS AG
AMS designation	Set CEM CERT II 7MB1957
Serial number of units under test	TÜV 1/TÜV 2
Measuring principle	electrochemical

**Test report**

Test laboratory	EuL/21253799/B TÜV Rheinland
-----------------	---------------------------------

**Measured component**

Certification range	O <sub>2</sub> 0 - 25 Vol.-%
---------------------	---------------------------------

**Evaluation of the cross-sensitivity (CS)**

(system with largest CS)

Sum of positive CS at zero point	0.00 Vol.-%
Sum of negative CS at zero point	-0.10 Vol.-%
Sum of positive CS at span point	0.00 Vol.-%
Sum of negative CS at span point	0.00 Vol.-%
Maximum sum of cross-sensitivities	-0.10 Vol.-%
Uncertainty of cross-sensitivity	u <sub>i</sub> -0.058 Vol.-%

**Calculation of the combined standard uncertainty**

**Tested parameter**

			u <sup>2</sup>
Lack of fit	u <sub>D</sub> 0.038 Vol.-%		0.001 (Vol.-%) <sup>2</sup>
Zero drift from field test	u <sub>lof</sub> -0.058 Vol.-%		0.003 (Vol.-%) <sup>2</sup>
Span drift from field test	u <sub>d,z</sub> 0.058 Vol.-%		0.003 (Vol.-%) <sup>2</sup>
Influence of ambient temperature at span	u <sub>d,s</sub> 0.104 Vol.-%		0.011 (Vol.-%) <sup>2</sup>
Influence of supply voltage	u <sub>f</sub> 0.064 Vol.-%		0.004 (Vol.-%) <sup>2</sup>
Cross-sensitivity (interference)	u <sub>v</sub> 0.021 Vol.-%		0.000 (Vol.-%) <sup>2</sup>
Influence of sample gas flow	u <sub>i</sub> -0.058 Vol.-%		0.003 (Vol.-%) <sup>2</sup>
Uncertainty of reference material at 70% of certification range	u <sub>n</sub> 0.006 Vol.-%		0.000 (Vol.-%) <sup>2</sup>
	u <sub>rm</sub> 0.202 Vol.-%		0.041 (Vol.-%) <sup>2</sup>

\* The larger value is used :  
"Repeatability standard deviation at set point" or  
"Standard deviation from paired measurements under field conditions"

Combined standard uncertainty (u <sub>c</sub> )	$u_c = \sqrt{\sum (u_{max,i})^2}$	0.26 Vol.-%
Total expanded uncertainty	U = u <sub>c</sub> * k = u <sub>c</sub> * 1.96	0.51 Vol.-%

**Relative total expanded uncertainty**

Requirement of 2010/75/EU	<b>U in % of the range 25 Vol.-%</b>	<b>2.0</b>
Requirement of EN 15267-3	<b>U in % of the range 25 Vol.-%</b>	<b>10.0 **</b>
	<b>U in % of the range 25 Vol.-%</b>	<b>7.5</b>

\*\* The EU-directive 2010/75/EC on industrial emissions does not define requirements for this component.  
A value of 10.0 % was used instead.

**Calculation of overall uncertainty according to EN 14181 and EN 15267-3**

**Measuring system**

Manufacturer	SIEMENS AG
AMS designation	Set CEM CERT II 7MB1957
Serial number of units under test	TÜV 1/TÜV 2
Measuring principle	UV Absorption

**Test report**

Test laboratory	EuL/21253799/B TÜV Rheinland
-----------------	---------------------------------

**Measured component**

Certification range	SO <sub>2</sub> 0 - 70 mg/m <sup>3</sup>
---------------------	---

**Evaluation of the cross-sensitivity (CS)**

(system with largest CS)

Sum of positive CS at zero point	0.54 mg/m <sup>3</sup>
Sum of negative CS at zero point	-0.61 mg/m <sup>3</sup>
Sum of positive CS at span point	2.20 mg/m <sup>3</sup>
Sum of negative CS at span point	-1.20 mg/m <sup>3</sup>
Maximum sum of cross-sensitivities	2.20 mg/m <sup>3</sup>
Uncertainty of cross-sensitivity	$u_i$ 1.269 mg/m <sup>3</sup>

**Calculation of the combined standard uncertainty**

**Tested parameter**

		$u^2$
Lack of fit	$u_D$ 0.203 mg/m <sup>3</sup>	0.041 (mg/m <sup>3</sup> ) <sup>2</sup>
Zero drift from field test	$u_{krf}$ -0.287 mg/m <sup>3</sup>	0.082 (mg/m <sup>3</sup> ) <sup>2</sup>
Span drift from field test	$u_{d,z}$ 0.323 mg/m <sup>3</sup>	0.104 (mg/m <sup>3</sup> ) <sup>2</sup>
Influence of ambient temperature at span	$u_{d,s}$ 0.970 mg/m <sup>3</sup>	0.941 (mg/m <sup>3</sup> ) <sup>2</sup>
Influence of supply voltage	$u_t$ 0.608 mg/m <sup>3</sup>	0.370 (mg/m <sup>3</sup> ) <sup>2</sup>
Cross-sensitivity (interference)	$u_v$ 0.176 mg/m <sup>3</sup>	0.031 (mg/m <sup>3</sup> ) <sup>2</sup>
Influence of sample gas flow	$u_i$ 1.269 mg/m <sup>3</sup>	1.610 (mg/m <sup>3</sup> ) <sup>2</sup>
Uncertainty of reference material at 70% of certification range	$u_o$ 0.289 mg/m <sup>3</sup>	0.084 (mg/m <sup>3</sup> ) <sup>2</sup>
	$u_{rm}$ 0.566 mg/m <sup>3</sup>	0.320 (mg/m <sup>3</sup> ) <sup>2</sup>

\* The larger value is used :

"Repeatability standard deviation at set point" or

"Standard deviation from paired measurements under field conditions"

Combined standard uncertainty ( $u_c$ )

$$u_c = \sqrt{\sum (u_{max,i})^2} \quad 1.89 \text{ mg/m}^3$$

Total expanded uncertainty

$$U = u_c * k = u_c * 1.96 \quad 3.71 \text{ mg/m}^3$$

**Relative total expanded uncertainty**

**U in % of the ELV 46.6 mg/m<sup>3</sup> 8.0**

**Requirement of 2010/75/EU**

**U in % of the ELV 46.6 mg/m<sup>3</sup> 20.0**

Requirement of EN 15267-3

U in % of the ELV 46.6 mg/m<sup>3</sup> 15.0