

# CONFIRMATION

Notification: 0000040214\_02\_02  
on changes according to EN 15267 regarding certificate 0000040214\_02 dated 01 July 2020

---

**AMS designation:** Modell 5014i Beta mit PM<sub>10</sub> Pre-separator for PM<sub>10</sub>

**Manufacturer:** Thermo Fisher Scientific  
27, Forge Parkway  
Franklin, MA 02038  
USA

---

**German Federal Environmental Agency (UBA)**

**Announcement regarding the uniform practice in  
monitoring emissions and ambient air  
dated 21 February 2023  
Publication: BAnz AT 20.03.2023 B6**

**IV. Notifications as regards the uniform practice in monitoring  
emissions and ambient air:**

**87 Notification as regards Federal Environment Agency (UBA) notices  
of 27 February 2014 (BAnz AT 01.04.2014 B12, chapter IV number 7.2) and  
of 29 June 2021 (BAnz AT 05.08.2021 B5, chapter IV notification 5)**

The current software version of the measuring system Model 5014i Beta with PM<sub>10</sub> pre-  
separator for suspended particulate matter PM<sub>10</sub> from Thermo Fisher Scientific is:


V 03.00.05.268


Version V 03.00.03.261 can also be used.

Statement issued by TÜV Rheinland Energy GmbH dated 16 February 2022

---

TÜV Rheinland Energy GmbH  
Cologne, 25 April 2023

  
i. V. Dipl.-Ing. Guido Baum

  
i. A. Dipl.-Ing. Carsten Röllig

[www.umwelt-tuv.eu](http://www.umwelt-tuv.eu)  
TRE@umwelt-tuv.eu  
Tel. +49 221 806-5200

TÜV Rheinland Energy GmbH  
Am Grauen Stein  
51105 Cologne

Test institute accredited to EN ISO/IEC 17025 by DAkkS (German Accreditation Body).  
This accreditation is limited to the accreditation scope defined in the enclosure to the certificate D-PL-11120-02-00.