

# CONFIRMATION

Notification: 0000040217\_02\_02  
on changes according to EN 15267 regarding certificate 0000040217\_02 dated 01 July 2020

---

**AMS designation:** Modell 42i for NO, NO<sub>2</sub> and NO<sub>x</sub>

**Manufacturer:** Thermo Fisher Scientific  
27, Forge Parkway  
Franklin, MA 02038  
USA

---

## German Federal Environmental Agency (UBA)

**Announcement regarding the uniform practice in  
monitoring emissions and ambient air  
dated 28 June 2022  
Publication: BAnz AT 28.07.2022 B4**

**III. Notifications as regards the uniform practice in monitoring  
emissions and ambient air:**

**45 Notification as regards Federal Environment Agency (UBA) notices  
of 21 February 2006 (BAnz. p. 2653, chapter IV number 4.1) and  
of 29 June 2021 (BAnz AT 05.08.2021 B5, chapter IV notification 8)**


The following hardware changes have been introduced for the Model 42i measuring device for NO, NO<sub>2</sub> and NO<sub>x</sub> from Thermo Fisher Scientific:


- The measuring device can now also be equipped with the Arcturus CPU (53281) processor board.
- The measuring device can now also be equipped with a SUNON housing fan (part no. PMD2408PMB-A).
- The pump head of the measuring device can also be equipped with a new intermediate plate made of coated PTFE and a valve plate with a stainless steel flap valve.

Statement issued by TÜV Rheinland Energy GmbH dated 18 May 2022

---

TÜV Rheinland Energy GmbH  
Cologne, 01 August 2022

  
i. V. Dipl.-Ing. Guido Baum

  
i. A. Dipl.-Ing. Carsten Röllig

[www.umwelt-tuv.eu](http://www.umwelt-tuv.eu)  
TRE@umwelt-tuv.eu  
Tel. +49 221 806-5200

TÜV Rheinland Energy GmbH  
Am Grauen Stein  
51105 Cologne

Test institute accredited to EN ISO/IEC 17025 by DAkkS (German Accreditation Body).  
This accreditation is limited to the accreditation scope defined in the enclosure to the certificate D-PL-11120-02-00.