

CONFIRMATION

Notification: 0000040219_02_03
on changes according to EN 15267 regarding certificate 0000040219_02 dated 01 July 2020

AMS designation: Modell 48i for CO

Manufacturer: Thermo Fisher Scientific
27, Forge Parkway
Franklin, MA 02038
USA

German Federal Environmental Agency (UBA)

**Announcement regarding the uniform practice in
monitoring emissions and ambient air
dated 21 February 2023
Publication: BAnz AT 20.03.2023 B6**


**IV. Notifications as regards the uniform practice in monitoring
emissions and ambient air:**


**85 Notification as regards Federal Environment Agency (UBA) notices
of 21 February 2006 (BAnz. p. 2653, chapter IV number 2.2) and
of 28 June 2022 (BAnz AT 28.07.2022 B4, chapter III notification 47)**

The Model 48i measuring system for CO from Thermo Fisher Scientific can also be equipped with the alternative IR source of type 66785 15R from the manufacturer Ohmite Dove Electronics.

Statement issued by TÜV Rheinland Energy GmbH dated 16 September 2022

TÜV Rheinland Energy GmbH
Cologne, 25 April 2023


i. V. Dipl.-Ing. Guido Baum


i. A. Dipl.-Ing. Carsten Röllig

www.umwelt-tuv.eu
TRE@umwelt-tuv.eu
Tel. +49 221 806-5200

TÜV Rheinland Energy GmbH
Am Grauen Stein
51105 Cologne

Test institute accredited to EN ISO/IEC 17025 by DAkkS (German Accreditation Body).
This accreditation is limited to the accreditation scope defined in the enclosure to the certificate D-PL-11120-02-00.